

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050621\
 Data File : PP035614.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 May 2021 17:10
 Operator : DD\AJ
 Sample : PB136200BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 18:01:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 06 11:27:40 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.784	4.203	935864	662789	21.390	21.849
2)	SA Decachlor...	10.637	9.456	879898	442658	22.308	22.943
Target Compounds							
3)	L1 AR-1016-1	6.090	5.446	690876	436565	469.056	476.376
4)	L1 AR-1016-2	6.113	5.466	1023699	669035	464.153	440.264
5)	L1 AR-1016-3	6.178	5.657	652940	352960	467.949	468.970
6)	L1 AR-1016-4	6.284	5.704	536232	274487	472.643	452.147
7)	L1 AR-1016-5	6.596	5.933	497633	354914	432.748	464.383
8)	L2 AR-1221-1	5.032	4.454	86424	46957	166.146	125.762
9)	L2 AR-1221-2	5.130	4.556	86041	57599	218.470	205.569
10)	L2 AR-1221-3	5.215	4.644	373266	261551	312.084	298.989
11)	L3 AR-1232-1	5.215	4.644	373266	261551	406.188	388.593
12)	L3 AR-1232-2	5.796	5.466	517208	669035	1038.224	1015.382
13)	L3 AR-1232-3	6.113	5.657	1023699	352960	1148.077	1165.215
14)	L3 AR-1232-4	6.284	5.749	536232	339595	1186.500	1251.684
15)	L3 AR-1232-5	6.381	5.933	409286	354914	1264.763	1228.604
16)	L4 AR-1242-1	6.090	5.446	690876	436565	604.736	615.154
17)	L4 AR-1242-2	6.113	5.466	1023699	669035	596.241	537.931
18)	L4 AR-1242-3	6.178	5.657	652940	352960	601.669	603.210
19)	L4 AR-1242-4	6.284	5.749	536232	339595	611.911	578.860
20)	L4 AR-1242-5	7.060	6.306	67331	268935	76.256	397.814 #
21)	L5 AR-1248-1	6.090	5.446	690876	436565	829.732	822.820
22)	L5 AR-1248-2	6.381	5.704	409286	274487	329.107	339.061
23)	L5 AR-1248-3	6.596	5.749	497633	339595	327.852	395.053
24)	L5 AR-1248-4	7.021	5.933	78482	354914	52.517	361.165 #
25)	L5 AR-1248-5	7.060	6.352	67331	41637	45.587	45.854
26)	L6 AR-1254-1	6.990	6.306	358277	268935	213.567	189.352
27)	L6 AR-1254-2	7.217	6.462	393886	253344	156.358	201.149 #
28)	L6 AR-1254-3	7.597	6.895	251656	435842	97.196	229.649 #
29)	L6 AR-1254-4	7.889	7.117	163164	46401	95.509	42.178 #
30)	L6 AR-1254-5	8.316	7.539	1259960	794529	609.140	473.911
31)	L7 AR-1260-1	7.762	7.013	946605	599171	436.874	442.325
32)	L7 AR-1260-2	8.027	7.207	1066657	707809	433.668	448.723
33)	L7 AR-1260-3	8.391	7.362	678515	647263	418.557	433.692
34)	L7 AR-1260-4	8.619	7.839	818617	469788	413.534	436.074
35)	L7 AR-1260-5	8.933	8.083	1583615	1042448	428.685	438.346
36)	L8 AR-1262-1	8.391	7.539	678515	794529	258.813	919.433 #
37)	L8 AR-1262-2	8.933	8.083	1583615	1042448	379.429	379.415
38)	L8 AR-1262-3	9.249f	8.364	1036671	210130	337.873	174.223 #
39)	L8 AR-1262-4	9.298f	8.429	597874	762759	250.128	357.425 #
40)	L8 AR-1262-5	9.943	8.921	416458	246219	271.988	240.483
41)	L9 AR-1268-1	9.249f	8.364	1036671	210130	190.347	67.779 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050621\
 Data File : PP035614.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 May 2021 17:10
 Operator : DD\AJ
 Sample : PB136200BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 18:01:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 06 11:27:40 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.298f	8.429	597874	762759	117.283	272.172 #
43) L9	AR-1268-3	9.530	8.631	32175	19934	6.931	8.185
44) L9	AR-1268-4	9.943	8.921	416458	246219	250.710	246.614
45) L9	AR-1268-5	10.329	9.209	147760	75191	10.909	11.099



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050621\
Data File : PP035614.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 May 2021 17:10
Operator : DD\AJ
Sample : PB136200BS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 18:01:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050521.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 06 11:27:40 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.784 min
Delta R.T.: -0.003 min
Response: 935864
Conc: 21.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.203 min
Delta R.T.: 0.000 min
Response: 662789
Conc: 21.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.637 min
Delta R.T.: -0.006 min
Response: 879898
Conc: 22.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.456 min
Delta R.T.: -0.001 min
Response: 442658
Conc: 22.94 ng/ml



#3 AR-1016-1


R.T.: 6.090 min
Delta R.T.: -0.003 min
Response: 690876
Conc: 469.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.446 min
Delta R.T.: -0.002 min
Response: 436565
Conc: 476.38 ng/ml




#4 AR-1016-2

R.T.: 6.113 min
Delta R.T.: -0.003 min
Response: 1023699
Conc: 464.15 ng/ml



#4 AR-1016-2


R.T.: 5.466 min
Delta R.T.: -0.003 min
Response: 669035
Conc: 440.26 ng/ml



#5 AR-1016-3


R.T.: 6.178 min
Delta R.T.: -0.004 min
Response: 652940
Conc: 467.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.657 min
Delta R.T.: -0.002 min
Response: 352960
Conc: 468.97 ng/ml




#6 AR-1016-4

R.T.: 6.284 min
Delta R.T.: -0.003 min
Response: 536232
Conc: 472.64 ng/ml



#6 AR-1016-4


R.T.: 5.704 min
Delta R.T.: -0.002 min
Response: 274487
Conc: 452.15 ng/ml



#7 AR-1016-5


R.T.: 6.596 min
Delta R.T.: -0.003 min
Response: 497633
Conc: 432.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.933 min
Delta R.T.: -0.001 min
Response: 354914
Conc: 464.38 ng/ml




#8 AR-1221-1

R.T.: 5.032 min
Delta R.T.: -0.003 min
Response: 86424
Conc: 166.15 ng/ml



#8 AR-1221-1


R.T.: 4.454 min
Delta R.T.: -0.004 min
Response: 46957
Conc: 125.76 ng/ml



#9 AR-1221-2


R.T.: 5.130 min
Delta R.T.: -0.003 min
Response: 86041
Conc: 218.47 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.556 min
Delta R.T.: -0.001 min
Response: 57599
Conc: 205.57 ng/ml




#10 AR-1221-3

R.T.: 5.215 min
Delta R.T.: -0.003 min
Response: 373266
Conc: 312.08 ng/ml



#10 AR-1221-3


R.T.: 4.644 min
Delta R.T.: -0.002 min
Response: 261551
Conc: 298.99 ng/ml



#11 AR-1232-1


R.T.: 5.215 min
Delta R.T.: -0.003 min
Response: 373266
Conc: 406.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.644 min
Delta R.T.: -0.002 min
Response: 261551
Conc: 388.59 ng/ml




#12 AR-1232-2

R.T.: 5.796 min
Delta R.T.: -0.002 min
Response: 517208
Conc: 1038.22 ng/ml



#12 AR-1232-2


R.T.: 5.466 min
Delta R.T.: -0.003 min
Response: 669035
Conc: 1015.38 ng/ml



#13 AR-1232-3


R.T.: 6.113 min
Delta R.T.: -0.003 min
Response: 1023699
Conc: 1148.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.657 min
Delta R.T.: -0.002 min
Response: 352960
Conc: 1165.21 ng/ml




#14 AR-1232-4

R.T.: 6.284 min
Delta R.T.: -0.003 min
Response: 536232
Conc: 1186.50 ng/ml



#14 AR-1232-4


R.T.: 5.749 min
Delta R.T.: -0.002 min
Response: 339595
Conc: 1251.68 ng/ml



#15 AR-1232-5


R.T.: 6.381 min
Delta R.T.: -0.003 min
Response: 409286
Conc: 1264.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.933 min
Delta R.T.: -0.001 min
Response: 354914
Conc: 1228.60 ng/ml




#16 AR-1242-1

R.T.: 6.090 min
Delta R.T.: -0.003 min
Response: 690876
Conc: 604.74 ng/ml



#16 AR-1242-1


R.T.: 5.446 min
Delta R.T.: -0.003 min
Response: 436565
Conc: 615.15 ng/ml



#17 AR-1242-2


R.T.: 6.113 min
Delta R.T.: -0.003 min
Response: 1023699
Conc: 596.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.466 min
Delta R.T.: -0.002 min
Response: 669035
Conc: 537.93 ng/ml




#18 AR-1242-3

R.T.: 6.178 min
Delta R.T.: -0.003 min
Response: 652940
Conc: 601.67 ng/ml



#18 AR-1242-3


R.T.: 5.657 min
Delta R.T.: -0.003 min
Response: 352960
Conc: 603.21 ng/ml



#19 AR-1242-4


R.T.: 6.284 min
Delta R.T.: -0.003 min
Response: 536232
Conc: 611.91 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.749 min
Delta R.T.: -0.002 min
Response: 339595
Conc: 578.86 ng/ml




#20 AR-1242-5

R.T.: 7.060 min
Delta R.T.: -0.003 min
Response: 67331
Conc: 76.26 ng/ml



#20 AR-1242-5


R.T.: 6.306 min
Delta R.T.: -0.004 min
Response: 268935
Conc: 397.81 ng/ml



#21 AR-1248-1


R.T.: 6.090 min
Delta R.T.: -0.003 min
Response: 690876
Conc: 829.73 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.446 min
Delta R.T.: -0.003 min
Response: 436565
Conc: 822.82 ng/ml




#22 AR-1248-2

R.T.: 6.381 min
Delta R.T.: -0.002 min
Response: 409286
Conc: 329.11 ng/ml



#22 AR-1248-2


R.T.: 5.704 min
Delta R.T.: -0.001 min
Response: 274487
Conc: 339.06 ng/ml



#23 AR-1248-3


R.T.: 6.596 min
Delta R.T.: -0.002 min
Response: 497633
Conc: 327.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.749 min
Delta R.T.: -0.002 min
Response: 339595
Conc: 395.05 ng/ml




#24 AR-1248-4

R.T.: 7.021 min
Delta R.T.: -0.003 min
Response: 78482
Conc: 52.52 ng/ml



#24 AR-1248-4


R.T.: 5.933 min
Delta R.T.: -0.001 min
Response: 354914
Conc: 361.17 ng/ml



#25 AR-1248-5


R.T.: 7.060 min
Delta R.T.: -0.004 min
Response: 67331
Conc: 45.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.352 min
Delta R.T.: -0.002 min
Response: 41637
Conc: 45.85 ng/ml




#26 AR-1254-1

R.T.: 6.990 min
Delta R.T.: -0.003 min
Response: 358277
Conc: 213.57 ng/ml



#26 AR-1254-1


R.T.: 6.306 min
Delta R.T.: -0.001 min
Response: 268935
Conc: 189.35 ng/ml



#27 AR-1254-2


R.T.: 7.217 min
Delta R.T.: -0.004 min
Response: 393886
Conc: 156.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.462 min
Delta R.T.: -0.002 min
Response: 253344
Conc: 201.15 ng/ml




#28 AR-1254-3

R.T.: 7.597 min
Delta R.T.: -0.003 min
Response: 251656
Conc: 97.20 ng/ml



#28 AR-1254-3


R.T.: 6.895 min
Delta R.T.: 0.013 min
Response: 435842
Conc: 229.65 ng/ml



#29 AR-1254-4


R.T.: 7.889 min
Delta R.T.: -0.005 min
Response: 163164
Conc: 95.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 7.117 min
Delta R.T.: -0.002 min
Response: 46401
Conc: 42.18 ng/ml




#30 AR-1254-5

R.T.: 8.316 min
Delta R.T.: -0.004 min
Response: 1259960
Conc: 609.14 ng/ml



#30 AR-1254-5


R.T.: 7.539 min
Delta R.T.: -0.002 min
Response: 794529
Conc: 473.91 ng/ml



#31 AR-1260-1


R.T.: 7.762 min
Delta R.T.: -0.004 min
Response: 946605
Conc: 436.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 7.013 min
Delta R.T.: -0.001 min
Response: 599171
Conc: 442.32 ng/ml




#32 AR-1260-2

R.T.: 8.027 min
Delta R.T.: -0.003 min
Response: 1066657
Conc: 433.67 ng/ml



#32 AR-1260-2


R.T.: 7.207 min
Delta R.T.: -0.001 min
Response: 707809
Conc: 448.72 ng/ml



#33 AR-1260-3


R.T.: 8.391 min
Delta R.T.: -0.003 min
Response: 678515
Conc: 418.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.362 min
Delta R.T.: -0.001 min
Response: 647263
Conc: 433.69 ng/ml




#34 AR-1260-4

R.T.: 8.619 min
Delta R.T.: -0.004 min
Response: 818617
Conc: 413.53 ng/ml



#34 AR-1260-4


R.T.: 7.839 min
Delta R.T.: -0.002 min
Response: 469788
Conc: 436.07 ng/ml



#35 AR-1260-5


R.T.: 8.933 min
Delta R.T.: -0.003 min
Response: 1583615
Conc: 428.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 8.083 min
Delta R.T.: -0.001 min
Response: 1042448
Conc: 438.35 ng/ml




#36 AR-1262-1

R.T.: 8.391 min
Delta R.T.: 0.000 min
Response: 678515
Conc: 258.81 ng/ml



#36 AR-1262-1


R.T.: 7.539 min
Delta R.T.: 0.000 min
Response: 794529
Conc: 919.43 ng/ml



#37 AR-1262-2


R.T.: 8.933 min
Delta R.T.: -0.002 min
Response: 1583615
Conc: 379.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 8.083 min
Delta R.T.: 0.000 min
Response: 1042448
Conc: 379.42 ng/ml




#38 AR-1262-3

R.T.: 9.249 min
Delta R.T.: 0.020 min
Response: 1036671
Conc: 337.87 ng/ml



#38 AR-1262-3


R.T.: 8.364 min
Delta R.T.: 0.000 min
Response: 210130
Conc: 174.22 ng/ml



#39 AR-1262-4


R.T.: 9.298 min
Delta R.T.: -0.020 min
Response: 597874
Conc: 250.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.429 min
Delta R.T.: 0.000 min
Response: 762759
Conc: 357.43 ng/ml




#40 AR-1262-5

R.T.: 9.943 min
Delta R.T.: -0.001 min
Response: 416458
Conc: 271.99 ng/ml



#40 AR-1262-5


R.T.: 8.921 min
Delta R.T.: 0.002 min
Response: 246219
Conc: 240.48 ng/ml



#41 AR-1268-1


R.T.: 9.249 min
Delta R.T.: 0.022 min
Response: 1036671
Conc: 190.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.364 min
Delta R.T.: 0.000 min
Response: 210130
Conc: 67.78 ng/ml




#42 AR-1268-2

R.T.: 9.298 min
Delta R.T.: -0.020 min
Response: 597874
Conc: 117.28 ng/ml



#42 AR-1268-2


R.T.: 8.429 min
Delta R.T.: -0.001 min
Response: 762759
Conc: 272.17 ng/ml



#43 AR-1268-3


R.T.: 9.530 min
Delta R.T.: 0.002 min
Response: 32175
Conc: 6.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.631 min
Delta R.T.: 0.000 min
Response: 19934
Conc: 8.18 ng/ml




#44 AR-1268-4

R.T.: 9.943 min
Delta R.T.: 0.000 min
Response: 416458
Conc: 250.71 ng/ml



#44 AR-1268-4


R.T.: 8.921 min
Delta R.T.: 0.000 min
Response: 246219
Conc: 246.61 ng/ml



#45 AR-1268-5

R.T.: 10.329 min
Delta R.T.: 0.000 min
Response: 147760
Conc: 10.91 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.209 min
Delta R.T.: 0.000 min
Response: 75191
Conc: 11.10 ng/ml