

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050621\
 Data File : PP035594.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 May 2021 9:53
 Operator : DD\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 13:42:32 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 06 11:27:40 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.785	4.203	2081932	1505515	47.585	49.631
2) SA Decachlor...	10.636	9.454	1906686	962475	48.340	49.885
Target Compounds						
16) L4 AR-1242-1	6.090	5.447	523992	331928	458.660	467.713
17) L4 AR-1242-2	6.115	5.467	812136	597915	473.018	480.748
18) L4 AR-1242-3	6.179	5.659	537990	280207	495.745	478.875
19) L4 AR-1242-4	6.285	5.749	428782	282617	489.297	481.736
20) L4 AR-1242-5	7.061	6.308	423064	323999	479.145	479.266



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050621\
Data File : PP035594.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 May 2021 9:53
Operator : DD\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 13:42:32 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050521.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 06 11:27:40 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.785 min
Delta R.T.: -0.002 min
Response: 2081932
Conc: 47.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.203 min
Delta R.T.: 0.000 min
Response: 1505515
Conc: 49.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.636 min
Delta R.T.: -0.007 min
Response: 1906686
Conc: 48.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.454 min
Delta R.T.: -0.003 min
Response: 962475
Conc: 49.89 ng/ml



#16 AR-1242-1


R.T.: 6.090 min
Delta R.T.: -0.003 min
Response: 523992
Conc: 458.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.447 min
Delta R.T.: -0.002 min
Response: 331928
Conc: 467.71 ng/ml




#17 AR-1242-2

R.T.: 6.115 min
Delta R.T.: -0.002 min
Response: 812136
Conc: 473.02 ng/ml



#17 AR-1242-2


R.T.: 5.467 min
Delta R.T.: 0.000 min
Response: 597915
Conc: 480.75 ng/ml



#18 AR-1242-3


R.T.: 6.179 min
Delta R.T.: -0.002 min
Response: 537990
Conc: 495.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.659 min
Delta R.T.: -0.001 min
Response: 280207
Conc: 478.87 ng/ml




#19 AR-1242-4

R.T.: 6.285 min
Delta R.T.: -0.001 min
Response: 428782
Conc: 489.30 ng/ml



#19 AR-1242-4


R.T.: 5.749 min
Delta R.T.: -0.001 min
Response: 282617
Conc: 481.74 ng/ml



#20 AR-1242-5

R.T.: 7.061 min
Delta R.T.: -0.002 min
Response: 423064
Conc: 479.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 6.308 min
Delta R.T.: -0.002 min
Response: 323999
Conc: 479.27 ng/ml