

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050621\
 Data File : PP035611.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 May 2021 16:20
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 17:12:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 06 11:27:40 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.785	4.202	2162353	1595712	49.423	52.604
2) SA Decachlor...	10.637	9.455	1899126	925169	48.148	47.952
Target Compounds						
26) L6 AR-1254-1	6.991	6.307	782093	689459	466.201	485.436
27) L6 AR-1254-2	7.218	6.463	1193141	598581	473.633	475.261
28) L6 AR-1254-3	7.597	6.881	1239619	891631	478.773	469.808
29) L6 AR-1254-4	7.892	7.117	890188	558810	521.074	507.953
30) L6 AR-1254-5	8.317	7.540	967633	780184	467.812	465.355



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050621\
Data File : PP035611.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 May 2021 16:20
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 17:12:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050521.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 06 11:27:40 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.785 min
Delta R.T.: -0.002 min
Response: 2162353
Conc: 49.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.202 min
Delta R.T.: 0.000 min
Response: 1595712
Conc: 52.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.637 min
Delta R.T.: -0.006 min
Response: 1899126
Conc: 48.15 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.455 min
Delta R.T.: -0.002 min
Response: 925169
Conc: 47.95 ng/ml



#26 AR-1254-1


R.T.: 6.991 min
Delta R.T.: -0.003 min
Response: 782093
Conc: 466.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.307 min
Delta R.T.: 0.000 min
Response: 689459
Conc: 485.44 ng/ml




#27 AR-1254-2

R.T.: 7.218 min
Delta R.T.: -0.003 min
Response: 1193141
Conc: 473.63 ng/ml



#27 AR-1254-2


R.T.: 6.463 min
Delta R.T.: -0.002 min
Response: 598581
Conc: 475.26 ng/ml



#28 AR-1254-3


R.T.: 7.597 min
Delta R.T.: -0.003 min
Response: 1239619
Conc: 478.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.881 min
Delta R.T.: -0.002 min
Response: 891631
Conc: 469.81 ng/ml




#29 AR-1254-4

R.T.: 7.892 min
Delta R.T.: -0.002 min
Response: 890188
Conc: 521.07 ng/ml



#29 AR-1254-4


R.T.: 7.117 min
Delta R.T.: -0.002 min
Response: 558810
Conc: 507.95 ng/ml



#30 AR-1254-5

R.T.: 8.317 min
Delta R.T.: -0.003 min
Response: 967633
Conc: 467.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.540 min
Delta R.T.: -0.001 min
Response: 780184
Conc: 465.35 ng/ml