

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050622\
 Data File : PP047125.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2022 18:44
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 21:12:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 06 10:17:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.854	3.121	130.9E6	186.9E6	54.233	56.983
2) SA Decachlor...	9.923	8.429	75434215	130.7E6	51.263	54.774
Target Compounds						
21) L5 AR-1248-1	5.096	4.249	22211238	37446998	531.432	544.954
22) L5 AR-1248-2	5.391	4.499f	29930363	51565692	506.404	551.551
23) L5 AR-1248-3	5.609f	4.541	34663943	53634550	496.253	549.408
24) L5 AR-1248-4	6.047f	4.721f	36587662	61490145	477.139	522.353
25) L5 AR-1248-5	6.087f	5.137	35235451	61585705	453.022	505.378



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050622\
Data File : PP047125.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2022 18:44
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 21:12:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 06 10:17:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.854 min
Delta R.T.: -0.011 min
Response: 130916149
Conc: 54.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.121 min
Delta R.T.: -0.004 min
Response: 186924036
Conc: 56.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.923 min
Delta R.T.: -0.009 min
Response: 75434215
Conc: 51.26 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.429 min
Delta R.T.: -0.008 min
Response: 130652634
Conc: 54.77 ng/ml



#21 AR-1248-1


R.T.: 5.096 min
Delta R.T.: -0.009 min
Response: 22211238
Conc: 531.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.249 min
Delta R.T.: -0.001 min
Response: 37446998
Conc: 544.95 ng/ml




#22 AR-1248-2

R.T.: 5.391 min
Delta R.T.: -0.015 min
Response: 29930363
Conc: 506.40 ng/ml



#22 AR-1248-2


R.T.: 4.499 min
Delta R.T.: -0.016 min
Response: 51565692
Conc: 551.55 ng/ml



#23 AR-1248-3


R.T.: 5.609 min
Delta R.T.: -0.015 min
Response: 34663943
Conc: 496.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.541 min
Delta R.T.: -0.009 min
Response: 53634550
Conc: 549.41 ng/ml




#24 AR-1248-4

R.T.: 6.047 min
Delta R.T.: -0.016 min
Response: 36587662
Conc: 477.14 ng/ml



#24 AR-1248-4


R.T.: 4.721 min
Delta R.T.: -0.016 min
Response: 61490145
Conc: 522.35 ng/ml



#25 AR-1248-5

R.T.: 6.087 min
Delta R.T.: -0.017 min
Response: 35235451
Conc: 453.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.137 min
Delta R.T.: -0.013 min
Response: 61585705
Conc: 505.38 ng/ml