

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050823\
 Data File : PP057629.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 May 2023 14:49
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 08 17:37:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 17:36:20 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.389	3.623	210.6E6	220.9E6	93.735	94.733
2)	SA Decachlor...	10.220	8.687	100.8E6	141.7E6	94.652	91.799
Target Compounds							
26)	L6 AR-1254-1	6.434	5.531	60189583	100.4E6	905.037	915.174
27)	L6 AR-1254-2	6.654	5.680	84602582	88603290	909.663	907.388
28)	L6 AR-1254-3	7.023	6.088	82672586	142.4E6	921.261	922.165
29)	L6 AR-1254-4	7.312	6.319	53395818	83263702	915.617	909.806
30)	L6 AR-1254-5	7.736	6.740	62170523	128.7E6	929.500	909.722



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050823\
Data File : PP057629.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 May 2023 14:49
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 17:37:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 17:36:20 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.389 min
Delta R.T.: 0.001 min
Response: 210556745
Conc: 93.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.623 min
Delta R.T.: 0.000 min
Response: 220907199
Conc: 94.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.220 min
Delta R.T.: 0.001 min
Response: 100782878
Conc: 94.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.687 min
Delta R.T.: 0.000 min
Response: 141673196
Conc: 91.80 ng/ml



#26 AR-1254-1


R.T.: 6.434 min
Delta R.T.: 0.001 min
Response: 60189583
Conc: 905.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#26 AR-1254-1

R.T.: 5.531 min
Delta R.T.: 0.000 min
Response: 100363528
Conc: 915.17 ng/ml




#27 AR-1254-2

R.T.: 6.654 min
Delta R.T.: 0.000 min
Response: 84602582
Conc: 909.66 ng/ml



#27 AR-1254-2


R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 88603290
Conc: 907.39 ng/ml



#28 AR-1254-3


R.T.: 7.023 min
Delta R.T.: 0.000 min
Response: 82672586
Conc: 921.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 6.088 min
Delta R.T.: 0.000 min
Response: 142447860
Conc: 922.17 ng/ml




#29 AR-1254-4

R.T.: 7.312 min
Delta R.T.: 0.000 min
Response: 53395818
Conc: 915.62 ng/ml



#29 AR-1254-4


R.T.: 6.319 min
Delta R.T.: 0.000 min
Response: 83263702
Conc: 909.81 ng/ml



#30 AR-1254-5

R.T.: 7.736 min
Delta R.T.: 0.001 min
Response: 62170523
Conc: 929.50 ng/ml

Instrument :
ECD_P
ClientSampleId :



#30 AR-1254-5

R.T.: 6.740 min
Delta R.T.: 0.000 min
Response: 128747627
Conc: 909.72 ng/ml