

Data Path : P:\HPCHEM1\ECD_P\Data\PP050918\
 Data File : PP025595.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 May 2018 20:19
 Operator : UA/AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 06:00:02 2018
 Quant Method : P:\HPCHEM1\ECD_P\methods\PP043018.M
 Quant Title : 8080.M
 QLast Update : Tue May 01 05:53:49 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.411	11.115	418.8E6	450.6E6	732.047	741.623
Target Compounds							
1) T	Dalapon	2.907	2.454	533.4E6	531.5E6	667.094	659.833
2) T	3,5-DICHL...	9.819	9.242	583.6E6	555.6E6	683.243	690.992
3) T	4-Nitroph...	10.988	10.276	255.4E6	241.4E6	680.242	681.670
5) T	DICAMBA	11.728	11.409	1758.1E6	1615.5E6	706.608	709.130
6) T	MCPD	12.067	11.618	121.3E6	98932757	76.174	68.435
7) T	MCPA	12.323	12.009	164.9E6	143.2E6	69.331	67.678
8) T	DICHLORPROP	12.918	12.520	428.5E6	392.5E6	690.924	699.447
9) T	2,4-D	13.279	13.004	540.4E6	522.9E6	690.593	693.869
10) T	Pentachlo...	13.663	13.552	7049.3E6	6184.7E6	713.192	713.430
11) T	2,4,5-TP ...	14.536	14.156	2971.8E6	2578.0E6	714.132	714.054
12) T	2,4,5-T	14.936	14.780	2740.0E6	2075.2E6	709.808	696.531
13) T	2,4-DB	15.683	15.378	329.1E6	291.2E6	687.446	688.635
14) T	DINOSEB	17.150	15.775	1974.4E6	1638.5E6	730.218	731.601
15) T	Picloram	16.957	17.109	3808.3E6	3097.8E6	646.232	645.290
16) T	DCPA	17.519	16.990	3653.9E6	3064.5E6	731.632	724.130



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD_P\Data\PP050918\
Data File : PP025595.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 May 2018 20:19
Operator : UA/AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 06:00:02 2018
Quant Method : P:\HPCHEM1\ECD_P\methods\PP043018.M
Quant Title : 8080.M
QLast Update : Tue May 01 05:53:49 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.907 min
Delta R.T.: 0.003 min
Response: 533413179
Conc: 667.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#1 Dalapon

R.T.: 2.454 min
Delta R.T.: 0.000 min
Response: 531507221
Conc: 659.83 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 9.819 min
Delta R.T.: -0.002 min
Response: 583631756
Conc: 683.24 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.242 min
Delta R.T.: 0.000 min
Response: 555633433
Conc: 690.99 ng/ml



#3 4-Nitrophenol


R.T.: 10.988 min
Delta R.T.: -0.003 min
Response: 255377239
Conc: 680.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#3 4-Nitrophenol

R.T.: 10.276 min
Delta R.T.: 0.000 min
Response: 241394142
Conc: 681.67 ng/ml




#4 2,4-DCAA

R.T.: 11.411 min
Delta R.T.: -0.004 min
Response: 418827701
Conc: 732.05 ng/ml



#4 2,4-DCAA


R.T.: 11.115 min
Delta R.T.: -0.001 min
Response: 450627819
Conc: 741.62 ng/ml



#5 DICAMBA


R.T.: 11.728 min
Delta R.T.: -0.003 min
Response: 1758087384
Conc: 706.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#5 DICAMBA

R.T.: 11.409 min
Delta R.T.: 0.000 min
Response: 1615474699
Conc: 709.13 ng/ml




#6 MCP

R.T.: 12.067 min
Delta R.T.: 0.002 min
Response: 121318605
Conc: 76.17 ug/ml



#6 MCP


R.T.: 11.618 min
Delta R.T.: 0.007 min
Response: 98932757
Conc: 68.43 ug/ml



#7 MCPA


R.T.: 12.323 min
Delta R.T.: 0.004 min
Response: 164881643
Conc: 69.33 ug/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#7 MCPA

R.T.: 12.009 min
Delta R.T.: 0.008 min
Response: 143175050
Conc: 67.68 ug/ml




#8 DICHLORPROP

R.T.: 12.918 min
Delta R.T.: -0.004 min
Response: 428458599
Conc: 690.92 ng/ml



#8 DICHLORPROP


R.T.: 12.520 min
Delta R.T.: 0.000 min
Response: 392505183
Conc: 699.45 ng/ml



#9 2,4-D


R.T.: 13.279 min
Delta R.T.: -0.005 min
Response: 540423300
Conc: 690.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#9 2,4-D

R.T.: 13.004 min
Delta R.T.: -0.001 min
Response: 522934807
Conc: 693.87 ng/ml




#10 Pentachlorophenol

R.T.: 13.663 min
Delta R.T.: -0.004 min
Response: 7049300436
Conc: 713.19 ng/ml



#10 Pentachlorophenol


R.T.: 13.552 min
Delta R.T.: 0.000 min
Response: 6184738196
Conc: 713.43 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 14.536 min
Delta R.T.: -0.004 min
Response: 2971799765
Conc: 714.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#11 2,4,5-TP (SILVEX)

R.T.: 14.156 min
Delta R.T.: 0.000 min
Response: 2578002682
Conc: 714.05 ng/ml




#12 2,4,5-T

R.T.: 14.936 min
Delta R.T.: -0.005 min
Response: 2740011036
Conc: 709.81 ng/ml



#12 2,4,5-T


R.T.: 14.780 min
Delta R.T.: 0.000 min
Response: 2075218029
Conc: 696.53 ng/ml



#13 2,4-DB


R.T.: 15.683 min
Delta R.T.: -0.004 min
Response: 329116207
Conc: 687.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#13 2,4-DB

R.T.: 15.378 min
Delta R.T.: -0.002 min
Response: 291194997
Conc: 688.63 ng/ml




#14 DINOSEB

R.T.: 17.150 min
Delta R.T.: -0.004 min
Response: 1974430592
Conc: 730.22 ng/ml



#14 DINOSEB


R.T.: 15.775 min
Delta R.T.: -0.001 min
Response: 1638539534
Conc: 731.60 ng/ml



#15 Picloram


R.T.: 16.957 min
Delta R.T.: -0.004 min
Response: 3808251279
Conc: 646.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750




#15 Picloram

R.T.: 17.109 min
Delta R.T.: -0.001 min
Response: 3097751329
Conc: 645.29 ng/ml



#16 DCPA

R.T.: 17.519 min
Delta R.T.: -0.002 min
Response: 3653941171
Conc: 731.63 ng/ml



#16 DCPA

R.T.: 16.990 min
Delta R.T.: 0.000 min
Response: 3064473136
Conc: 724.13 ng/ml