

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050922\
 Data File : PP047295.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2022 23:24
 Operator : YP\AJ
 Sample : N2619-08
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SB04-14-15-FT

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/10/2022
 Supervised By :mohammad ahmed 05/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 06:32:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 06 10:17:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.851	3.115	52367781	64208939	21.694	19.574m
2) SA Decachlor...	9.918	8.424	21578767	39332896	14.664	16.490

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP050922\
Data File : PP047295.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2022 23:24
Operator : YP\AJ
Sample : N2619-08
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
SB04-14-15-FT


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/10/2022
Supervised By :mohammad ahmed 05/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 06:32:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042222.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 06 10:17:17 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.851 min
Delta R.T.: -0.014 min
Response: 52367781
Conc: 21.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
SB04-14-15-FT


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/10/2022
Supervised By :mohammad ahmed 05/10/2022




#1 Tetrachloro-m-xylene

R.T.: 3.115 min
Delta R.T.: -0.009 min
Response: 64208939
Conc: 19.57 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.918 min
Delta R.T.: 0.000 min
Response: 21578767
Conc: 14.66 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.424 min
Delta R.T.: -0.012 min
Response: 39332896
Conc: 16.49 ng/ml