

Data Path : P:\HPCHEM1\ECD P\Data\PP051018\
 Data File : PP025609.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 May 2018 9:53
 Operator : UA/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 14:50:10 2018
 Quant Method : P:\HPCHEM1\ECD_P\methods\PP043018.M
 Quant Title : 8080.M
 QLast Update : Tue May 01 05:53:49 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	nq/ml	nq/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.415	11.116	292.6E6	281.3E6	511.413	462.905
Target Compounds							
6) T	MCPD	12.109	0.000	3881216	0	2.437	N.D. #
11) T	2,4,5-TP ...	14.543	0.000	2715006	0	0.652	N.D. #
12) T	2,4,5-T	14.948	14.785	6027484	8479109	1.561	2.846 #
15) T	Picloram	16.975	17.114	9471557	8704093	1.607	1.813
16) T	DCPA	17.525	0.000	5737678	0	1.149	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD P\Data\PP051018\
Data File : PP025609.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 May 2018 9:53
Operator : UA/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 14:50:10 2018
Quant Method : P:\HPCHEM1\ECD_P\methods\PP043018.M
Quant Title : 8080.M
QLast Update : Tue May 01 05:53:49 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 11.415 min
Delta R.T.: 0.000 min
Response: 292595787
Conc: 511.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#4 2,4-DCAA

R.T.: 11.116 min
Delta R.T.: 0.000 min
Response: 281272413
Conc: 462.91 ng/ml




#6 MCP

R.T.: 12.109 min
Delta R.T.: 0.043 min
Response: 3881216
Conc: 2.44 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 11.611 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 14.543 min
Delta R.T.: 0.003 min
Response: 2715006
Conc: 0.65 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T.: 14.156 min
Response: 0
Conc: N.D.




#12 2,4,5-T

R.T.: 14.948 min
Delta R.T.: 0.007 min
Response: 6027484
Conc: 1.56 ng/ml



#12 2,4,5-T


R.T.: 14.785 min
Delta R.T.: 0.005 min
Response: 8479109
Conc: 2.85 ng/ml



#15 Picloram


R.T.: 16.975 min
Delta R.T.: 0.014 min
Response: 9471557
Conc: 1.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 Picloram

R.T.: 17.114 min
Delta R.T.: 0.004 min
Response: 8704093
Conc: 1.81 ng/ml



#16 DCPA

R.T.: 17.525 min
Delta R.T.: 0.003 min
Response: 5737678
Conc: 1.15 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 16.990 min
Response: 0
Conc: N.D.