

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP051021.M
 Title : GC EXTRACTABLES
 Last Update : Tue May 11 06:10:08 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP035728.D 1000 =PP035726.D 750 =PP035727.D
 250 =PP035729.D 50 =PP035730.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.380	2.191	2.307	2.460	2.683	2.404	E4	7.68
2) SA	Decachlorobiphenyl	1.243	1.126	1.188	1.281	1.272	1.222	E4	5.30
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	1.112	0.990	1.050	1.169	1.531	1.170	E3	18.15
37) L8	AR-1262-2	1.763	1.592	1.685	1.868	1.975	1.777	E3	8.47
38) L8	AR-1262-3	1.224	1.119	1.170	1.291	1.432	1.248	E3	9.74
39) L8	AR-1262-4	0.971	0.880	0.900	0.981	1.125	0.972	E3	9.90
40) L8	AR-1262-5	6.036	5.468	5.688	6.060	6.162	5.883	E2	4.98
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP035728.D 1000 =PP035726.D 750 =PP035727.D
 250 =PP035729.D 50 =PP035730.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.427	3.226	3.377	3.404	3.637	3.414	E4	4.32
2) SA	Decachlorobiphenyl	2.219	2.043	2.132	2.307	2.525	2.245	E4	8.23
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP051021.M
 Title : GC EXTRACTABLES
 Last Update : Tue May 11 06:10:08 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP035728.D 1000 =PP035726.D 750 =PP035727.D
 250 =PP035729.D 50 =PP035730.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	9.428	8.612	9.065	9.975	9.801	9.376 E2	5.89
37) L8	AR-1262-2	3.100	2.898	2.992	3.235	3.483	3.142 E3	7.27
38) L8	AR-1262-3	1.302	1.191	1.243	1.366	1.516	1.323 E3	9.51
39) L8	AR-1262-4	2.370	2.194	2.296	2.458	2.739	2.411 E3	8.60
40) L8	AR-1262-5	1.043	0.970	1.004	1.399	1.533	1.190 E3	21.65
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range