

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP051021.M
 Title : GC EXTRACTABLES
 Last Update : Tue May 11 06:16:03 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP035733.D 1000 =PP035731.D 750 =PP035732.D
 250 =PP035734.D 50 =PP035735.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.307	2.125	2.209	2.439	2.598	2.336 E4	8.04
2) SA	Decachlorobiphenyl	2.014	1.834	1.910	2.145	2.325	2.046 E4	9.53
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	1.945	1.790	1.863	2.091	2.377	2.013 E3	11.52
42) L9	AR-1268-2	1.827	1.675	1.753	1.959	2.169	1.876 E3	10.35
43) L9	AR-1268-3	1.562	1.422	1.474	1.669	1.877	1.601 E3	11.29
44) L9	AR-1268-4	6.568	5.898	6.110	6.806	7.582	6.593 E2	10.01
45) L9	AR-1268-5	4.255	3.953	4.075	4.555	5.069	4.381 E3	10.18

Signal #2 Calibration Files

500 =PP035733.D 1000 =PP035731.D 750 =PP035732.D
 250 =PP035734.D 50 =PP035735.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.315	3.139	3.218	3.389	3.557	3.324 E4	4.86
2) SA	Decachlorobiphenyl	3.603	3.350	3.473	3.782	4.099	3.661 E4	7.99
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP051021.M
 Title : GC EXTRACTABLES
 Last Update : Tue May 11 06:16:03 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP035733.D 1000 =PP035731.D 750 =PP035732.D
 250 =PP035734.D 50 =PP035735.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	3.452	3.203	3.312	3.659	3.933	3.512 E3	8.28
42) L9	AR-1268-2	3.235	3.011	3.099	3.410	3.531	3.257 E3	6.60
43) L9	AR-1268-3	2.657	2.480	2.558	2.814	3.064	2.715 E3	8.53
44) L9	AR-1268-4	1.110	1.027	1.069	1.165	1.451	1.165 E3	14.43
45) L9	AR-1268-5	7.719	7.374	7.570	7.935	8.880	7.896 E3	7.44

(#) = Out of Range