

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051021\
 Data File : PP035737.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 May 2021 00:02
 Operator : DD\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051021AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 05:02:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 11 05:01:24 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.777	4.198	1229312	1768023	50.736	50.401
2)	SA Decachlor...	10.621	9.439	635972	1133565	55.678	51.527
Target Compounds							
16)	L4 AR-1242-1	6.084	5.443	262211	326857	532.762	545.888
17)	L4 AR-1242-2	6.107	5.463	432607	693911	532.635	513.396
18)	L4 AR-1242-3	6.172	5.653	277887	295409	534.124	534.496
19)	L4 AR-1242-4	6.279	5.742	217081	315933	509.622	531.811
20)	L4 AR-1242-5	7.054	6.302	214369	373323	544.984	528.892



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051021\
Data File : PP035737.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 May 2021 00:02
Operator : DD\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051021AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 05:02:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 11 05:01:24 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.777 min
Delta R.T.: 0.000 min
Response: 1229312
Conc: 50.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP051021AR1242




#1 Tetrachloro-m-xylene

R.T.: 4.198 min
Delta R.T.: 0.000 min
Response: 1768023
Conc: 50.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.621 min
Delta R.T.: -0.006 min
Response: 635972
Conc: 55.68 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.439 min
Delta R.T.: -0.002 min
Response: 1133565
Conc: 51.53 ng/ml



#16 AR-1242-1


R.T.: 6.084 min
Delta R.T.: -0.002 min
Response: 262211
Conc: 532.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP051021AR1242




#16 AR-1242-1

R.T.: 5.443 min
Delta R.T.: -0.003 min
Response: 326857
Conc: 545.89 ng/ml




#17 AR-1242-2

R.T.: 6.107 min
Delta R.T.: -0.003 min
Response: 432607
Conc: 532.63 ng/ml



#17 AR-1242-2


R.T.: 5.463 min
Delta R.T.: -0.004 min
Response: 693911
Conc: 513.40 ng/ml



#18 AR-1242-3


R.T.: 6.172 min
Delta R.T.: -0.002 min
Response: 277887
Conc: 534.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP051021AR1242




#18 AR-1242-3

R.T.: 5.653 min
Delta R.T.: 0.000 min
Response: 295409
Conc: 534.50 ng/ml




#19 AR-1242-4

R.T.: 6.279 min
Delta R.T.: -0.003 min
Response: 217081
Conc: 509.62 ng/ml



#19 AR-1242-4


R.T.: 5.742 min
Delta R.T.: 0.000 min
Response: 315933
Conc: 531.81 ng/ml



#20 AR-1242-5

R.T.: 7.054 min
Delta R.T.: -0.004 min
Response: 214369
Conc: 544.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP051021AR1242



#20 AR-1242-5

R.T.: 6.302 min
Delta R.T.: -0.003 min
Response: 373323
Conc: 528.89 ng/ml