

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051022\
 Data File : PP047381.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 May 2022 03:47
 Operator : YP\AJ
 Sample : N2743-07 10X
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SOM-2022-31A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 06:07:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 10 18:34:51 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.850	3.114	5094713	7215148	2.004	2.191
2)	SA Decachlor...	9.914f	8.423	3652773	6995615	2.553	2.570
Target Compounds							
3)	L1 AR-1016-1	5.106	4.265f	82546	293322	1.120	2.514 #
4)	L1 AR-1016-2	5.125	4.265	111610	293322	1.000	1.692 #
5)	L1 AR-1016-3	5.187	4.445	245601	98072	3.581	1.084 #
6)	L1 AR-1016-4	5.292	4.514f	341643	400821	5.848	5.489
7)	L1 AR-1016-5	5.618	4.723	148641	278373	2.723	2.963
8)	L2 AR-1221-1	4.084	3.343	58840	301164	2.066	7.553 #
9)	L2 AR-1221-2	4.173	3.424	183760	782901	8.953	24.413 #
10)	L2 AR-1221-3	4.258	3.495	2199737	498319	34.102	5.441 #
11)	L3 AR-1232-1	4.258	3.495	2199737	498319	36.094	5.653 #
12)	L3 AR-1232-2	4.807	4.265	212386	293322	7.010	3.378 #
13)	L3 AR-1232-3	5.125	4.445	111610	98072	2.017	2.207
14)	L3 AR-1232-4	5.292	4.550	341643	202876	12.747	5.168 #
15)	L3 AR-1232-5	5.384	4.723	122375	278373	6.581	6.292
16)	L4 AR-1242-1	5.088	4.265f	35455	293322	0.616	3.211 #
17)	L4 AR-1242-2	5.125	4.265	111610	293322	1.280	2.164 #
18)	L4 AR-1242-3	5.187	4.445	245601	98072	4.741	1.390 #
19)	L4 AR-1242-4	5.292	4.550	341643	202876	6.879	2.949 #
20)	L4 AR-1242-5	6.073	5.083	405827	177178	8.891	1.987 #
21)	L5 AR-1248-1	5.088	4.265f	35455	293322	0.794	4.053 #
22)	L5 AR-1248-2	5.384	4.514f	122375	400821	1.950	3.970 #
23)	L5 AR-1248-3	5.600	4.550	341256	202876	4.640	1.927 #
24)	L5 AR-1248-4	6.047	4.723	343919	278373	4.210	2.223 #
25)	L5 AR-1248-5	6.073	5.136	405827	125706	5.035	0.983 #
26)	L6 AR-1254-1	6.016	5.083	492580	177178	6.297	0.950 #
27)	L6 AR-1254-2	6.252	0.000	412736	0	3.394	N.D. #
28)	L6 AR-1254-3	6.659	5.687	529047	75224	4.032	0.294 #
29)	L6 AR-1254-4	6.963	5.951f	166550	190481	1.749	1.040 #
30)	L6 AR-1254-5	7.422	0.000	446992	0	4.582	N.D. #
31)	L7 AR-1260-1	6.838	5.818	145795	158823	1.612	0.932 #
32)	L7 AR-1260-2	7.120	6.010f	385712	813993	3.322	3.356
33)	L7 AR-1260-3	7.525	6.174	473849	828262	4.598	4.145
34)	L7 AR-1260-4	7.770	6.694	550343	412414	5.695	2.334 #
35)	L7 AR-1260-5	8.091	6.965	799935	597014	4.472	1.269 #
36)	L8 AR-1262-1	7.507	6.434f	767623	1033695	5.022	3.177 #
37)	L8 AR-1262-2	8.091	6.694	799935	412414	3.748	1.688 #
38)	L8 AR-1262-3	8.404	7.272	345656	1045536	2.449	4.853 #
39)	L8 AR-1262-4	8.513	7.335	372090	975817	3.338	2.702
40)	L8 AR-1262-5	9.173	7.873	92922	2128438	1.305	12.363 #
41)	L9 AR-1268-1	8.404	7.272	345656	1045536	1.326	1.617

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051022\
Data File : PP047381.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 May 2022 03:47
Operator : YP\AJ
Sample : N2743-07 10X
Misc :
ALS Vial : 48 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 06:07:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 10 18:34:51 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.513	7.335	372090	975817	1.614	1.841
43) L9	AR-1268-3	8.744	0.000	210287	0	1.046	N.D. #
44) L9	AR-1268-4	9.173	7.873	92922	2128438	1.159	10.765 #
45) L9	AR-1268-5	9.585	8.172	147518	3483901	0.259	2.800 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051022\
Data File : PP047381.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 May 2022 03:47
Operator : YP\AJ
Sample : N2743-07 10X
Misc :
ALS Vial : 48 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 06:07:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 10 18:34:51 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.850 min
Delta R.T.: -0.010 min
Response: 5094713
Conc: 2.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#1 Tetrachloro-m-xylene

R.T.: 3.114 min
Delta R.T.: -0.001 min
Response: 7215148
Conc: 2.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.914 min
Delta R.T.: -0.019 min
Response: 3652773
Conc: 2.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.423 min
Delta R.T.: -0.004 min
Response: 6995615
Conc: 2.57 ng/ml



#3 AR-1016-1


R.T.: 5.106 min
Delta R.T.: 0.000 min
Response: 82546
Conc: 1.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#3 AR-1016-1

R.T.: 4.265 min
Delta R.T.: 0.018 min
Response: 293322
Conc: 2.51 ng/ml




#4 AR-1016-2

R.T.: 5.125 min
Delta R.T.: -0.003 min
Response: 111610
Conc: 1.00 ng/ml



#4 AR-1016-2


R.T.: 4.265 min
Delta R.T.: 0.000 min
Response: 293322
Conc: 1.69 ng/ml



#5 AR-1016-3


R.T.: 5.187 min
Delta R.T.: -0.007 min
Response: 245601
Conc: 3.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#5 AR-1016-3

R.T.: 4.445 min
Delta R.T.: -0.003 min
Response: 98072
Conc: 1.08 ng/ml




#6 AR-1016-4

R.T.: 5.292 min
Delta R.T.: -0.007 min
Response: 341643
Conc: 5.85 ng/ml



#6 AR-1016-4


R.T.: 4.514 min
Delta R.T.: 0.017 min
Response: 400821
Conc: 5.49 ng/ml



#7 AR-1016-5


R.T.: 5.618 min
Delta R.T.: 0.000 min
Response: 148641
Conc: 2.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#7 AR-1016-5

R.T.: 4.723 min
Delta R.T.: 0.004 min
Response: 278373
Conc: 2.96 ng/ml




#8 AR-1221-1

R.T.: 4.084 min
Delta R.T.: 0.002 min
Response: 58840
Conc: 2.07 ng/ml



#8 AR-1221-1


R.T.: 3.343 min
Delta R.T.: 0.004 min
Response: 301164
Conc: 7.55 ng/ml



#9 AR-1221-2


R.T.: 4.173 min
Delta R.T.: -0.002 min
Response: 183760
Conc: 8.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#9 AR-1221-2

R.T.: 3.424 min
Delta R.T.: -0.005 min
Response: 782901
Conc: 24.41 ng/ml




#10 AR-1221-3

R.T.: 4.258 min
Delta R.T.: 0.004 min
Response: 2199737
Conc: 34.10 ng/ml



#10 AR-1221-3


R.T.: 3.495 min
Delta R.T.: -0.011 min
Response: 498319
Conc: 5.44 ng/ml



#11 AR-1232-1


R.T.: 4.258 min
Delta R.T.: 0.011 min
Response: 2199737
Conc: 36.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#11 AR-1232-1

R.T.: 3.495 min
Delta R.T.: -0.010 min
Response: 498319
Conc: 5.65 ng/ml




#12 AR-1232-2

R.T.: 4.807 min
Delta R.T.: 0.000 min
Response: 212386
Conc: 7.01 ng/ml



#12 AR-1232-2


R.T.: 4.265 min
Delta R.T.: 0.002 min
Response: 293322
Conc: 3.38 ng/ml



#13 AR-1232-3


R.T.: 5.125 min
Delta R.T.: 0.007 min
Response: 111610
Conc: 2.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#13 AR-1232-3

R.T.: 4.445 min
Delta R.T.: -0.001 min
Response: 98072
Conc: 2.21 ng/ml




#14 AR-1232-4

R.T.: 5.292 min
Delta R.T.: 0.003 min
Response: 341643
Conc: 12.75 ng/ml



#14 AR-1232-4


R.T.: 4.550 min
Delta R.T.: 0.013 min
Response: 202876
Conc: 5.17 ng/ml



#15 AR-1232-5


R.T.: 5.384 min
Delta R.T.: -0.005 min
Response: 122375
Conc: 6.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#15 AR-1232-5

R.T.: 4.723 min
Delta R.T.: 0.006 min
Response: 278373
Conc: 6.29 ng/ml




#16 AR-1242-1

R.T.: 5.088 min
Delta R.T.: -0.005 min
Response: 35455
Conc: 0.62 ng/ml



#16 AR-1242-1


R.T.: 4.265 min
Delta R.T.: 0.020 min
Response: 293322
Conc: 3.21 ng/ml



#17 AR-1242-2


R.T.: 5.125 min
Delta R.T.: 0.007 min
Response: 111610
Conc: 1.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#17 AR-1242-2

R.T.: 4.265 min
Delta R.T.: 0.002 min
Response: 293322
Conc: 2.16 ng/ml




#18 AR-1242-3

R.T.: 5.187 min
Delta R.T.: 0.003 min
Response: 245601
Conc: 4.74 ng/ml



#18 AR-1242-3


R.T.: 4.445 min
Delta R.T.: -0.002 min
Response: 98072
Conc: 1.39 ng/ml



#19 AR-1242-4


R.T.: 5.292 min
Delta R.T.: 0.004 min
Response: 341643
Conc: 6.88 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#19 AR-1242-4

R.T.: 4.550 min
Delta R.T.: 0.013 min
Response: 202876
Conc: 2.95 ng/ml




#20 AR-1242-5

R.T.: 6.073 min
Delta R.T.: -0.013 min
Response: 405827
Conc: 8.89 ng/ml



#20 AR-1242-5


R.T.: 5.083 min
Delta R.T.: -0.008 min
Response: 177178
Conc: 1.99 ng/ml



#21 AR-1248-1


R.T.: 5.088 min
Delta R.T.: -0.007 min
Response: 35455
Conc: 0.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#21 AR-1248-1

R.T.: 4.265 min
Delta R.T.: 0.020 min
Response: 293322
Conc: 4.05 ng/ml




#22 AR-1248-2

R.T.: 5.384 min
Delta R.T.: -0.005 min
Response: 122375
Conc: 1.95 ng/ml



#22 AR-1248-2


R.T.: 4.514 min
Delta R.T.: 0.019 min
Response: 400821
Conc: 3.97 ng/ml



#23 AR-1248-3


R.T.: 5.600 min
Delta R.T.: -0.007 min
Response: 341256
Conc: 4.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#23 AR-1248-3

R.T.: 4.550 min
Delta R.T.: 0.013 min
Response: 202876
Conc: 1.93 ng/ml




#24 AR-1248-4

R.T.: 6.047 min
Delta R.T.: 0.001 min
Response: 343919
Conc: 4.21 ng/ml



#24 AR-1248-4


R.T.: 4.723 min
Delta R.T.: 0.006 min
Response: 278373
Conc: 2.22 ng/ml



#25 AR-1248-5


R.T.: 6.073 min
Delta R.T.: -0.013 min
Response: 405827
Conc: 5.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#25 AR-1248-5

R.T.: 5.136 min
Delta R.T.: 0.003 min
Response: 125706
Conc: 0.98 ng/ml




#26 AR-1254-1

R.T.: 6.016 min
Delta R.T.: 0.002 min
Response: 492580
Conc: 6.30 ng/ml



#26 AR-1254-1


R.T.: 5.083 min
Delta R.T.: -0.007 min
Response: 177178
Conc: 0.95 ng/ml



#27 AR-1254-2


R.T.: 6.252 min
Delta R.T.: 0.002 min
Response: 412736
Conc: 3.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 5.249 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 6.659 min
Delta R.T.: 0.007 min
Response: 529047
Conc: 4.03 ng/ml



#28 AR-1254-3


R.T.: 5.687 min
Delta R.T.: 0.009 min
Response: 75224
Conc: 0.29 ng/ml



#29 AR-1254-4


R.T.: 6.963 min
Delta R.T.: 0.002 min
Response: 166550
Conc: 1.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#29 AR-1254-4

R.T.: 5.951 min
Delta R.T.: 0.025 min
Response: 190481
Conc: 1.04 ng/ml




#30 AR-1254-5

R.T.: 7.422 min
Delta R.T.: 0.000 min
Response: 446992
Conc: 4.58 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 6.374 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 6.838 min
Delta R.T.: -0.003 min
Response: 145795
Conc: 1.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#31 AR-1260-1

R.T.: 5.818 min
Delta R.T.: -0.002 min
Response: 158823
Conc: 0.93 ng/ml




#32 AR-1260-2

R.T.: 7.120 min
Delta R.T.: -0.004 min
Response: 385712
Conc: 3.32 ng/ml



#32 AR-1260-2


R.T.: 6.010 min
Delta R.T.: -0.016 min
Response: 813993
Conc: 3.36 ng/ml



#33 AR-1260-3


R.T.: 7.525 min
Delta R.T.: 0.008 min
Response: 473849
Conc: 4.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#33 AR-1260-3

R.T.: 6.174 min
Delta R.T.: -0.012 min
Response: 828262
Conc: 4.14 ng/ml




#34 AR-1260-4

R.T.: 7.770 min
Delta R.T.: 0.006 min
Response: 550343
Conc: 5.70 ng/ml



#34 AR-1260-4


R.T.: 6.694 min
Delta R.T.: -0.002 min
Response: 412414
Conc: 2.33 ng/ml



#35 AR-1260-5


R.T.: 8.091 min
Delta R.T.: -0.013 min
Response: 799935
Conc: 4.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#35 AR-1260-5

R.T.: 6.965 min
Delta R.T.: 0.002 min
Response: 597014
Conc: 1.27 ng/ml




#36 AR-1262-1

R.T.: 7.507 min
Delta R.T.: 0.002 min
Response: 767623
Conc: 5.02 ng/ml



#36 AR-1262-1


R.T.: 6.434 min
Delta R.T.: 0.020 min
Response: 1033695
Conc: 3.18 ng/ml



#37 AR-1262-2


R.T.: 8.091 min
Delta R.T.: 0.000 min
Response: 799935
Conc: 3.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#37 AR-1262-2

R.T.: 6.694 min
Delta R.T.: 0.003 min
Response: 412414
Conc: 1.69 ng/ml




#38 AR-1262-3

R.T.: 8.404 min
Delta R.T.: -0.011 min
Response: 345656
Conc: 2.45 ng/ml



#38 AR-1262-3


R.T.: 7.272 min
Delta R.T.: 0.009 min
Response: 1045536
Conc: 4.85 ng/ml



#39 AR-1262-4


R.T.: 8.513 min
Delta R.T.: 0.006 min
Response: 372090
Conc: 3.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#39 AR-1262-4

R.T.: 7.335 min
Delta R.T.: 0.000 min
Response: 975817
Conc: 2.70 ng/ml




#40 AR-1262-5

R.T.: 9.173 min
Delta R.T.: -0.004 min
Response: 92922
Conc: 1.31 ng/ml



#40 AR-1262-5


R.T.: 7.873 min
Delta R.T.: -0.003 min
Response: 2128438
Conc: 12.36 ng/ml



#41 AR-1268-1


R.T.: 8.404 min
Delta R.T.: -0.008 min
Response: 345656
Conc: 1.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#41 AR-1268-1

R.T.: 7.272 min
Delta R.T.: 0.010 min
Response: 1045536
Conc: 1.62 ng/ml




#42 AR-1268-2

R.T.: 8.513 min
Delta R.T.: 0.004 min
Response: 372090
Conc: 1.61 ng/ml



#42 AR-1268-2


R.T.: 7.335 min
Delta R.T.: 0.001 min
Response: 975817
Conc: 1.84 ng/ml



#43 AR-1268-3


R.T.: 8.744 min
Delta R.T.: 0.006 min
Response: 210287
Conc: 1.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.549 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.173 min
Delta R.T.: 0.000 min
Response: 92922
Conc: 1.16 ng/ml



#44 AR-1268-4


R.T.: 7.873 min
Delta R.T.: -0.002 min
Response: 2128438
Conc: 10.76 ng/ml



#45 AR-1268-5

R.T.: 9.585 min
Delta R.T.: 0.000 min
Response: 147518
Conc: 0.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
SOM-2022-31A



#45 AR-1268-5

R.T.: 8.172 min
Delta R.T.: 0.000 min
Response: 3483901
Conc: 2.80 ng/ml