

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051222\
 Data File : PP047521.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2022 01:28
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 03:39:14 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 11 13:10:20 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.849	3.114	129.1E6	165.6E6	50.792	50.306
2)	SA Decachlor...	9.907	8.416	73880968	133.1E6	51.644	48.895
Target Compounds							
3)	L1 AR-1016-1	5.092	4.242	38173844	58564179	518.158	501.910
4)	L1 AR-1016-2	5.113	4.259	57874584	86882912	518.401	501.234
5)	L1 AR-1016-3	5.180	4.443	34292936	44918066	499.984	496.472
6)	L1 AR-1016-4	5.285	4.492	29305361	36385605	501.617	498.291
7)	L1 AR-1016-5	5.603f	4.714	28090129	47134036	514.673	501.624
31)	L7 AR-1260-1	6.825	5.812	42388685	86656549	468.582	508.683
32)	L7 AR-1260-2	7.107	6.018	56461950	120.3E6	486.327	496.125
33)	L7 AR-1260-3	7.501	6.178	47493472	93415836	460.815	467.446
34)	L7 AR-1260-4	7.747	6.687	45027898	83553777	465.974	472.876
35)	L7 AR-1260-5	8.086	6.955	89688697	222.6E6	501.382	473.129



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051222\
Data File : PP047521.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2022 01:28
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 03:39:14 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 11 13:10:20 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.849 min
Delta R.T.: -0.011 min
Response: 129129761
Conc: 50.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.114 min
Delta R.T.: -0.001 min
Response: 165634286
Conc: 50.31 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.907 min
Delta R.T.: -0.007 min
Response: 73880968
Conc: 51.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.416 min
Delta R.T.: -0.010 min
Response: 133088850
Conc: 48.90 ng/ml



#3 AR-1016-1


R.T.: 5.092 min
Delta R.T.: -0.002 min
Response: 38173844
Conc: 518.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.242 min
Delta R.T.: -0.005 min
Response: 58564179
Conc: 501.91 ng/ml




#4 AR-1016-2

R.T.: 5.113 min
Delta R.T.: -0.014 min
Response: 57874584
Conc: 518.40 ng/ml



#4 AR-1016-2


R.T.: 4.259 min
Delta R.T.: -0.005 min
Response: 86882912
Conc: 501.23 ng/ml



#5 AR-1016-3


R.T.: 5.180 min
Delta R.T.: -0.014 min
Response: 34292936
Conc: 499.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.443 min
Delta R.T.: -0.006 min
Response: 44918066
Conc: 496.47 ng/ml




#6 AR-1016-4

R.T.: 5.285 min
Delta R.T.: -0.014 min
Response: 29305361
Conc: 501.62 ng/ml



#6 AR-1016-4


R.T.: 4.492 min
Delta R.T.: -0.005 min
Response: 36385605
Conc: 498.29 ng/ml



#7 AR-1016-5


R.T.: 5.603 min
Delta R.T.: -0.015 min
Response: 28090129
Conc: 514.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 4.714 min
Delta R.T.: -0.006 min
Response: 47134036
Conc: 501.62 ng/ml




#31 AR-1260-1

R.T.: 6.825 min
Delta R.T.: -0.004 min
Response: 42388685
Conc: 468.58 ng/ml



#31 AR-1260-1


R.T.: 5.812 min
Delta R.T.: -0.008 min
Response: 86656549
Conc: 508.68 ng/ml



#32 AR-1260-2


R.T.: 7.107 min
Delta R.T.: -0.004 min
Response: 56461950
Conc: 486.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.018 min
Delta R.T.: -0.008 min
Response: 120337694
Conc: 496.13 ng/ml




#33 AR-1260-3

R.T.: 7.501 min
Delta R.T.: -0.003 min
Response: 47493472
Conc: 460.82 ng/ml



#33 AR-1260-3


R.T.: 6.178 min
Delta R.T.: -0.008 min
Response: 93415836
Conc: 467.45 ng/ml



#34 AR-1260-4


R.T.: 7.747 min
Delta R.T.: -0.005 min
Response: 45027898
Conc: 465.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.687 min
Delta R.T.: -0.009 min
Response: 83553777
Conc: 472.88 ng/ml



#35 AR-1260-5

R.T.: 8.086 min
Delta R.T.: -0.005 min
Response: 89688697
Conc: 501.38 ng/ml



#35 AR-1260-5

R.T.: 6.955 min
Delta R.T.: -0.009 min
Response: 222563509
Conc: 473.13 ng/ml