

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051321\
 Data File : PP035844.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 May 2021 18:05
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 18:51:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 11 06:42:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.775	4.195	1087206	1625443	46.401	47.817
2)	SA Decachlor...	10.613	9.434	621691	1050949	55.233	49.696
Target Compounds							
3)	L1 AR-1016-1	6.079	5.436	342891	457719	534.045	570.619
4)	L1 AR-1016-2	6.102	5.457	541461	885811	523.908	511.257
5)	L1 AR-1016-3	6.167	5.648	358060	370475	525.216	512.118
6)	L1 AR-1016-4	6.272	5.692	288001	307753	524.764	472.220
7)	L1 AR-1016-5	6.583	5.922	245944	399640	484.280	507.648
8)	L2 AR-1221-1	5.019	4.448	40969	58421	149.626	152.898
9)	L2 AR-1221-2	5.120	4.549	60336	82059	268.859	294.034
10)	L2 AR-1221-3	5.205	4.637	210505	322784	339.792	361.867
11)	L3 AR-1232-1	5.205	4.637	210505	322784	440.851	471.905
12)	L3 AR-1232-2	5.785	5.457	295453	885811	1136.684	1238.794
13)	L3 AR-1232-3	6.102	5.648	541461	370475	1207.031	1338.454
14)	L3 AR-1232-4	6.272	5.738	288001	391418	1258.675	1423.290
15)	L3 AR-1232-5	6.368	5.922	202758	399640	1426.755	1378.779
16)	L4 AR-1242-1	6.079	5.436	342891	457719	696.689	764.444
17)	L4 AR-1242-2	6.102	5.457	541461	885811	666.658	655.375
18)	L4 AR-1242-3	6.167	5.648	358060	370475	688.222	670.315
19)	L4 AR-1242-4	6.272	5.738	288001	391418	676.113	658.874
20)	L4 AR-1242-5	7.049	6.294	34014	318713	86.474	451.524 #
21)	L5 AR-1248-1	6.079	5.436	342891	457719	913.051	991.564
22)	L5 AR-1248-2	6.368	5.692	202758	307753	374.678	353.304
23)	L5 AR-1248-3	6.583	5.738	245944	391418	359.944	435.511
24)	L5 AR-1248-4	7.008	5.922	44204	399640	65.828	384.562 #
25)	L5 AR-1248-5	7.049	6.341	34014	51697	48.996	45.113
26)	L6 AR-1254-1	6.977	6.294	181096	318713	258.570	195.450
27)	L6 AR-1254-2	7.205	6.450	202008	303615	183.553	220.830
28)	L6 AR-1254-3	7.585	6.880	112393	530756	94.980	255.719 #
29)	L6 AR-1254-4	7.876	7.106	72682	46769	91.161	32.179 #
30)	L6 AR-1254-5	8.302	7.526	643265	968230	681.713	504.645 #
31)	L7 AR-1260-1	7.748	6.999	450379	720304	499.555	502.410
32)	L7 AR-1260-2	8.013	7.193	511935	873261	525.530	505.887
33)	L7 AR-1260-3	8.376	7.347	337668	782409	504.627	484.900
34)	L7 AR-1260-4	8.605	7.823	433056	586596	552.617	508.490
35)	L7 AR-1260-5	8.919	8.068	754477	1322073	572.014	515.251
36)	L8 AR-1262-1	8.376	7.526	337668	968230	303.524	1026.999 #
37)	L8 AR-1262-2	8.919	8.068	754477	1322073	427.950	426.446
38)	L8 AR-1262-3	9.233f	8.348	494393	269737	403.782	207.201 #
39)	L8 AR-1262-4	9.283f	8.413	282774	982755	291.291	414.635 #
40)	L8 AR-1262-5	9.924	8.905	173920	307989	288.116	295.167
41)	L9 AR-1268-1	9.233f	8.348	494393	269737	245.576	76.808 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051321\
 Data File : PP035844.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 May 2021 18:05
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 18:51:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 11 06:42:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.283f	8.413	282774	982755	150.701	301.711 #
43) L9	AR-1268-3	9.512	8.613	21338	23032	13.330	8.484 #
44) L9	AR-1268-4	9.924	8.905	173920	307989	263.805	264.481
45) L9	AR-1268-5	10.306	9.190	75783	139366	17.297	17.651



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051321\
Data File : PP035844.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 May 2021 18:05
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 18:51:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051021.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 11 06:42:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.775 min
Delta R.T.: -0.004 min
Response: 1087206
Conc: 46.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.195 min
Delta R.T.: -0.002 min
Response: 1625443
Conc: 47.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.613 min
Delta R.T.: -0.012 min
Response: 621691
Conc: 55.23 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.434 min
Delta R.T.: -0.008 min
Response: 1050949
Conc: 49.70 ng/ml



#3 AR-1016-1


R.T.: 6.079 min
Delta R.T.: -0.006 min
Response: 342891
Conc: 534.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.436 min
Delta R.T.: -0.007 min
Response: 457719
Conc: 570.62 ng/ml




#4 AR-1016-2

R.T.: 6.102 min
Delta R.T.: -0.007 min
Response: 541461
Conc: 523.91 ng/ml



#4 AR-1016-2


R.T.: 5.457 min
Delta R.T.: -0.007 min
Response: 885811
Conc: 511.26 ng/ml



#5 AR-1016-3


R.T.: 6.167 min
Delta R.T.: -0.007 min
Response: 358060
Conc: 525.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.648 min
Delta R.T.: -0.006 min
Response: 370475
Conc: 512.12 ng/ml




#6 AR-1016-4

R.T.: 6.272 min
Delta R.T.: -0.008 min
Response: 288001
Conc: 524.76 ng/ml



#6 AR-1016-4


R.T.: 5.692 min
Delta R.T.: -0.006 min
Response: 307753
Conc: 472.22 ng/ml



#7 AR-1016-5


R.T.: 6.583 min
Delta R.T.: -0.006 min
Response: 245944
Conc: 484.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.922 min
Delta R.T.: -0.006 min
Response: 399640
Conc: 507.65 ng/ml




#8 AR-1221-1

R.T.: 5.019 min
Delta R.T.: -0.009 min
Response: 40969
Conc: 149.63 ng/ml



#8 AR-1221-1


R.T.: 4.448 min
Delta R.T.: -0.006 min
Response: 58421
Conc: 152.90 ng/ml



#9 AR-1221-2


R.T.: 5.120 min
Delta R.T.: -0.005 min
Response: 60336
Conc: 268.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.549 min
Delta R.T.: -0.005 min
Response: 82059
Conc: 294.03 ng/ml




#10 AR-1221-3

R.T.: 5.205 min
Delta R.T.: -0.006 min
Response: 210505
Conc: 339.79 ng/ml



#10 AR-1221-3


R.T.: 4.637 min
Delta R.T.: -0.005 min
Response: 322784
Conc: 361.87 ng/ml



#11 AR-1232-1


R.T.: 5.205 min
Delta R.T.: -0.005 min
Response: 210505
Conc: 440.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.637 min
Delta R.T.: -0.005 min
Response: 322784
Conc: 471.91 ng/ml




#12 AR-1232-2

R.T.: 5.785 min
Delta R.T.: -0.005 min
Response: 295453
Conc: 1136.68 ng/ml



#12 AR-1232-2


R.T.: 5.457 min
Delta R.T.: -0.008 min
Response: 885811
Conc: 1238.79 ng/ml



#13 AR-1232-3


R.T.: 6.102 min
Delta R.T.: -0.008 min
Response: 541461
Conc: 1207.03 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.648 min
Delta R.T.: -0.007 min
Response: 370475
Conc: 1338.45 ng/ml




#14 AR-1232-4

R.T.: 6.272 min
Delta R.T.: -0.007 min
Response: 288001
Conc: 1258.68 ng/ml



#14 AR-1232-4


R.T.: 5.738 min
Delta R.T.: -0.006 min
Response: 391418
Conc: 1423.29 ng/ml



#15 AR-1232-5


R.T.: 6.368 min
Delta R.T.: -0.006 min
Response: 202758
Conc: 1426.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.922 min
Delta R.T.: -0.006 min
Response: 399640
Conc: 1378.78 ng/ml




#16 AR-1242-1

R.T.: 6.079 min
Delta R.T.: -0.006 min
Response: 342891
Conc: 696.69 ng/ml



#16 AR-1242-1


R.T.: 5.436 min
Delta R.T.: -0.006 min
Response: 457719
Conc: 764.44 ng/ml



#17 AR-1242-2


R.T.: 6.102 min
Delta R.T.: -0.007 min
Response: 541461
Conc: 666.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.457 min
Delta R.T.: -0.005 min
Response: 885811
Conc: 655.37 ng/ml




#18 AR-1242-3

R.T.: 6.167 min
Delta R.T.: -0.007 min
Response: 358060
Conc: 688.22 ng/ml



#18 AR-1242-3


R.T.: 5.648 min
Delta R.T.: -0.005 min
Response: 370475
Conc: 670.31 ng/ml



#19 AR-1242-4


R.T.: 6.272 min
Delta R.T.: -0.008 min
Response: 288001
Conc: 676.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.738 min
Delta R.T.: -0.005 min
Response: 391418
Conc: 658.87 ng/ml




#20 AR-1242-5

R.T.: 7.049 min
Delta R.T.: -0.007 min
Response: 34014
Conc: 86.47 ng/ml



#20 AR-1242-5


R.T.: 6.294 min
Delta R.T.: -0.009 min
Response: 318713
Conc: 451.52 ng/ml



#21 AR-1248-1


R.T.: 6.079 min
Delta R.T.: -0.007 min
Response: 342891
Conc: 913.05 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.436 min
Delta R.T.: -0.009 min
Response: 457719
Conc: 991.56 ng/ml




#22 AR-1248-2

R.T.: 6.368 min
Delta R.T.: -0.007 min
Response: 202758
Conc: 374.68 ng/ml



#22 AR-1248-2


R.T.: 5.692 min
Delta R.T.: -0.005 min
Response: 307753
Conc: 353.30 ng/ml



#23 AR-1248-3


R.T.: 6.583 min
Delta R.T.: -0.007 min
Response: 245944
Conc: 359.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.738 min
Delta R.T.: -0.005 min
Response: 391418
Conc: 435.51 ng/ml




#24 AR-1248-4

R.T.: 7.008 min
Delta R.T.: -0.007 min
Response: 44204
Conc: 65.83 ng/ml



#24 AR-1248-4


R.T.: 5.922 min
Delta R.T.: -0.005 min
Response: 399640
Conc: 384.56 ng/ml



#25 AR-1248-5


R.T.: 7.049 min
Delta R.T.: -0.008 min
Response: 34014
Conc: 49.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.341 min
Delta R.T.: -0.008 min
Response: 51697
Conc: 45.11 ng/ml




#26 AR-1254-1

R.T.: 6.977 min
Delta R.T.: -0.007 min
Response: 181096
Conc: 258.57 ng/ml



#26 AR-1254-1


R.T.: 6.294 min
Delta R.T.: -0.006 min
Response: 318713
Conc: 195.45 ng/ml



#27 AR-1254-2


R.T.: 7.205 min
Delta R.T.: -0.007 min
Response: 202008
Conc: 183.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.450 min
Delta R.T.: -0.006 min
Response: 303615
Conc: 220.83 ng/ml




#28 AR-1254-3

R.T.: 7.585 min
Delta R.T.: -0.007 min
Response: 112393
Conc: 94.98 ng/ml



#28 AR-1254-3


R.T.: 6.880 min
Delta R.T.: 0.006 min
Response: 530756
Conc: 255.72 ng/ml



#29 AR-1254-4


R.T.: 7.876 min
Delta R.T.: -0.010 min
Response: 72682
Conc: 91.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 7.106 min
Delta R.T.: -0.007 min
Response: 46769
Conc: 32.18 ng/ml




#30 AR-1254-5

R.T.: 8.302 min
Delta R.T.: -0.008 min
Response: 643265
Conc: 681.71 ng/ml



#30 AR-1254-5


R.T.: 7.526 min
Delta R.T.: -0.006 min
Response: 968230
Conc: 504.64 ng/ml



#31 AR-1260-1


R.T.: 7.748 min
Delta R.T.: -0.007 min
Response: 450379
Conc: 499.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.999 min
Delta R.T.: -0.006 min
Response: 720304
Conc: 502.41 ng/ml




#32 AR-1260-2

R.T.: 8.013 min
Delta R.T.: -0.006 min
Response: 511935
Conc: 525.53 ng/ml



#32 AR-1260-2


R.T.: 7.193 min
Delta R.T.: -0.007 min
Response: 873261
Conc: 505.89 ng/ml



#33 AR-1260-3


R.T.: 8.376 min
Delta R.T.: -0.005 min
Response: 337668
Conc: 504.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.347 min
Delta R.T.: -0.007 min
Response: 782409
Conc: 484.90 ng/ml




#34 AR-1260-4

R.T.: 8.605 min
Delta R.T.: -0.006 min
Response: 433056
Conc: 552.62 ng/ml



#34 AR-1260-4


R.T.: 7.823 min
Delta R.T.: -0.007 min
Response: 586596
Conc: 508.49 ng/ml



#35 AR-1260-5


R.T.: 8.919 min
Delta R.T.: -0.007 min
Response: 754477
Conc: 572.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 8.068 min
Delta R.T.: -0.007 min
Response: 1322073
Conc: 515.25 ng/ml




#36 AR-1262-1

R.T.: 8.376 min
Delta R.T.: -0.005 min
Response: 337668
Conc: 303.52 ng/ml



#36 AR-1262-1


R.T.: 7.526 min
Delta R.T.: -0.005 min
Response: 968230
Conc: 1027.00 ng/ml



#37 AR-1262-2


R.T.: 8.919 min
Delta R.T.: -0.005 min
Response: 754477
Conc: 427.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 8.068 min
Delta R.T.: -0.005 min
Response: 1322073
Conc: 426.45 ng/ml




#38 AR-1262-3

R.T.: 9.233 min
Delta R.T.: 0.016 min
Response: 494393
Conc: 403.78 ng/ml



#38 AR-1262-3


R.T.: 8.348 min
Delta R.T.: -0.005 min
Response: 269737
Conc: 207.20 ng/ml



#39 AR-1262-4


R.T.: 9.283 min
Delta R.T.: -0.023 min
Response: 282774
Conc: 291.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.413 min
Delta R.T.: -0.005 min
Response: 982755
Conc: 414.63 ng/ml




#40 AR-1262-5

R.T.: 9.924 min
Delta R.T.: -0.007 min
Response: 173920
Conc: 288.12 ng/ml



#40 AR-1262-5


R.T.: 8.905 min
Delta R.T.: -0.006 min
Response: 307989
Conc: 295.17 ng/ml



#41 AR-1268-1


R.T.: 9.233 min
Delta R.T.: 0.017 min
Response: 494393
Conc: 245.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.348 min
Delta R.T.: -0.005 min
Response: 269737
Conc: 76.81 ng/ml




#42 AR-1268-2

R.T.: 9.283 min
Delta R.T.: -0.025 min
Response: 282774
Conc: 150.70 ng/ml



#42 AR-1268-2


R.T.: 8.413 min
Delta R.T.: -0.006 min
Response: 982755
Conc: 301.71 ng/ml



#43 AR-1268-3


R.T.: 9.512 min
Delta R.T.: -0.004 min
Response: 21338
Conc: 13.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.613 min
Delta R.T.: -0.006 min
Response: 23032
Conc: 8.48 ng/ml




#44 AR-1268-4

R.T.: 9.924 min
Delta R.T.: -0.007 min
Response: 173920
Conc: 263.81 ng/ml



#44 AR-1268-4


R.T.: 8.905 min
Delta R.T.: -0.006 min
Response: 307989
Conc: 264.48 ng/ml



#45 AR-1268-5

R.T.: 10.306 min
Delta R.T.: -0.010 min
Response: 75783
Conc: 17.30 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.190 min
Delta R.T.: -0.005 min
Response: 139366
Conc: 17.65 ng/ml