

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051322\
 Data File : PP047544.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2022 19:20
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 14 06:58:34 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 11 13:10:20 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.848	3.114	130.3E6	170.9E6	51.242	51.908
2)	SA Decachlor...	9.907	8.416	73820725	148.6E6	51.602	54.585
Target Compounds							
26)	L6 AR-1254-1	6.010	5.085	38917100	91995844	497.494	493.206
27)	L6 AR-1254-2	6.247	5.245	62404876	80597702	513.159	489.079
28)	L6 AR-1254-3	6.647	5.673	62343039	126.8E6	475.186	494.809
29)	L6 AR-1254-4	6.960	5.922	50114141	90878286	526.310	496.307
30)	L6 AR-1254-5	7.417	6.368	50419854	124.0E6	516.849	516.373



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051322\
Data File : PP047544.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2022 19:20
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 14 06:58:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 11 13:10:20 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.848 min
Delta R.T.: -0.012 min
Response: 130274821
Conc: 51.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.114 min
Delta R.T.: -0.002 min
Response: 170909809
Conc: 51.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.907 min
Delta R.T.: -0.007 min
Response: 73820725
Conc: 51.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.416 min
Delta R.T.: -0.011 min
Response: 148576804
Conc: 54.59 ng/ml



#26 AR-1254-1


R.T.: 6.010 min
Delta R.T.: -0.004 min
Response: 38917100
Conc: 497.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.085 min
Delta R.T.: -0.005 min
Response: 91995844
Conc: 493.21 ng/ml




#27 AR-1254-2

R.T.: 6.247 min
Delta R.T.: -0.003 min
Response: 62404876
Conc: 513.16 ng/ml



#27 AR-1254-2


R.T.: 5.245 min
Delta R.T.: -0.004 min
Response: 80597702
Conc: 489.08 ng/ml



#28 AR-1254-3


R.T.: 6.647 min
Delta R.T.: -0.005 min
Response: 62343039
Conc: 475.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.673 min
Delta R.T.: -0.005 min
Response: 126753971
Conc: 494.81 ng/ml




#29 AR-1254-4

R.T.: 6.960 min
Delta R.T.: -0.002 min
Response: 50114141
Conc: 526.31 ng/ml



#29 AR-1254-4


R.T.: 5.922 min
Delta R.T.: -0.005 min
Response: 90878286
Conc: 496.31 ng/ml



#30 AR-1254-5

R.T.: 7.417 min
Delta R.T.: -0.005 min
Response: 50419854
Conc: 516.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.368 min
Delta R.T.: -0.005 min
Response: 123970518
Conc: 516.37 ng/ml