

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051324\
 Data File : PP064854.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2024 16:55
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 17:26:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 13 17:18:12 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.260	3.372	49827234	58722943	25.773	25.654
2) SA Decachlor...	9.911	8.268	19822405	46276863	27.552	27.024
Target Compounds						
3) L1 AR-1016-1	5.430	4.432	11542532	20456548	277.358	267.817
4) L1 AR-1016-2	5.451	4.449	17009750	28546980	272.143	264.421
5) L1 AR-1016-3	5.513	4.621	10890903	16084350	273.496	269.061
6) L1 AR-1016-4	5.612	4.665	8077584	13426837	261.187	276.836
7) L1 AR-1016-5	5.908	4.873	7684833	17207070	255.592	273.318
31) L7 AR-1260-1	7.036	5.892	13318729	32051004	261.676	269.750
32) L7 AR-1260-2	7.293	6.079	13743238	35853105	270.657	268.316
33) L7 AR-1260-3	7.653	6.229	10731978	35021199	273.626	266.457
34) L7 AR-1260-4	7.877	6.696	12130392	28329334	273.202	268.632
35) L7 AR-1260-5	8.187	6.938	19305267	60405287	264.822	261.588



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051324\
Data File : PP064854.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2024 16:55
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 17:26:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051324.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 13 17:18:12 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.260 min
Delta R.T.: 0.000 min
Response: 49827234
Conc: 25.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.372 min
Delta R.T.: 0.000 min
Response: 58722943
Conc: 25.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.911 min
Delta R.T.: 0.000 min
Response: 19822405
Conc: 27.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.268 min
Delta R.T.: 0.000 min
Response: 46276863
Conc: 27.02 ng/ml



#3 AR-1016-1


R.T.: 5.430 min
Delta R.T.: 0.000 min
Response: 11542532
Conc: 277.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#3 AR-1016-1

R.T.: 4.432 min
Delta R.T.: 0.000 min
Response: 20456548
Conc: 267.82 ng/ml




#4 AR-1016-2

R.T.: 5.451 min
Delta R.T.: 0.000 min
Response: 17009750
Conc: 272.14 ng/ml



#4 AR-1016-2


R.T.: 4.449 min
Delta R.T.: 0.000 min
Response: 28546980
Conc: 264.42 ng/ml



#5 AR-1016-3


R.T.: 5.513 min
Delta R.T.: 0.000 min
Response: 10890903
Conc: 273.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#5 AR-1016-3

R.T.: 4.621 min
Delta R.T.: 0.000 min
Response: 16084350
Conc: 269.06 ng/ml




#6 AR-1016-4

R.T.: 5.612 min
Delta R.T.: 0.000 min
Response: 8077584
Conc: 261.19 ng/ml



#6 AR-1016-4


R.T.: 4.665 min
Delta R.T.: 0.000 min
Response: 13426837
Conc: 276.84 ng/ml



#7 AR-1016-5


R.T.: 5.908 min
Delta R.T.: 0.000 min
Response: 7684833
Conc: 255.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#7 AR-1016-5

R.T.: 4.873 min
Delta R.T.: 0.000 min
Response: 17207070
Conc: 273.32 ng/ml




#31 AR-1260-1

R.T.: 7.036 min
Delta R.T.: 0.000 min
Response: 13318729
Conc: 261.68 ng/ml



#31 AR-1260-1


R.T.: 5.892 min
Delta R.T.: 0.000 min
Response: 32051004
Conc: 269.75 ng/ml



#32 AR-1260-2


R.T.: 7.293 min
Delta R.T.: 0.000 min
Response: 13743238
Conc: 270.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#32 AR-1260-2

R.T.: 6.079 min
Delta R.T.: 0.000 min
Response: 35853105
Conc: 268.32 ng/ml




#33 AR-1260-3

R.T.: 7.653 min
Delta R.T.: 0.000 min
Response: 10731978
Conc: 273.63 ng/ml



#33 AR-1260-3


R.T.: 6.229 min
Delta R.T.: 0.000 min
Response: 35021199
Conc: 266.46 ng/ml



#34 AR-1260-4


R.T.: 7.877 min
Delta R.T.: 0.000 min
Response: 12130392
Conc: 273.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#34 AR-1260-4

R.T.: 6.696 min
Delta R.T.: 0.000 min
Response: 28329334
Conc: 268.63 ng/ml



#35 AR-1260-5

R.T.: 8.187 min
Delta R.T.: 0.000 min
Response: 19305267
Conc: 264.82 ng/ml



#35 AR-1260-5

R.T.: 6.938 min
Delta R.T.: 0.000 min
Response: 60405287
Conc: 261.59 ng/ml