

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051324\
 Data File : PP064864.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2024 19:38
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 14 01:52:59 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 14 01:50:21 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.257	3.371	149.5E6	195.1E6	74.495	74.823
2)	SA Decachlor...	9.910	8.266	55006853	131.2E6	72.110	71.823
Target Compounds							
21)	L5 AR-1248-1	5.424	4.427	20121132	37266734	783.931	747.735
22)	L5 AR-1248-2	5.699	4.660	30041293	52064072	788.179	717.545
23)	L5 AR-1248-3	5.903	4.700	31114556	55246125	762.215	717.119
24)	L5 AR-1248-4	6.308	4.867	29035703	64224319	735.823	722.133
25)	L5 AR-1248-5	6.346	5.253	29590724	60344254	738.835	729.264



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051324\
Data File : PP064864.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2024 19:38
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 14 01:52:59 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051324.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 14 01:50:21 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.257 min
Delta R.T.: 0.000 min
Response: 149531772
Conc: 74.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.371 min
Delta R.T.: 0.000 min
Response: 195052467
Conc: 74.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.910 min
Delta R.T.: 0.001 min
Response: 55006853
Conc: 72.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.266 min
Delta R.T.: 0.000 min
Response: 131168964
Conc: 71.82 ng/ml



#21 AR-1248-1


R.T.: 5.424 min
Delta R.T.: 0.000 min
Response: 20121132
Conc: 783.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 4.427 min
Delta R.T.: -0.001 min
Response: 37266734
Conc: 747.74 ng/ml




#22 AR-1248-2

R.T.: 5.699 min
Delta R.T.: -0.002 min
Response: 30041293
Conc: 788.18 ng/ml



#22 AR-1248-2


R.T.: 4.660 min
Delta R.T.: -0.001 min
Response: 52064072
Conc: 717.55 ng/ml



#23 AR-1248-3


R.T.: 5.903 min
Delta R.T.: 0.000 min
Response: 31114556
Conc: 762.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 4.700 min
Delta R.T.: -0.001 min
Response: 55246125
Conc: 717.12 ng/ml




#24 AR-1248-4

R.T.: 6.308 min
Delta R.T.: 0.000 min
Response: 29035703
Conc: 735.82 ng/ml



#24 AR-1248-4


R.T.: 4.867 min
Delta R.T.: -0.002 min
Response: 64224319
Conc: 722.13 ng/ml



#25 AR-1248-5

R.T.: 6.346 min
Delta R.T.: 0.000 min
Response: 29590724
Conc: 738.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 5.253 min
Delta R.T.: -0.001 min
Response: 60344254
Conc: 729.26 ng/ml