

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051324\
 Data File : PP064860.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2024 18:32
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 23:21:06 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 13 23:18:57 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.256	3.371	101.4E6	132.2E6	50.000	50.000
2) SA Decachlor...	9.910	8.266	37632172	90064176	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.424	4.428	18251330	33918903	500.000	500.000
17) L4 AR-1242-2	5.445	4.445	27556542	46936456	500.000	500.000
18) L4 AR-1242-3	5.508	4.617	17755643	26530192	500.000	500.000
19) L4 AR-1242-4	5.607	4.700	14500793	26905389	500.000	500.000
20) L4 AR-1242-5	6.346	5.214	12115984	31090461	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051324\
Data File : PP064860.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2024 18:32
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 23:21:06 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051324.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 13 23:18:57 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.256 min
Delta R.T.: 0.000 min
Response: 101447611
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.371 min
Delta R.T.: 0.000 min
Response: 132200527
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.910 min
Delta R.T.: 0.000 min
Response: 37632172
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.266 min
Delta R.T.: 0.000 min
Response: 90064176
Conc: 50.00 ng/ml



#16 AR-1242-1


R.T.: 5.424 min
Delta R.T.: 0.000 min
Response: 18251330
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#16 AR-1242-1

R.T.: 4.428 min
Delta R.T.: 0.000 min
Response: 33918903
Conc: 500.00 ng/ml




#17 AR-1242-2

R.T.: 5.445 min
Delta R.T.: 0.000 min
Response: 27556542
Conc: 500.00 ng/ml



#17 AR-1242-2


R.T.: 4.445 min
Delta R.T.: 0.000 min
Response: 46936456
Conc: 500.00 ng/ml



#18 AR-1242-3


R.T.: 5.508 min
Delta R.T.: 0.000 min
Response: 17755643
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#18 AR-1242-3

R.T.: 4.617 min
Delta R.T.: 0.000 min
Response: 26530192
Conc: 500.00 ng/ml




#19 AR-1242-4

R.T.: 5.607 min
Delta R.T.: 0.000 min
Response: 14500793
Conc: 500.00 ng/ml



#19 AR-1242-4


R.T.: 4.700 min
Delta R.T.: 0.000 min
Response: 26905389
Conc: 500.00 ng/ml



#20 AR-1242-5

R.T.: 6.346 min
Delta R.T.: 0.000 min
Response: 12115984
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500



#20 AR-1242-5

R.T.: 5.214 min
Delta R.T.: 0.000 min
Response: 31090461
Conc: 500.00 ng/ml