

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051420\
 Data File : PP030634.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 May 2020 22:47
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 02:39:08 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051420.M
 Quant Title : 8080.M
 QLast Update : Fri May 15 02:35:56 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 T	Dalapon	682.500	692.053	-1.4	102	0.00
2 T	3,5-DICHLORO BENZOIC ACID	697.500	705.597	-1.2	102	0.00
3 T	4-Nitrophenol	682.500	701.956	-2.9	103	0.00
4 S	2,4-DCAA	750.000	768.706	-2.5	104	0.00
5 T	DICAMBA	705.000	718.844	-2.0	103	0.00
6 T	MCP P	70.500	67.550	4.2	102	0.00
7 T	MCPA	69.750	66.056	5.3	102	0.00
8 T	DICHLORPROP	705.000	718.934	-2.0	104	0.00
9 T	2,4-D	705.000	705.039	-0.0	104	0.00
10 T	Pentachlorophenol	712.500	738.650	-3.7	104	0.00
11 T	2,4,5-TP (SILVEX)	712.500	740.435	-3.9	105	0.00
12 T	2,4,5-T	712.500	745.605	-4.6	105	0.00
13 T	2,4-DB	712.500	749.845	-5.2	105	0.00
14 T	DINOSEB	705.000	727.553	-3.2	105	0.00
15 T	Picloram	712.500	749.925	-5.3	106	0.00
16 T	DCPA	720.000	753.173	-4.6	105	0.00

Signal #2

1 T	Dalapon	682.500	695.916	-2.0	102	0.00
2 T	3,5-DICHLORO BENZOIC ACID	697.500	708.129	-1.5	103	0.00
3 T	4-Nitrophenol	682.500	739.146	-8.3	105	0.00
4 S	2,4-DCAA	750.000	772.835	-3.0	104	0.00
5 T	DICAMBA	705.000	723.766	-2.7	103	0.00
6 T	MCP P	70.500	75.542	-7.2	103	0.00
7 T	MCPA	69.750	72.021	-3.3	103	0.00
8 T	DICHLORPROP	705.000	721.156	-2.3	104	0.00
9 T	2,4-D	705.000	743.728	-5.5	105	0.00
10 T	Pentachlorophenol	712.500	736.657	-3.4	103	0.00
11 T	2,4,5-TP (SILVEX)	712.500	741.340	-4.0	104	0.00
12 T	2,4,5-T	712.500	765.172	-7.4	106	0.00
13 T	2,4-DB	712.500	753.357	-5.7	105	0.00
14 T	DINOSEB	705.000	726.372	-3.0	104	0.00
15 T	Picloram	712.500	769.733	-8.0	107	0.00
16 T	DCPA	720.000	747.346	-3.8	104	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051420\
Data File : PP030634.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 May 2020 22:47
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 02:39:08 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051420.M
Quant Title : 8080.M
QLast Update : Fri May 15 02:35:56 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0