

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051424\
 Data File : PP064885.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 May 2024 09:03
 Operator : YP\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 14 14:19:55 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 14 06:05:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

25)	L5	AR-1248-5	0.000	5.253	0	531624	N.D.	6.452 #
28)	L6	AR-1254-3	6.878	0.000	939003	0	14.823	N.D. #
30)	L6	AR-1254-5	7.591f	0.000	317615	0	7.051	N.D. #
32)	L7	AR-1260-2	0.000	6.066	0	257262	N.D.	1.900 #
33)	L7	AR-1260-3	7.663	0.000	50828	0	1.311	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051424\
Data File : PP064885.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 May 2024 09:03
Operator : YP\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 14 14:19:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051324.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 14 06:05:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 4.260 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.372 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.910 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 8.267 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 6.349 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
HEXANE




#25 AR-1248-5

R.T.: 5.253 min
Delta R.T.: -0.004 min
Response: 531624
Conc: 6.45 ng/ml




#28 AR-1254-3

R.T.: 6.878 min
Delta R.T.: 0.009 min
Response: 939003
Conc: 14.82 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 5.757 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 7.591 min
Delta R.T.: 0.016 min
Response: 317615
Conc: 7.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEXANE




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 6.398 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.293 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 6.066 min
Delta R.T.: -0.013 min
Response: 257262
Conc: 1.90 ng/ml



#33 AR-1260-3

R.T.: 7.663 min
Delta R.T.: 0.011 min
Response: 50828
Conc: 1.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEXANE



#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.230 min
Response: 0
Conc: N.D.