

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051424\
 Data File : PP064886.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 May 2024 09:19
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 14 14:20:17 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051324.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 14 06:05:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.279	3.392	93698666	129.3E6	49.266	57.446
2) SA Decachlor...	9.910	8.267	37085296	90263285	51.648	51.535
Target Compounds						
3) L1 AR-1016-1	5.440	4.447	20540042	39153047	498.974	510.012
4) L1 AR-1016-2	5.461	4.464	30730894	50729496	497.902	471.404m
5) L1 AR-1016-3	5.523	4.635	18859661	31242122	480.862	527.175
6) L1 AR-1016-4	5.621	4.678	14587485	25374026	486.555	522.314
7) L1 AR-1016-5	5.915	4.884	14771952	33346392	506.469	527.215
31) L7 AR-1260-1	7.039	5.897	27696195	62809503	554.888	522.856
32) L7 AR-1260-2	7.296	6.083	27783931	70954316	550.606	524.161
33) L7 AR-1260-3	7.655	6.233	21285221	69714709	549.135	526.486
34) L7 AR-1260-4	7.879	6.698	24257681	56115603	552.930	527.253
35) L7 AR-1260-5	8.188	6.939	39521328	123.2E6	545.085	532.092



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051424\
Data File : PP064886.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 May 2024 09:19
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 14 14:20:17 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051324.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 14 06:05:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.279 min
Delta R.T.: 0.019 min
Response: 93698666
Conc: 49.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.392 min
Delta R.T.: 0.019 min
Response: 129319546
Conc: 57.45 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.910 min
Delta R.T.: 0.000 min
Response: 37085296
Conc: 51.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.267 min
Delta R.T.: 0.000 min
Response: 90263285
Conc: 51.53 ng/ml



#3 AR-1016-1


R.T.: 5.440 min
Delta R.T.: 0.011 min
Response: 20540042
Conc: 498.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.447 min
Delta R.T.: 0.015 min
Response: 39153047
Conc: 510.01 ng/ml




#4 AR-1016-2

R.T.: 5.461 min
Delta R.T.: 0.010 min
Response: 30730894
Conc: 497.90 ng/ml



#4 AR-1016-2


R.T.: 4.464 min
Delta R.T.: 0.014 min
Response: 50729496
Conc: 471.40 ng/ml m



#5 AR-1016-3


R.T.: 5.523 min
Delta R.T.: 0.010 min
Response: 18859661
Conc: 480.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.635 min
Delta R.T.: 0.013 min
Response: 31242122
Conc: 527.18 ng/ml




#6 AR-1016-4

R.T.: 5.621 min
Delta R.T.: 0.009 min
Response: 14587485
Conc: 486.56 ng/ml



#6 AR-1016-4


R.T.: 4.678 min
Delta R.T.: 0.013 min
Response: 25374026
Conc: 522.31 ng/ml



#7 AR-1016-5


R.T.: 5.915 min
Delta R.T.: 0.007 min
Response: 14771952
Conc: 506.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 4.884 min
Delta R.T.: 0.011 min
Response: 33346392
Conc: 527.21 ng/ml




#31 AR-1260-1

R.T.: 7.039 min
Delta R.T.: 0.003 min
Response: 27696195
Conc: 554.89 ng/ml



#31 AR-1260-1


R.T.: 5.897 min
Delta R.T.: 0.005 min
Response: 62809503
Conc: 522.86 ng/ml



#32 AR-1260-2


R.T.: 7.296 min
Delta R.T.: 0.003 min
Response: 27783931
Conc: 550.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.083 min
Delta R.T.: 0.004 min
Response: 70954316
Conc: 524.16 ng/ml




#33 AR-1260-3

R.T.: 7.655 min
Delta R.T.: 0.002 min
Response: 21285221
Conc: 549.14 ng/ml



#33 AR-1260-3


R.T.: 6.233 min
Delta R.T.: 0.003 min
Response: 69714709
Conc: 526.49 ng/ml



#34 AR-1260-4


R.T.: 7.879 min
Delta R.T.: 0.002 min
Response: 24257681
Conc: 552.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.698 min
Delta R.T.: 0.003 min
Response: 56115603
Conc: 527.25 ng/ml



#35 AR-1260-5

R.T.: 8.188 min
Delta R.T.: 0.001 min
Response: 39521328
Conc: 545.09 ng/ml



#35 AR-1260-5

R.T.: 6.939 min
Delta R.T.: 0.001 min
Response: 123233101
Conc: 532.09 ng/ml