

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051622\
 Data File : PP047591.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 May 2022 10:04
 Operator : YP\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 16 13:09:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 11 13:10:20 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.844f	3.107	217472	117767	0.086	0.036 #
2)	SA Decachlor...	9.923	8.413	259037	193850	0.181	0.071 #
Target Compounds							
3)	L1 AR-1016-1	5.097	0.000	128219	0	1.740	N.D. #
4)	L1 AR-1016-2	5.119	0.000	178062	0	1.595	N.D. #
5)	L1 AR-1016-3	5.188	4.437	66481	44757	0.969	0.495 #
6)	L1 AR-1016-4	5.288	4.514f	257072	85151	4.400	1.166 #
7)	L1 AR-1016-5	5.600	0.000	257186	0	4.712	N.D. #
8)	L2 AR-1221-1	4.090	3.352	560238	221357	19.670	5.552 #
9)	L2 AR-1221-2	4.180	3.446f	281368	307994	13.709	9.604 #
10)	L2 AR-1221-3	4.248	0.000	452635	0	7.017	N.D. #
11)	L3 AR-1232-1	4.248	0.000	452635	0	7.427	N.D. #
12)	L3 AR-1232-2	4.811	0.000	132246	0	4.365	N.D. #
13)	L3 AR-1232-3	5.119	4.437	178062	44757	3.217	1.007 #
14)	L3 AR-1232-4	5.288	4.554f	257072	124768	9.592	3.178 #
15)	L3 AR-1232-5	5.388	0.000	189124	0	10.170	N.D. #
16)	L4 AR-1242-1	5.097	0.000	128219	0	2.227	N.D. #
17)	L4 AR-1242-2	5.119	0.000	178062	0	2.043	N.D. #
18)	L4 AR-1242-3	5.180	4.437	134406	44757	2.594	0.634 #
19)	L4 AR-1242-4	5.288	4.554f	257072	124768	5.176	1.814 #
20)	L4 AR-1242-5	6.099	5.108f	661424	378512	14.490	4.246 #
21)	L5 AR-1248-1	5.097	0.000	128219	0	2.953	N.D. #
22)	L5 AR-1248-2	5.388	4.514f	189124	85151	3.018	0.849 #
23)	L5 AR-1248-3	5.600	4.554f	257186	124768	3.499	1.201 #
24)	L5 AR-1248-4	6.049	0.000	232259	0	2.870	N.D. #
25)	L5 AR-1248-5	6.099	5.108f	661424	378512	8.434	2.937 #
26)	L6 AR-1254-1	6.022	5.108f	337854	378512	4.319	2.029 #
27)	L6 AR-1254-2	6.255	0.000	496471	0	4.083	N.D. #
28)	L6 AR-1254-3	6.658	0.000	321650	0	2.452	N.D. #
29)	L6 AR-1254-4	6.976	5.924	1388817	4105328	14.586	22.420 #
30)	L6 AR-1254-5	7.417	0.000	746341	0	7.651	N.D. #
31)	L7 AR-1260-1	6.824	0.000	510028	0	5.638	N.D. #
32)	L7 AR-1260-2	7.101	6.013	299670	543284	2.581	2.240 #
33)	L7 AR-1260-3	7.497	0.000	1516424	0	14.713	N.D. #
34)	L7 AR-1260-4	7.761	0.000	2508960	0	25.964	N.D. #
35)	L7 AR-1260-5	8.085	6.955	2936343	1108460	16.415	2.356 #
36)	L8 AR-1262-1	7.497	0.000	1516424	0	9.921	N.D. #
37)	L8 AR-1262-2	8.085	0.000	2936343	0	13.759	N.D. #
38)	L8 AR-1262-3	8.410	0.000	65749	0	0.466	N.D. #
39)	L8 AR-1262-4	8.503	7.331	208047	3661617	1.867	10.139 #
40)	L8 AR-1262-5	9.177	7.895f	139190	6270658	1.956	36.423 #
41)	L9 AR-1268-1	8.410	0.000	65749	0	0.252	N.D. #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051622\
 Data File : PP047591.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 May 2022 10:04
 Operator : YP\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 16 13:09:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 11 13:10:20 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.512	7.331	162162	3661617	0.703	6.907 #
43) L9	AR-1268-3	8.721f	7.557	5020948	9363227	24.967	20.241
44) L9	AR-1268-4	9.177	7.895f	139190	6270658	1.737	31.715 #
45) L9	AR-1268-5	9.589	0.000	277153	0	0.487	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051622\
Data File : PP047591.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 May 2022 10:04
Operator : YP\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 16 13:09:42 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 11 13:10:20 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.844 min
Delta R.T.: -0.016 min
Response: 217472
Conc: 0.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.107 min
Delta R.T.: -0.008 min
Response: 117767
Conc: 0.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.923 min
Delta R.T.: 0.009 min
Response: 259037
Conc: 0.18 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.413 min
Delta R.T.: -0.014 min
Response: 193850
Conc: 0.07 ng/ml



#3 AR-1016-1


R.T.: 5.097 min
Delta R.T.: 0.000 min
Response: 128219
Conc: 1.74 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.247 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.119 min
Delta R.T.: -0.002 min
Response: 178062
Conc: 1.59 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.264 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.188 min
Delta R.T.: 0.000 min
Response: 66481
Conc: 0.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.437 min
Delta R.T.: -0.012 min
Response: 44757
Conc: 0.49 ng/ml




#6 AR-1016-4

R.T.: 5.288 min
Delta R.T.: -0.004 min
Response: 257072
Conc: 4.40 ng/ml



#6 AR-1016-4


R.T.: 4.514 min
Delta R.T.: 0.017 min
Response: 85151
Conc: 1.17 ng/ml



#7 AR-1016-5


R.T.: 5.600 min
Delta R.T.: -0.010 min
Response: 257186
Conc: 4.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 4.720 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.090 min
Delta R.T.: 0.007 min
Response: 560238
Conc: 19.67 ng/ml



#8 AR-1221-1


R.T.: 3.352 min
Delta R.T.: 0.013 min
Response: 221357
Conc: 5.55 ng/ml



#9 AR-1221-2


R.T.: 4.180 min
Delta R.T.: 0.005 min
Response: 281368
Conc: 13.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.446 min
Delta R.T.: 0.018 min
Response: 307994
Conc: 9.60 ng/ml




#10 AR-1221-3

R.T.: 4.248 min
Delta R.T.: -0.006 min
Response: 452635
Conc: 7.02 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.506 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.248 min
Delta R.T.: 0.001 min
Response: 452635
Conc: 7.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.505 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 4.811 min
Delta R.T.: 0.004 min
Response: 132246
Conc: 4.36 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T. : 4.263 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.119 min
Delta R.T.: 0.001 min
Response: 178062
Conc: 3.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.437 min
Delta R.T.: -0.010 min
Response: 44757
Conc: 1.01 ng/ml




#14 AR-1232-4

R.T.: 5.288 min
Delta R.T.: 0.000 min
Response: 257072
Conc: 9.59 ng/ml



#14 AR-1232-4


R.T.: 4.554 min
Delta R.T.: 0.016 min
Response: 124768
Conc: 3.18 ng/ml



#15 AR-1232-5


R.T.: 5.388 min
Delta R.T.: -0.002 min
Response: 189124
Conc: 10.17 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 4.717 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.097 min
Delta R.T.: 0.004 min
Response: 128219
Conc: 2.23 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.245 min
Response: 0
Conc: N.D.



#17 AR-1242-2


R.T.: 5.119 min
Delta R.T.: 0.002 min
Response: 178062
Conc: 2.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.263 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.180 min
Delta R.T.: -0.004 min
Response: 134406
Conc: 2.59 ng/ml



#18 AR-1242-3


R.T.: 4.437 min
Delta R.T.: -0.010 min
Response: 44757
Conc: 0.63 ng/ml



#19 AR-1242-4


R.T.: 5.288 min
Delta R.T.: 0.000 min
Response: 257072
Conc: 5.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.554 min
Delta R.T.: 0.017 min
Response: 124768
Conc: 1.81 ng/ml




#20 AR-1242-5

R.T.: 6.099 min
Delta R.T.: 0.013 min
Response: 661424
Conc: 14.49 ng/ml



#20 AR-1242-5


R.T.: 5.108 min
Delta R.T.: 0.017 min
Response: 378512
Conc: 4.25 ng/ml



#21 AR-1248-1


R.T.: 5.097 min
Delta R.T.: 0.002 min
Response: 128219
Conc: 2.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.244 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.388 min
Delta R.T.: -0.002 min
Response: 189124
Conc: 3.02 ng/ml



#22 AR-1248-2


R.T.: 4.514 min
Delta R.T.: 0.020 min
Response: 85151
Conc: 0.85 ng/ml



#23 AR-1248-3


R.T.: 5.600 min
Delta R.T.: -0.007 min
Response: 257186
Conc: 3.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.554 min
Delta R.T.: 0.018 min
Response: 124768
Conc: 1.20 ng/ml




#24 AR-1248-4

R.T.: 6.049 min
Delta R.T.: 0.004 min
Response: 232259
Conc: 2.87 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 4.716 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.099 min
Delta R.T.: 0.013 min
Response: 661424
Conc: 8.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.108 min
Delta R.T.: -0.024 min
Response: 378512
Conc: 2.94 ng/ml




#26 AR-1254-1

R.T.: 6.022 min
Delta R.T.: 0.008 min
Response: 337854
Conc: 4.32 ng/ml



#26 AR-1254-1


R.T.: 5.108 min
Delta R.T.: 0.018 min
Response: 378512
Conc: 2.03 ng/ml



#27 AR-1254-2


R.T.: 6.255 min
Delta R.T.: 0.005 min
Response: 496471
Conc: 4.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 5.249 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 6.658 min
Delta R.T.: 0.006 min
Response: 321650
Conc: 2.45 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 5.678 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.976 min
Delta R.T.: 0.014 min
Response: 1388817
Conc: 14.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.924 min
Delta R.T.: -0.003 min
Response: 4105328
Conc: 22.42 ng/ml




#30 AR-1254-5

R.T.: 7.417 min
Delta R.T.: -0.005 min
Response: 746341
Conc: 7.65 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T. : 6.374 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 6.824 min
Delta R.T.: -0.005 min
Response: 510028
Conc: 5.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 5.820 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 7.101 min
Delta R.T.: -0.010 min
Response: 299670
Conc: 2.58 ng/ml



#32 AR-1260-2


R.T.: 6.013 min
Delta R.T.: -0.013 min
Response: 543284
Conc: 2.24 ng/ml



#33 AR-1260-3


R.T.: 7.497 min
Delta R.T.: -0.007 min
Response: 1516424
Conc: 14.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.186 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 7.761 min
Delta R.T.: 0.009 min
Response: 2508960
Conc: 25.96 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 6.696 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.085 min
Delta R.T.: -0.006 min
Response: 2936343
Conc: 16.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 6.955 min
Delta R.T.: -0.009 min
Response: 1108460
Conc: 2.36 ng/ml




#36 AR-1262-1

R.T.: 7.497 min
Delta R.T.: -0.008 min
Response: 1516424
Conc: 9.92 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 6.413 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.085 min
Delta R.T.: -0.006 min
Response: 2936343
Conc: 13.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 6.691 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.410 min
Delta R.T.: -0.005 min
Response: 65749
Conc: 0.47 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 7.263 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 8.503 min
Delta R.T.: -0.004 min
Response: 208047
Conc: 1.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.331 min
Delta R.T.: -0.003 min
Response: 3661617
Conc: 10.14 ng/ml




#40 AR-1262-5

R.T.: 9.177 min
Delta R.T.: 0.000 min
Response: 139190
Conc: 1.96 ng/ml



#40 AR-1262-5


R.T.: 7.895 min
Delta R.T.: 0.019 min
Response: 6270658
Conc: 36.42 ng/ml



#41 AR-1268-1


R.T.: 8.410 min
Delta R.T.: -0.002 min
Response: 65749
Conc: 0.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 7.262 min
Response: 0
Conc: N.D.




#42 AR-1268-2

R.T.: 8.512 min
Delta R.T.: 0.003 min
Response: 162162
Conc: 0.70 ng/ml



#42 AR-1268-2


R.T.: 7.331 min
Delta R.T.: -0.003 min
Response: 3661617
Conc: 6.91 ng/ml



#43 AR-1268-3


R.T.: 8.721 min
Delta R.T.: -0.018 min
Response: 5020948
Conc: 24.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.557 min
Delta R.T.: 0.008 min
Response: 9363227
Conc: 20.24 ng/ml




#44 AR-1268-4

R.T.: 9.177 min
Delta R.T.: 0.004 min
Response: 139190
Conc: 1.74 ng/ml



#44 AR-1268-4


R.T.: 7.895 min
Delta R.T.: 0.020 min
Response: 6270658
Conc: 31.71 ng/ml



#45 AR-1268-5

R.T.: 9.589 min
Delta R.T.: 0.003 min
Response: 277153
Conc: 0.49 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 8.172 min
Response: 0
Conc: N.D.