

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051624\
 Data File : PP065016.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 May 2024 21:05
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 17 04:54:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 17 04:50:34 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.262	3.374	134.3E6	186.1E6	72.976	73.998
2) SA Decachlor...	9.908	8.264	130.6E6	265.6E6	71.148	73.424
Target Compounds						
41) L9 AR-1268-1	8.483	7.215	106.9E6	247.4E6	724.015	742.399
42) L9 AR-1268-2	8.573	7.278	93236176	227.0E6	724.204	741.269
43) L9 AR-1268-3	8.791	7.481	86694880	204.2E6	722.588	734.850
44) L9 AR-1268-4	9.195	7.768	34348658	79417886	715.660	701.663
45) L9 AR-1268-5	9.589	8.037	237.1E6	565.3E6	736.009	743.828



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051624\
Data File : PP065016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 May 2024 21:05
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 17 04:54:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051624.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 17 04:50:34 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.262 min
Delta R.T.: 0.002 min
Response: 134306785
Conc: 72.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.374 min
Delta R.T.: 0.002 min
Response: 186089509
Conc: 74.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.908 min
Delta R.T.: 0.002 min
Response: 130637374
Conc: 71.15 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.264 min
Delta R.T.: 0.002 min
Response: 265565488
Conc: 73.42 ng/ml



#41 AR-1268-1


R.T.: 8.483 min
Delta R.T.: 0.000 min
Response: 106944755
Conc: 724.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750




#41 AR-1268-1

R.T.: 7.215 min
Delta R.T.: 0.001 min
Response: 247403804
Conc: 742.40 ng/ml




#42 AR-1268-2

R.T.: 8.573 min
Delta R.T.: 0.000 min
Response: 93236176
Conc: 724.20 ng/ml



#42 AR-1268-2


R.T.: 7.278 min
Delta R.T.: 0.001 min
Response: 227044315
Conc: 741.27 ng/ml



#43 AR-1268-3


R.T.: 8.791 min
Delta R.T.: 0.000 min
Response: 86694880
Conc: 722.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750




#43 AR-1268-3

R.T.: 7.481 min
Delta R.T.: 0.002 min
Response: 204186381
Conc: 734.85 ng/ml




#44 AR-1268-4

R.T.: 9.195 min
Delta R.T.: 0.000 min
Response: 34348658
Conc: 715.66 ng/ml



#44 AR-1268-4


R.T.: 7.768 min
Delta R.T.: 0.002 min
Response: 79417886
Conc: 701.66 ng/ml



#45 AR-1268-5

R.T.: 9.589 min
Delta R.T.: 0.002 min
Response: 237113073
Conc: 736.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750



#45 AR-1268-5

R.T.: 8.037 min
Delta R.T.: 0.002 min
Response: 565305661
Conc: 743.83 ng/ml