

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051624\
 Data File : PP065002.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 May 2024 17:17
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 17 02:15:16 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 17 02:14:27 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.257	3.371	48194523	65353111	25.661	25.570
2) SA Decachlor...	9.909	8.262	32364795	57171905	27.434	26.894
Target Compounds						
16) L4 AR-1242-1	5.423	4.427	13925392	18416856	270.404	267.527
17) L4 AR-1242-2	5.445	4.444	20902132	25364118	269.691	264.345
18) L4 AR-1242-3	5.507	4.615	13480680	14229044	272.666	267.843
19) L4 AR-1242-4	5.605	4.698	10596265	14789643	267.515	274.189
20) L4 AR-1242-5	6.345	5.211	11344593	17466564	267.664	268.713



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051624\
Data File : PP065002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 May 2024 17:17
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 17 02:15:16 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051624.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 17 02:14:27 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.257 min
Delta R.T.: 0.000 min
Response: 48194523
Conc: 25.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.371 min
Delta R.T.: 0.000 min
Response: 65353111
Conc: 25.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.909 min
Delta R.T.: 0.000 min
Response: 32364795
Conc: 27.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.262 min
Delta R.T.: 0.000 min
Response: 57171905
Conc: 26.89 ng/ml



#16 AR-1242-1


R.T.: 5.423 min
Delta R.T.: 0.000 min
Response: 13925392
Conc: 270.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.427 min
Delta R.T.: 0.000 min
Response: 18416856
Conc: 267.53 ng/ml




#17 AR-1242-2

R.T.: 5.445 min
Delta R.T.: 0.000 min
Response: 20902132
Conc: 269.69 ng/ml



#17 AR-1242-2


R.T.: 4.444 min
Delta R.T.: 0.000 min
Response: 25364118
Conc: 264.35 ng/ml



#18 AR-1242-3


R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 13480680
Conc: 272.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.615 min
Delta R.T.: 0.000 min
Response: 14229044
Conc: 267.84 ng/ml




#19 AR-1242-4

R.T.: 5.605 min
Delta R.T.: 0.000 min
Response: 10596265
Conc: 267.51 ng/ml



#19 AR-1242-4


R.T.: 4.698 min
Delta R.T.: 0.000 min
Response: 14789643
Conc: 274.19 ng/ml



#20 AR-1242-5

R.T.: 6.345 min
Delta R.T.: 0.000 min
Response: 11344593
Conc: 267.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.211 min
Delta R.T.: 0.000 min
Response: 17466564
Conc: 268.71 ng/ml