

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051723\
 Data File : PP057685.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 May 2023 21:02
 Operator : YP\AJ
 Sample : 02812-16
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 WG-0516-B2-0-10

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 18 06:09:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.386	3.623	45220900	44208017	20.613	19.489
2) SA Decachlor...	10.212	8.686	20234535	26360096	17.855	16.793

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051723\
Data File : PP057685.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 May 2023 21:02
Operator : YP\AJ
Sample : 02812-16
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
WG-0516-B2-0-10

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 18 06:09:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.386 min
Delta R.T.: -0.004 min
Response: 45220900
Conc: 20.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
WG-0516-B2-0-10




#1 Tetrachloro-m-xylene

R.T.: 3.623 min
Delta R.T.: 0.000 min
Response: 44208017
Conc: 19.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.212 min
Delta R.T.: -0.011 min
Response: 20234535
Conc: 17.86 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.686 min
Delta R.T.: 0.000 min
Response: 26360096
Conc: 16.79 ng/ml