

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051723\
 Data File : PP057689.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 May 2023 22:29
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 18 06:11:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.387	3.624	110.8E6	113.2E6	50.526	49.903
2) SA Decachlor...	10.212	8.686	61144261	75086094	53.955	47.833
Target Compounds						
16) L4 AR-1242-1	5.566	4.721	24525764	28678530	514.836	492.665
17) L4 AR-1242-2	5.589	4.740	35458229	38978721	518.495	488.344
18) L4 AR-1242-3	5.651	4.918	22934366	23797035	526.737	507.419
19) L4 AR-1242-4	5.750	5.002	18390442	23290383	521.907	501.944
20) L4 AR-1242-5	6.494	5.531	16836451	27496674	519.623	491.805



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051723\
Data File : PP057689.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 May 2023 22:29
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 18 06:11:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.387 min
Delta R.T.: -0.003 min
Response: 110845400
Conc: 50.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 113196129
Conc: 49.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.212 min
Delta R.T.: -0.011 min
Response: 61144261
Conc: 53.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.686 min
Delta R.T.: 0.000 min
Response: 75086094
Conc: 47.83 ng/ml



#16 AR-1242-1


R.T.: 5.566 min
Delta R.T.: -0.002 min
Response: 24525764
Conc: 514.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.721 min
Delta R.T.: 0.000 min
Response: 28678530
Conc: 492.66 ng/ml




#17 AR-1242-2

R.T.: 5.589 min
Delta R.T.: -0.002 min
Response: 35458229
Conc: 518.49 ng/ml



#17 AR-1242-2


R.T.: 4.740 min
Delta R.T.: 0.000 min
Response: 38978721
Conc: 488.34 ng/ml



#18 AR-1242-3


R.T.: 5.651 min
Delta R.T.: -0.003 min
Response: 22934366
Conc: 526.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.918 min
Delta R.T.: 0.000 min
Response: 23797035
Conc: 507.42 ng/ml




#19 AR-1242-4

R.T.: 5.750 min
Delta R.T.: -0.003 min
Response: 18390442
Conc: 521.91 ng/ml



#19 AR-1242-4


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 23290383
Conc: 501.94 ng/ml



#20 AR-1242-5

R.T.: 6.494 min
Delta R.T.: -0.003 min
Response: 16836451
Conc: 519.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.531 min
Delta R.T.: 0.000 min
Response: 27496674
Conc: 491.81 ng/ml