

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051920\
 Data File : PP030665.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 May 2020 15:47
 Operator : DD\AJ
 Sample : INDB201
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 07:35:43 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051920CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed May 20 07:33:55 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	2.637	2.739	15097335	5557332	10.312	10.120
27)	SA Decachlor...	7.347	7.776	17782475	10850642	20.680	21.113
Target Compounds							
5)	MB Aldrin	3.973	4.200	20076384	8324709	10.346	10.253
6)	B beta-BHC	3.370	3.577	9514460	3632180	10.410	10.034
7)	B delta-BHC	3.520	3.844	21764456	8647262	10.344	10.035
8)	B Heptachlo...	4.555	4.774	18825627	7806100	10.556	10.361
10)	B trans-Chl...	4.679	4.958	18590235	7560163	10.450	10.272
11)	B cis-Chlor...	4.811	5.101	17738566	7570001	10.241	10.313
12)	B 4,4'-DDE	4.890	5.269	34334720	14170568	20.682	20.638
15)	B Endosulfa...	5.588	5.881	28044842	12976142	20.799	21.049
18)	B Endrin al...	5.958	6.173	22359820	10824578	21.134	20.811
19)	B Endosulfa...	6.326	6.403	24562037	12984401	20.716	21.194
21)	B Endrin ke...	6.555	6.842	25565603	14126201	20.432	21.167

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051920\
Data File : PP030665.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 May 2020 15:47
Operator : DD\AJ
Sample : INDB201
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 07:35:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051920CLP.M
Quant Title : GC Extractables
QLast Update : Wed May 20 07:33:55 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm




Instrument :
ECD_P
ClientSampled :



#1 Tetrachloro-m-xylene


R.T.: 2.637 min
Delta R.T.: 0.000 min
Response: 15097335
Conc: 10.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.739 min
Delta R.T.: 0.000 min
Response: 5557332
Conc: 10.12 ng/ml




#5 Aldrin

R.T.: 3.973 min
Delta R.T.: 0.000 min
Response: 20076384
Conc: 10.35 ng/ml



#5 Aldrin


R.T.: 4.200 min
Delta R.T.: 0.000 min
Response: 8324709
Conc: 10.25 ng/ml



#6 beta-BHC


R.T.: 3.370 min
Delta R.T.: 0.000 min
Response: 9514460
Conc: 10.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#6 beta-BHC

R.T.: 3.577 min
Delta R.T.: 0.000 min
Response: 3632180
Conc: 10.03 ng/ml




#7 delta-BHC

R.T.: 3.520 min
Delta R.T.: 0.000 min
Response: 21764456
Conc: 10.34 ng/ml



#7 delta-BHC


R.T.: 3.844 min
Delta R.T.: 0.000 min
Response: 8647262
Conc: 10.04 ng/ml



#8 Heptachlor epoxide


R.T.: 4.555 min
Delta R.T.: 0.000 min
Response: 18825627
Conc: 10.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 4.774 min
Delta R.T.: 0.000 min
Response: 7806100
Conc: 10.36 ng/ml




#10 trans-Chlordane

R.T.: 4.679 min
Delta R.T.: 0.000 min
Response: 18590235
Conc: 10.45 ng/ml



#10 trans-Chlordane


R.T.: 4.958 min
Delta R.T.: 0.000 min
Response: 7560163
Conc: 10.27 ng/ml



#11 cis-Chlordane


R.T.: 4.811 min
Delta R.T.: 0.001 min
Response: 17738566
Conc: 10.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 cis-Chlordane

R.T.: 5.101 min
Delta R.T.: 0.000 min
Response: 7570001
Conc: 10.31 ng/ml




#12 4,4'-DDE

R.T.: 4.890 min
Delta R.T.: 0.000 min
Response: 34334720
Conc: 20.68 ng/ml



#12 4,4'-DDE


R.T.: 5.269 min
Delta R.T.: 0.000 min
Response: 14170568
Conc: 20.64 ng/ml



#15 Endosulfan II


R.T.: 5.588 min
Delta R.T.: 0.000 min
Response: 28044842
Conc: 20.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 Endosulfan II

R.T.: 5.881 min
Delta R.T.: 0.000 min
Response: 12976142
Conc: 21.05 ng/ml




#18 Endrin aldehyde

R.T.: 5.958 min
Delta R.T.: 0.000 min
Response: 22359820
Conc: 21.13 ng/ml



#18 Endrin aldehyde


R.T.: 6.173 min
Delta R.T.: 0.000 min
Response: 10824578
Conc: 20.81 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.326 min
Delta R.T.: 0.000 min
Response: 24562037
Conc: 20.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 6.403 min
Delta R.T.: 0.000 min
Response: 12984401
Conc: 21.19 ng/ml




#21 Endrin ketone

R.T.: 6.555 min
Delta R.T.: 0.000 min
Response: 25565603
Conc: 20.43 ng/ml



#21 Endrin ketone


R.T.: 6.842 min
Delta R.T.: 0.000 min
Response: 14126201
Conc: 21.17 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.347 min
Delta R.T.: 0.000 min
Response: 17782475
Conc: 20.68 ng/ml

Instrument :
ECD_P
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 7.776 min
Delta R.T.: 0.000 min
Response: 10850642
Conc: 21.11 ng/ml