

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051922\
 Data File : PP047774.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2022 13:03
 Operator : YP\AJ
 Sample : N2904-17 10X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 15:33:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 19 15:06:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.841	3.108	2870698	4936134	1.129	1.499 #
2)	SA Decachlor...	9.894f	8.407	1593902	5031604	1.114	1.849 #
Target Compounds							
3)	L1 AR-1016-1	5.096	4.259	138525	68364	1.880	0.586 #
4)	L1 AR-1016-2	5.124	4.259	110325	68364	0.988	0.394 #
5)	L1 AR-1016-3	5.190	4.455	65861	18415	0.960	0.204 #
6)	L1 AR-1016-4	5.289	0.000	252047	0	4.314	N.D. #
7)	L1 AR-1016-5	5.612	4.704f	155216	23674	2.844	0.252 #
8)	L2 AR-1221-1	4.089	3.328	158440	54115	5.563	1.357 #
9)	L2 AR-1221-2	4.172	3.434	128970	24765	6.284	0.772 #
10)	L2 AR-1221-3	4.251	3.520	90206	17569	1.398	0.192 #
11)	L3 AR-1232-1	4.243	3.520f	213579	17569	3.505	0.199 #
12)	L3 AR-1232-2	4.802	4.259	162410	68364	5.360	0.787 #
13)	L3 AR-1232-3	5.124	4.455	110325	18415	1.993	0.414 #
14)	L3 AR-1232-4	5.289	4.532	252047	171846	9.404	4.378 #
15)	L3 AR-1232-5	5.395	4.704	220390	23674	11.852	0.535 #
16)	L4 AR-1242-1	5.096	4.259	138525	68364	2.406	0.748 #
17)	L4 AR-1242-2	5.124	4.259	110325	68364	1.266	0.504 #
18)	L4 AR-1242-3	5.178	4.455	167312	18415	3.229	0.261 #
19)	L4 AR-1242-4	5.289	4.532	252047	171846	5.075	2.498 #
20)	L4 AR-1242-5	6.094	5.072f	261334	49621	5.725	0.557 #
21)	L5 AR-1248-1	5.096	4.231	138525	71844	3.190	1.005 #
22)	L5 AR-1248-2	5.395	0.000	220390	0	3.517	N.D. #
23)	L5 AR-1248-3	5.612	4.532	155216	171846	2.111	1.654
24)	L5 AR-1248-4	6.052	4.704	131181	23674	1.621	0.192 #
25)	L5 AR-1248-5	6.094	0.000	261334	0	3.332	N.D. #
26)	L6 AR-1254-1	5.998f	5.072f	747268	49621	9.553	0.266 #
27)	L6 AR-1254-2	6.252	5.241	333955	626631	2.746	3.802 #
28)	L6 AR-1254-3	6.653	0.000	640252	0	4.880	N.D. #
29)	L6 AR-1254-4	6.970	5.951f	127261	754148	1.337	4.119 #
30)	L6 AR-1254-5	7.425	6.362	522956	505480	5.361	2.105 #
31)	L7 AR-1260-1	6.822	5.803	526908	249652	5.825	1.465 #
32)	L7 AR-1260-2	7.112	6.009	20161	910072	0.174	3.752 #
33)	L7 AR-1260-3	7.507	6.171	650630	530897	6.313	2.657 #
34)	L7 AR-1260-4	7.771f	6.680	330372	1577788	3.419	8.930 #
35)	L7 AR-1260-5	8.084	6.948	445339	2575518	2.490	5.475 #
36)	L8 AR-1262-1	7.507	6.403	650630	346186	4.257	1.064 #
37)	L8 AR-1262-2	8.084	6.680	445339	1577788	2.087	6.457 #
38)	L8 AR-1262-3	8.422	7.267	444268	826853	3.148	3.838
39)	L8 AR-1262-4	8.497	7.336	643104	49593	5.770	0.137 #
40)	L8 AR-1262-5	9.183	0.000	246976	0	3.470	N.D. #
41)	L9 AR-1268-1	8.404	7.267	467647	826853	1.794	1.279 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051922\
 Data File : PP047774.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2022 13:03
 Operator : YP\AJ
 Sample : N2904-17 10X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 15:33:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 19 15:06:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.497	7.336	643104	49593	2.789	0.094 #
43) L9	AR-1268-3	8.739	7.543	162491	3322121	0.808	7.182 #
44) L9	AR-1268-4	9.183	0.000	246976	0	3.081	N.D. #
45) L9	AR-1268-5	9.582	8.164	103934	3899381	0.183	3.134 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051922\
Data File : PP047774.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2022 13:03
Operator : YP\AJ
Sample : N2904-17 10X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 15:33:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 15:06:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.841 min
Delta R.T.: -0.009 min
Response: 2870698
Conc: 1.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.108 min
Delta R.T.: -0.008 min
Response: 4936134
Conc: 1.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.894 min
Delta R.T.: -0.020 min
Response: 1593902
Conc: 1.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.407 min
Delta R.T.: 0.000 min
Response: 5031604
Conc: 1.85 ng/ml



#3 AR-1016-1


R.T.: 5.096 min
Delta R.T.: -0.002 min
Response: 138525
Conc: 1.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.259 min
Delta R.T.: 0.013 min
Response: 68364
Conc: 0.59 ng/ml




#4 AR-1016-2

R.T.: 5.124 min
Delta R.T.: 0.003 min
Response: 110325
Conc: 0.99 ng/ml



#4 AR-1016-2


R.T.: 4.259 min
Delta R.T.: -0.005 min
Response: 68364
Conc: 0.39 ng/ml



#5 AR-1016-3


R.T.: 5.190 min
Delta R.T.: 0.003 min
Response: 65861
Conc: 0.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.455 min
Delta R.T.: 0.007 min
Response: 18415
Conc: 0.20 ng/ml




#6 AR-1016-4

R.T.: 5.289 min
Delta R.T.: -0.003 min
Response: 252047
Conc: 4.31 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 4.497 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 5.612 min
Delta R.T.: 0.002 min
Response: 155216
Conc: 2.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.704 min
Delta R.T.: -0.016 min
Response: 23674
Conc: 0.25 ng/ml




#8 AR-1221-1

R.T.: 4.089 min
Delta R.T.: 0.006 min
Response: 158440
Conc: 5.56 ng/ml



#8 AR-1221-1


R.T.: 3.328 min
Delta R.T.: -0.011 min
Response: 54115
Conc: 1.36 ng/ml



#9 AR-1221-2


R.T.: 4.172 min
Delta R.T.: -0.003 min
Response: 128970
Conc: 6.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.434 min
Delta R.T.: 0.005 min
Response: 24765
Conc: 0.77 ng/ml




#10 AR-1221-3

R.T.: 4.251 min
Delta R.T.: -0.003 min
Response: 90206
Conc: 1.40 ng/ml



#10 AR-1221-3


R.T.: 3.520 min
Delta R.T.: 0.014 min
Response: 17569
Conc: 0.19 ng/ml



#11 AR-1232-1


R.T.: 4.243 min
Delta R.T.: -0.004 min
Response: 213579
Conc: 3.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.520 min
Delta R.T.: 0.015 min
Response: 17569
Conc: 0.20 ng/ml




#12 AR-1232-2

R.T.: 4.802 min
Delta R.T.: -0.004 min
Response: 162410
Conc: 5.36 ng/ml



#12 AR-1232-2


R.T.: 4.259 min
Delta R.T.: -0.003 min
Response: 68364
Conc: 0.79 ng/ml



#13 AR-1232-3


R.T.: 5.124 min
Delta R.T.: 0.006 min
Response: 110325
Conc: 1.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.455 min
Delta R.T.: 0.009 min
Response: 18415
Conc: 0.41 ng/ml




#14 AR-1232-4

R.T.: 5.289 min
Delta R.T.: 0.000 min
Response: 252047
Conc: 9.40 ng/ml



#14 AR-1232-4


R.T.: 4.532 min
Delta R.T.: -0.005 min
Response: 171846
Conc: 4.38 ng/ml



#15 AR-1232-5


R.T.: 5.395 min
Delta R.T.: 0.006 min
Response: 220390
Conc: 11.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.704 min
Delta R.T.: -0.013 min
Response: 23674
Conc: 0.54 ng/ml




#16 AR-1242-1

R.T.: 5.096 min
Delta R.T.: 0.003 min
Response: 138525
Conc: 2.41 ng/ml



#16 AR-1242-1


R.T.: 4.259 min
Delta R.T.: 0.014 min
Response: 68364
Conc: 0.75 ng/ml



#17 AR-1242-2


R.T.: 5.124 min
Delta R.T.: 0.006 min
Response: 110325
Conc: 1.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.259 min
Delta R.T.: -0.004 min
Response: 68364
Conc: 0.50 ng/ml




#18 AR-1242-3

R.T.: 5.178 min
Delta R.T.: -0.005 min
Response: 167312
Conc: 3.23 ng/ml



#18 AR-1242-3


R.T.: 4.455 min
Delta R.T.: 0.008 min
Response: 18415
Conc: 0.26 ng/ml



#19 AR-1242-4


R.T.: 5.289 min
Delta R.T.: 0.000 min
Response: 252047
Conc: 5.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.532 min
Delta R.T.: -0.005 min
Response: 171846
Conc: 2.50 ng/ml




#20 AR-1242-5

R.T.: 6.094 min
Delta R.T.: 0.008 min
Response: 261334
Conc: 5.73 ng/ml



#20 AR-1242-5


R.T.: 5.072 min
Delta R.T.: -0.019 min
Response: 49621
Conc: 0.56 ng/ml



#21 AR-1248-1


R.T.: 5.096 min
Delta R.T.: 0.001 min
Response: 138525
Conc: 3.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.231 min
Delta R.T.: -0.014 min
Response: 71844
Conc: 1.01 ng/ml




#22 AR-1248-2

R.T.: 5.395 min
Delta R.T.: 0.006 min
Response: 220390
Conc: 3.52 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T. : 4.494 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 5.612 min
Delta R.T.: 0.006 min
Response: 155216
Conc: 2.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.532 min
Delta R.T.: -0.004 min
Response: 171846
Conc: 1.65 ng/ml




#24 AR-1248-4

R.T.: 6.052 min
Delta R.T.: 0.007 min
Response: 131181
Conc: 1.62 ng/ml



#24 AR-1248-4


R.T.: 4.704 min
Delta R.T.: -0.012 min
Response: 23674
Conc: 0.19 ng/ml



#25 AR-1248-5


R.T.: 6.094 min
Delta R.T.: 0.007 min
Response: 261334
Conc: 3.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 5.133 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 5.998 min
Delta R.T.: -0.016 min
Response: 747268
Conc: 9.55 ng/ml



#26 AR-1254-1


R.T.: 5.072 min
Delta R.T.: -0.018 min
Response: 49621
Conc: 0.27 ng/ml



#27 AR-1254-2


R.T.: 6.252 min
Delta R.T.: 0.002 min
Response: 333955
Conc: 2.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.241 min
Delta R.T.: -0.008 min
Response: 626631
Conc: 3.80 ng/ml




#28 AR-1254-3

R.T.: 6.653 min
Delta R.T.: 0.000 min
Response: 640252
Conc: 4.88 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 5.678 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.970 min
Delta R.T.: 0.008 min
Response: 127261
Conc: 1.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.951 min
Delta R.T.: 0.024 min
Response: 754148
Conc: 4.12 ng/ml




#30 AR-1254-5

R.T.: 7.425 min
Delta R.T.: 0.004 min
Response: 522956
Conc: 5.36 ng/ml



#30 AR-1254-5


R.T.: 6.362 min
Delta R.T.: -0.012 min
Response: 505480
Conc: 2.11 ng/ml



#31 AR-1260-1


R.T.: 6.822 min
Delta R.T.: -0.007 min
Response: 526908
Conc: 5.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.803 min
Delta R.T.: -0.012 min
Response: 249652
Conc: 1.47 ng/ml




#32 AR-1260-2

R.T.: 7.112 min
Delta R.T.: 0.000 min
Response: 20161
Conc: 0.17 ng/ml



#32 AR-1260-2


R.T.: 6.009 min
Delta R.T.: -0.003 min
Response: 910072
Conc: 3.75 ng/ml



#33 AR-1260-3


R.T.: 7.507 min
Delta R.T.: 0.003 min
Response: 650630
Conc: 6.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.171 min
Delta R.T.: -0.011 min
Response: 530897
Conc: 2.66 ng/ml




#34 AR-1260-4

R.T.: 7.771 min
Delta R.T.: 0.019 min
Response: 330372
Conc: 3.42 ng/ml



#34 AR-1260-4


R.T.: 6.680 min
Delta R.T.: 0.000 min
Response: 1577788
Conc: 8.93 ng/ml



#35 AR-1260-5


R.T.: 8.084 min
Delta R.T.: -0.007 min
Response: 445339
Conc: 2.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 6.948 min
Delta R.T.: -0.011 min
Response: 2575518
Conc: 5.48 ng/ml




#36 AR-1262-1

R.T.: 7.507 min
Delta R.T.: 0.002 min
Response: 650630
Conc: 4.26 ng/ml



#36 AR-1262-1


R.T.: 6.403 min
Delta R.T.: -0.010 min
Response: 346186
Conc: 1.06 ng/ml



#37 AR-1262-2


R.T.: 8.084 min
Delta R.T.: -0.007 min
Response: 445339
Conc: 2.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.680 min
Delta R.T.: -0.011 min
Response: 1577788
Conc: 6.46 ng/ml




#38 AR-1262-3

R.T.: 8.422 min
Delta R.T.: 0.007 min
Response: 444268
Conc: 3.15 ng/ml



#38 AR-1262-3


R.T.: 7.267 min
Delta R.T.: 0.004 min
Response: 826853
Conc: 3.84 ng/ml



#39 AR-1262-4


R.T.: 8.497 min
Delta R.T.: 0.005 min
Response: 643104
Conc: 5.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.336 min
Delta R.T.: 0.002 min
Response: 49593
Conc: 0.14 ng/ml




#40 AR-1262-5

R.T.: 9.183 min
Delta R.T.: 0.006 min
Response: 246976
Conc: 3.47 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T.: 7.876 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.404 min
Delta R.T.: -0.008 min
Response: 467647
Conc: 1.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.267 min
Delta R.T.: 0.005 min
Response: 826853
Conc: 1.28 ng/ml




#42 AR-1268-2

R.T.: 8.497 min
Delta R.T.: -0.012 min
Response: 643104
Conc: 2.79 ng/ml



#42 AR-1268-2


R.T.: 7.336 min
Delta R.T.: 0.002 min
Response: 49593
Conc: 0.09 ng/ml



#43 AR-1268-3


R.T.: 8.739 min
Delta R.T.: 0.000 min
Response: 162491
Conc: 0.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.543 min
Delta R.T.: -0.007 min
Response: 3322121
Conc: 7.18 ng/ml




#44 AR-1268-4

R.T.: 9.183 min
Delta R.T.: 0.009 min
Response: 246976
Conc: 3.08 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 7.875 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.582 min
Delta R.T.: -0.004 min
Response: 103934
Conc: 0.18 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.164 min
Delta R.T.: -0.008 min
Response: 3899381
Conc: 3.13 ng/ml