

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052022\
 Data File : PP047796.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 May 2022 08:39
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 21 06:39:36 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 21 06:22:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.843	3.110	141.4E6	175.1E6	53.456	49.656
2) SA Decachlor...	9.898	8.407	81349735	149.2E6	53.710	51.200
Target Compounds						
26) L6 AR-1254-1	6.006	5.079	31331562	102.3E6	400.525	548.562 #
27) L6 AR-1254-2	6.241	5.238	64288824	80319468	528.650	487.391
28) L6 AR-1254-3	6.641	5.666	66347882	125.1E6	505.711	488.534
29) L6 AR-1254-4	6.953	5.915	61071651	89887800	641.388	490.897
30) L6 AR-1254-5	7.412	6.362	50844101	122.0E6	521.198	508.134

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052022\
Data File : PP047796.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 May 2022 08:39
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 06:39:36 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 21 06:22:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

