

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052022\
 Data File : PP047816.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 May 2022 16:05
 Operator : YP\AJ
 Sample : N2824-05DL 50X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SAS-1DDL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/23/2022
 Supervised By :Ankita Jodhani 05/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 21 06:53:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 21 06:22:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.839	3.109	1110668	1612447	0.420m	0.457
2) SA Decachlor...	9.892	8.403	1442025	748744	0.952m	0.257m#
Target Compounds						
26) L6 AR-1254-1	6.000	5.077	46591095	86838377	595.594	465.556m
27) L6 AR-1254-2	6.238	5.238	79196228	96379661	651.235	584.847
28) L6 AR-1254-3	6.637	5.665	87977998	173.1E6	670.579	675.630
29) L6 AR-1254-4	6.949	5.913	87790178	144.8E6	921.992	790.869
30) L6 AR-1254-5	7.408	6.360	119.3E6	282.0E6	1223.115	1174.817

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052022\
Data File : PP047816.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 May 2022 16:05
Operator : YP\AJ
Sample : N2824-05DL 50X
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
SAS-1DDL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 06:53:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 21 06:22:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.839 min
Delta R.T.: -0.011 min
Response: 1110668
Conc: 0.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
SAS-1DDL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/23/2022




#1 Tetrachloro-m-xylene

R.T.: 3.109 min
Delta R.T.: -0.007 min
Response: 1612447
Conc: 0.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.892 min
Delta R.T.: -0.006 min
Response: 1442025
Conc: 0.95 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.403 min
Delta R.T.: -0.004 min
Response: 748744
Conc: 0.26 ng/ml m




#26 AR-1254-1

R.T.: 6.000 min
Delta R.T.: -0.013 min
Response: 46591095
Conc: 595.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
SAS-1DDL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/23/2022




#26 AR-1254-1

R.T.: 5.077 min
Delta R.T.: -0.012 min
Response: 86838377
Conc: 465.56 ng/ml m




#27 AR-1254-2

R.T.: 6.238 min
Delta R.T.: -0.013 min
Response: 79196228
Conc: 651.23 ng/ml



#27 AR-1254-2

R.T.: 5.238 min
Delta R.T.: -0.012 min
Response: 96379661
Conc: 584.85 ng/ml




#28 AR-1254-3

R.T.: 6.637 min
Delta R.T.: -0.014 min
Response: 87977998
Conc: 670.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
SAS-1DDL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/23/2022




#28 AR-1254-3

R.T.: 5.665 min
Delta R.T.: -0.012 min
Response: 173074368
Conc: 675.63 ng/ml




#29 AR-1254-4

R.T.: 6.949 min
Delta R.T.: -0.014 min
Response: 87790178
Conc: 921.99 ng/ml



#29 AR-1254-4

R.T.: 5.913 min
Delta R.T.: -0.013 min
Response: 144815419
Conc: 790.87 ng/ml




#30 AR-1254-5

R.T.: 7.408 min
Delta R.T.: -0.014 min
Response: 119317775
Conc: 1223.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
SAS-1DDL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/23/2022



#30 AR-1254-5

R.T.: 6.360 min
Delta R.T.: -0.014 min
Response: 282049480
Conc: 1174.82 ng/ml