

Method Path : P:\HPCHEM1\ECD_P\methods\
 Method File : PP052118.M
 Title : 8080.M
 Last Update : Tue May 22 02:19:00 2018
 Response Via : Initial Calibration

Calibration Files

750 =PP025970.D 1000 =PP025971.D 1500 =PP025972.D
 500 =PP025969.D 200 =PP025968.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	8.347	8.088	7.993	8.333	8.765	8.305	E5	3.60
2) T	3,5-DICHLOROBENZOI...	8.645	8.050	8.154	8.707	9.179	8.547	E5	5.35
3) T	4-Nitrophenol	3.836	3.611	3.584	3.879	4.202	3.822	E5	6.53
4) S	2,4-DCAA	5.619	5.311	5.335	5.657	6.079	5.600	E5	5.55
5) T	DICAMBA	2.438	2.313	2.300	2.430	2.480	2.392	E6	3.37
6) T	MCPP	1.614	1.581	1.602	1.558	1.496	1.570	E6	2.96
7) T	MCPA	2.466	2.376	2.384	2.450	2.573	2.450	E6	3.23
8) T	DICHLORPROP	6.068	5.805	5.788	6.097	6.563	6.064	E5	5.17
9) T	2,4-D	7.794	7.438	7.331	7.881	8.631	7.815	E5	6.55
10) T	Pentachlorophenol	1.040	0.967	0.892	1.078	1.144	1.024	E7	9.55
11) T	2,4,5-TP (SILVEX)	4.189	3.943	3.778	4.224	4.458	4.118	E6	6.40
12) T	2,4,5-T	3.959	3.733	3.577	4.043	4.288	3.920	E6	7.04
13) T	2,4-DB	4.887	4.680	4.743	4.835	5.035	4.836	E5	2.84
14) T	DINOSEB	3.110	2.962	2.847	3.134	3.298	3.070	E6	5.63
15) T	Picloram	6.024	5.730	5.392	6.111	6.368	5.925	E6	6.33
16) T	DCPA	5.276	4.944	4.641	5.403	5.764	5.205	E6	8.28

Signal #2 Calibration Files

750 =PP025970.D 1000 =PP025971.D 1500 =PP025972.D
 500 =PP025969.D 200 =PP025968.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	8.290	7.932	7.763	8.334	8.949	8.254	E5	5.54
2) T	3,5-DICHLOROBENZOI...	7.954	7.653	7.428	8.041	8.667	7.948	E5	5.91
3) T	4-Nitrophenol	3.556	3.413	3.305	3.595	3.947	3.563	E5	6.84
4) S	2,4-DCAA	5.791	5.565	5.465	5.817	6.331	5.794	E5	5.79
5) T	DICAMBA	2.191	2.107	2.045	2.183	2.259	2.157	E6	3.84
6) T	MCPP	1.384	1.336	1.341	1.363	1.381	1.361	E6	1.64
7) T	MCPA	2.064	2.023	2.016	1.992	1.907	2.000	E6	2.91
8) T	DICHLORPROP	5.393	5.206	5.103	5.417	5.920	5.408	E5	5.82
9) T	2,4-D	7.303	7.042	6.870	7.359	7.945	7.304	E5	5.61
10) T	Pentachlorophenol	8.857	8.267	7.576	9.168	9.834	8.740	E6	9.86
11) T	2,4,5-TP (SILVEX)	3.544	3.379	3.177	3.593	3.784	3.495	E6	6.56
12) T	2,4,5-T	2.905	2.781	2.636	2.939	3.082	2.869	E6	5.88
13) T	2,4-DB	4.082	3.961	3.889	4.092	4.411	4.087	E5	4.90
14) T	DINOSEB	2.496	2.395	2.290	2.514	2.642	2.468	E6	5.38
15) T	Picloram	4.798	4.577	4.259	4.883	5.105	4.724	E6	6.81
16) T	DCPA	4.273	4.039	3.765	4.374	4.698	4.230	E6	8.31

(#) = Out of Range