

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052122\
 Data File : PP047829.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2022 12:01
 Operator : YP\AJ
 Sample : N2824-09DL 10X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 CS-7NDL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/23/2022
 Supervised By :Ankita Jodhani 05/24/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 14:47:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 21 06:22:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.840	3.107	6351964	9382319	2.401	2.660
2) SA Decachlor...	9.891	8.403	5560399	4216638	3.671	1.447m#
Target Compounds						
26) L6 AR-1254-1	6.000	5.077	34658655	49994421	443.056	268.029 #
27) L6 AR-1254-2	6.237	5.236	59610127	74899853	490.177	454.504
28) L6 AR-1254-3	6.636	5.664	80947934	157.8E6	616.995	616.069
29) L6 AR-1254-4	6.949	5.912	95385331	148.4E6	1001.758	810.400
30) L6 AR-1254-5	7.407	6.358f	171.2E6	431.3E6	1754.467	1796.375m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052122\
Data File : PP047829.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2022 12:01
Operator : YP\AJ
Sample : N2824-09DL 10X
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
CS-7NDL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/24/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 14:47:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 21 06:22:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.840 min
Delta R.T.: -0.010 min
Response: 6351964
Conc: 2.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
CS-7NDL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/24/2022




#1 Tetrachloro-m-xylene

R.T.: 3.107 min
Delta R.T.: -0.008 min
Response: 9382319
Conc: 2.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.891 min
Delta R.T.: -0.007 min
Response: 5560399
Conc: 3.67 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.403 min
Delta R.T.: -0.004 min
Response: 4216638
Conc: 1.45 ng/ml m




#26 AR-1254-1

R.T.: 6.000 min
Delta R.T.: -0.013 min
Response: 34658655
Conc: 443.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
CS-7NDL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/24/2022




#26 AR-1254-1

R.T.: 5.077 min
Delta R.T.: -0.013 min
Response: 49994421
Conc: 268.03 ng/ml




#27 AR-1254-2

R.T.: 6.237 min
Delta R.T.: -0.014 min
Response: 59610127
Conc: 490.18 ng/ml



#27 AR-1254-2

R.T.: 5.236 min
Delta R.T.: -0.013 min
Response: 74899853
Conc: 454.50 ng/ml




#28 AR-1254-3

R.T.: 6.636 min
Delta R.T.: -0.015 min
Response: 80947934
Conc: 616.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
CS-7NDL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/24/2022




#28 AR-1254-3

R.T.: 5.664 min
Delta R.T.: -0.013 min
Response: 157816730
Conc: 616.07 ng/ml




#29 AR-1254-4

R.T.: 6.949 min
Delta R.T.: -0.015 min
Response: 95385331
Conc: 1001.76 ng/ml



#29 AR-1254-4

R.T.: 5.912 min
Delta R.T.: -0.014 min
Response: 148391715
Conc: 810.40 ng/ml




#30 AR-1254-5

R.T.: 7.407 min
Delta R.T.: -0.015 min
Response: 171152410
Conc: 1754.47 ng/m

Instrument :
ECD_P
ClientSampleId :
CS-7NDL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/24/2022



#30 AR-1254-5

R.T.: 6.358 min
Delta R.T.: -0.015 min
Response: 431272856
Conc: 1796.38 ng/ml m