

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052122\
 Data File : PP047878.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 May 2022 06:28
 Operator : YP\AJ
 Sample : PB145020BSD
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB145020BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 23 01:51:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 21 06:22:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.839	3.106	63187685	78242010	23.887	22.186
2) SA Decachlor...	9.887	8.402	37072969	74177684	24.477	25.453
Target Compounds						
3) L1 AR-1016-1	5.079	4.232	3384957	5756395	45.946m	49.334m
4) L1 AR-1016-2	5.103	4.251	5466757	9399100	48.967	54.224
5) L1 AR-1016-3	5.169	4.434	3491088	4943084	50.899	54.635
6) L1 AR-1016-4	5.275	4.482	2949473	3886222	50.486	53.221
7) L1 AR-1016-5	5.591	4.705	3075563	5365892	56.351	57.106
31) L7 AR-1260-1	6.812	5.801	5072959	9792590	56.079	57.484
32) L7 AR-1260-2	7.095	6.008	6441925	12957252	55.487	53.420
33) L7 AR-1260-3	7.488	6.167	5954507	10550582	57.775	52.794
34) L7 AR-1260-4	7.735	6.675	6189265	8070467	64.050	45.675 #
35) L7 AR-1260-5	8.073	6.944	11146060	20864988	62.309	44.355 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052122\
Data File : PP047878.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 May 2022 06:28
Operator : YP\AJ
Sample : PB145020BSD
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB145020BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 01:51:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 21 06:22:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.839 min
Delta R.T.: -0.011 min
Response: 63187685
Conc: 23.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB145020BSD




#1 Tetrachloro-m-xylene

R.T.: 3.106 min
Delta R.T.: -0.009 min
Response: 78242010
Conc: 22.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.887 min
Delta R.T.: -0.011 min
Response: 37072969
Conc: 24.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.402 min
Delta R.T.: -0.005 min
Response: 74177684
Conc: 25.45 ng/ml



#3 AR-1016-1


R.T.: 5.079 min
Delta R.T.: -0.015 min
Response: 3384957
Conc: 45.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB145020BSD




#3 AR-1016-1

R.T.: 4.232 min
Delta R.T.: -0.013 min
Response: 5756395
Conc: 49.33 ng/ml m




#4 AR-1016-2

R.T.: 5.103 min
Delta R.T.: -0.014 min
Response: 5466757
Conc: 48.97 ng/ml



#4 AR-1016-2


R.T.: 4.251 min
Delta R.T.: -0.012 min
Response: 9399100
Conc: 54.22 ng/ml



#5 AR-1016-3


R.T.: 5.169 min
Delta R.T.: 0.001 min
Response: 3491088
Conc: 50.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB145020BSD




#5 AR-1016-3

R.T.: 4.434 min
Delta R.T.: -0.013 min
Response: 4943084
Conc: 54.64 ng/ml




#6 AR-1016-4

R.T.: 5.275 min
Delta R.T.: -0.013 min
Response: 2949473
Conc: 50.49 ng/ml



#6 AR-1016-4


R.T.: 4.482 min
Delta R.T.: -0.013 min
Response: 3886222
Conc: 53.22 ng/ml



#7 AR-1016-5


R.T.: 5.591 min
Delta R.T.: 0.000 min
Response: 3075563
Conc: 56.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB145020BSD




#7 AR-1016-5

R.T.: 4.705 min
Delta R.T.: -0.013 min
Response: 5365892
Conc: 57.11 ng/ml




#31 AR-1260-1

R.T.: 6.812 min
Delta R.T.: -0.002 min
Response: 5072959
Conc: 56.08 ng/ml



#31 AR-1260-1


R.T.: 5.801 min
Delta R.T.: 0.000 min
Response: 9792590
Conc: 57.48 ng/ml



#32 AR-1260-2


R.T.: 7.095 min
Delta R.T.: 0.000 min
Response: 6441925
Conc: 55.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB145020BSD




#32 AR-1260-2

R.T.: 6.008 min
Delta R.T.: 0.000 min
Response: 12957252
Conc: 53.42 ng/ml




#33 AR-1260-3

R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 5954507
Conc: 57.77 ng/ml



#33 AR-1260-3


R.T.: 6.167 min
Delta R.T.: 0.000 min
Response: 10550582
Conc: 52.79 ng/ml



#34 AR-1260-4


R.T.: 7.735 min
Delta R.T.: 0.000 min
Response: 6189265
Conc: 64.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB145020BSD




#34 AR-1260-4

R.T.: 6.675 min
Delta R.T.: 0.000 min
Response: 8070467
Conc: 45.68 ng/ml



#35 AR-1260-5

R.T.: 8.073 min
Delta R.T.: -0.003 min
Response: 11146060
Conc: 62.31 ng/ml



#35 AR-1260-5

R.T.: 6.944 min
Delta R.T.: -0.001 min
Response: 20864988
Conc: 44.36 ng/ml