

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052124\
 Data File : PP065121.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2024 20:26
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/22/2024
 Supervised By :Ankita Jodhani 05/22/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 01:22:16 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 21 18:19:41 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.470	3.755	82690827	77174627	52.392	50.836
2) SA Decachlor...	10.209	8.834	71780918	70118751	55.203	53.197
Target Compounds						
26) L6 AR-1254-1	6.488	5.668	31231492	39650987	505.844	497.016
27) L6 AR-1254-2	6.707	5.815	46168174	35044042	503.240	496.941
28) L6 AR-1254-3	7.072	6.224	47130174	55156175	506.656	500.457
29) L6 AR-1254-4	7.357	6.452	30996506	28981786	511.313	513.368
30) L6 AR-1254-5	7.774	6.873	36607054	47819422	521.101m	513.195

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052124\
Data File : PP065121.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2024 20:26
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 39 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/22/2024
Supervised By :Ankita Jodhani 05/22/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 01:22:16 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052124.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 21 18:19:41 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.470 min
Delta R.T.: 0.002 min
Response: 82690827
Conc: 52.39 ng/ml

Instrument :

ECD_P


ClientSampleId :

AR1254CCC500

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 05/22/2024
Supervised By :Ankita Jodhani 05/22/2024




#1 Tetrachloro-m-xylene

R.T.: 3.755 min
Delta R.T.: 0.000 min
Response: 77174627
Conc: 50.84 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.209 min
Delta R.T.: 0.001 min
Response: 71780918
Conc: 55.20 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.834 min
Delta R.T.: -0.002 min
Response: 70118751
Conc: 53.20 ng/ml




#26 AR-1254-1

R.T.: 6.488 min
Delta R.T.: 0.000 min
Response: 31231492
Conc: 505.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/22/2024
Supervised By :Ankita Jodhani 05/22/2024




#26 AR-1254-1

R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 39650987
Conc: 497.02 ng/ml




#27 AR-1254-2

R.T.: 6.707 min
Delta R.T.: 0.000 min
Response: 46168174
Conc: 503.24 ng/ml



#27 AR-1254-2

R.T.: 5.815 min
Delta R.T.: 0.000 min
Response: 35044042
Conc: 496.94 ng/ml




#28 AR-1254-3

R.T.: 7.072 min
Delta R.T.: 0.000 min
Response: 47130174
Conc: 506.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/22/2024
Supervised By :Ankita Jodhani 05/22/2024




#28 AR-1254-3

R.T.: 6.224 min
Delta R.T.: 0.000 min
Response: 55156175
Conc: 500.46 ng/ml




#29 AR-1254-4

R.T.: 7.357 min
Delta R.T.: 0.000 min
Response: 30996506
Conc: 511.31 ng/ml



#29 AR-1254-4

R.T.: 6.452 min
Delta R.T.: 0.000 min
Response: 28981786
Conc: 513.37 ng/ml




#30 AR-1254-5

R.T.: 7.774 min
Delta R.T.: 0.000 min
Response: 36607054
Conc: 521.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/22/2024
Supervised By :Ankita Jodhani 05/22/2024



#30 AR-1254-5

R.T.: 6.873 min
Delta R.T.: 0.000 min
Response: 47819422
Conc: 513.19 ng/ml