

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052225\
 Data File : PP072304.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 May 2025 16:56
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 22:40:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:29:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.797	105.8E6	89694184	52.611	56.116
2) SA Decachlor...	10.206	8.820	89885918	60033562	55.174	56.710
Target Compounds						
26) L6 AR-1254-1	6.503	5.683	43739665	53001182	508.581	530.254
27) L6 AR-1254-2	6.719	5.831	67466212	44964037	516.281	520.978
28) L6 AR-1254-3	7.082	6.234	69343067	69150684	525.314	537.633
29) L6 AR-1254-4	7.364	6.462	63733182	44776001	488.337	524.946
30) L6 AR-1254-5	7.781	6.880	59640203	61281469	536.108	540.632



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052225\
Data File : PP072304.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 May 2025 16:56
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 22:40:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.501 min
Delta R.T.: 0.003 min
Response: 105790408
Conc: 52.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.006 min
Response: 89694184
Conc: 56.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.206 min
Delta R.T.: -0.008 min
Response: 89885918
Conc: 55.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.820 min
Delta R.T.: -0.002 min
Response: 60033562
Conc: 56.71 ng/ml



#26 AR-1254-1


R.T.: 6.503 min
Delta R.T.: -0.005 min
Response: 43739665
Conc: 508.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.683 min
Delta R.T.: -0.003 min
Response: 53001182
Conc: 530.25 ng/ml




#27 AR-1254-2

R.T.: 6.719 min
Delta R.T.: -0.005 min
Response: 67466212
Conc: 516.28 ng/ml



#27 AR-1254-2


R.T.: 5.831 min
Delta R.T.: -0.003 min
Response: 44964037
Conc: 520.98 ng/ml



#28 AR-1254-3


R.T.: 7.082 min
Delta R.T.: -0.006 min
Response: 69343067
Conc: 525.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.234 min
Delta R.T.: -0.002 min
Response: 69150684
Conc: 537.63 ng/ml




#29 AR-1254-4

R.T.: 7.364 min
Delta R.T.: -0.006 min
Response: 63733182
Conc: 488.34 ng/ml



#29 AR-1254-4


R.T.: 6.462 min
Delta R.T.: -0.003 min
Response: 44776001
Conc: 524.95 ng/ml



#30 AR-1254-5

R.T.: 7.781 min
Delta R.T.: -0.006 min
Response: 59640203
Conc: 536.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.880 min
Delta R.T.: -0.003 min
Response: 61281469
Conc: 540.63 ng/ml