

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052422\
 Data File : PP047981.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 May 2022 21:23
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/09/2022
 Supervised By :Ankita Jodhani 06/09/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 25 05:18:53 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 23 09:44:20 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.839	3.107	134.6E6	161.2E6	50.876	45.714
2) SA Decachlor...	9.882	8.397	79546945	141.2E6	52.520	48.462
Target Compounds						
3) L1 AR-1016-1	5.079	4.231	39118197	57365833	530.977	491.640m
4) L1 AR-1016-2	5.102	4.249	60089789	85851319	538.243	495.283
5) L1 AR-1016-3	5.167	4.432	35106463	44828289	511.845	495.480
6) L1 AR-1016-4	5.273	4.480	29825362	36229600	510.518	496.155
7) L1 AR-1016-5	5.590	4.702	28939918	49524554	530.243	527.065
31) L7 AR-1260-1	6.811	5.797	46024358	87449985	508.772	513.341
32) L7 AR-1260-2	7.092	6.004	62470871	120.0E6	538.084	494.758
33) L7 AR-1260-3	7.485	6.164	51959044	99252516	504.143	496.652
34) L7 AR-1260-4	7.732	6.672	48232670	85848549	499.139	485.863
35) L7 AR-1260-5	8.070	6.940	102.9E6	225.2E6	575.109	478.720

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052422\
Data File : PP047981.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 May 2022 21:23
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/09/2022
Supervised By :Ankita Jodhani 06/09/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 25 05:18:53 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051022.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 23 09:44:20 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

