

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052423\
 Data File : PP057851.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 May 2023 20:17
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 24 21:27:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.384	3.622	111.7E6	116.1E6	50.925	51.190
2)	SA Decachlor...	10.209	8.681	61286348	66978755	54.080	42.669
Target Compounds							
3)	L1 AR-1016-1	5.563	4.718	807306	1031118	12.959	13.837
4)	L1 AR-1016-2	5.592	4.737	1111488	730068	12.363	7.063 #
5)	L1 AR-1016-3	5.649	4.916	403537	577694	7.137	9.799 #
6)	L1 AR-1016-4	5.747	4.957	442074	22855680	9.572	485.889 #
7)	L1 AR-1016-5	6.045	5.173	11115075	14914122	238.332	251.181
8)	L2 AR-1221-1	4.599	3.835	34443	75523	1.190	2.561 #
9)	L2 AR-1221-2	4.693	3.925	1782104	1919583	82.225	86.177
10)	L2 AR-1221-3	4.756	4.002	211712	277891	3.514	4.362
11)	L3 AR-1232-1	4.756	4.002	211712	277891	4.834	6.022
12)	L3 AR-1232-2	5.294	4.737	393775	730068	15.099	15.428
13)	L3 AR-1232-3	5.592	4.916	1111488	577694	27.234	21.224
14)	L3 AR-1232-4	5.747	4.999	442074	7469009	20.760	306.376 #
15)	L3 AR-1232-5	5.840	5.173	19456031	14914122	1065.334	539.346 #
16)	L4 AR-1242-1	5.563	4.718	807306	1031118	16.947	17.713
17)	L4 AR-1242-2	5.592	4.737	1111488	730068	16.253	9.147 #
18)	L4 AR-1242-3	5.649	4.916	403537	577694	9.268	12.318 #
19)	L4 AR-1242-4	5.747	4.999	442074	7469009	12.546	160.969 #
20)	L4 AR-1242-5	6.496	5.528	11138545	53274162	343.769	952.861 #
21)	L5 AR-1248-1	5.563	4.718	807306	1031118	19.184	20.038
22)	L5 AR-1248-2	5.840	4.957	19456031	22855680	298.403	318.311
23)	L5 AR-1248-3	6.045	4.999	11115075	7469009	160.063	98.720 #
24)	L5 AR-1248-4	6.452	5.173	14246065	14914122	222.905	167.196
25)	L5 AR-1248-5	6.496	5.570	11138545	6729820	173.608	80.944 #
26)	L6 AR-1254-1	6.427	5.528	33255551	53274162	495.480	477.113
27)	L6 AR-1254-2	6.648	5.677	51388225	47308944	543.226	474.021
28)	L6 AR-1254-3	7.017	6.084	54805822	70488198	601.677	450.540 #
29)	L6 AR-1254-4	7.305	6.315	36709235	40841078	616.319	442.712 #
30)	L6 AR-1254-5	7.728	6.736	40119775	62068965	584.394	428.772 #
31)	L7 AR-1260-1	7.183	6.215	27993561	36508476	354.813	332.064
32)	L7 AR-1260-2	7.443	6.405	26789686	29168123	334.377	229.499 #
33)	L7 AR-1260-3	7.789f	6.559	7770311	29546451	136.382	240.957 #
34)	L7 AR-1260-4	8.019f	7.035	15311481	3337965	230.444	38.200 #
35)	L7 AR-1260-5	8.355	7.277	5661117	6956368	52.721	37.662 #
36)	L8 AR-1262-1	7.789f	6.826	7770311	2945555	90.235	46.360 #
37)	L8 AR-1262-2	8.355	7.035	5661117	3337965	46.504	26.925 #
38)	L8 AR-1262-3	8.687	7.561	4333074	413486	49.839	4.407 #
39)	L8 AR-1262-4	8.798f	7.626	100880	7384532	1.505	43.348 #
40)	L8 AR-1262-5	9.429	8.126	168758	839012	4.408	11.400 #
41)	L9 AR-1268-1	8.687	7.561	4333074	413486	27.829	1.571 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052423\
 Data File : PP057851.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 May 2023 20:17
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 24 21:27:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.798f	7.626	100880	7384532	0.773	30.162 #
43) L9	AR-1268-3	9.013	7.837	61339	58358	0.475	0.288 #
44) L9	AR-1268-4	9.455	8.126	65045	839012	1.536	10.131 #
45) L9	AR-1268-5	9.868	8.422	297506	362290	0.910	0.701



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052423\
Data File : PP057851.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 May 2023 20:17
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 21:27:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.384 min
Delta R.T.: -0.006 min
Response: 111721524
Conc: 50.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: -0.002 min
Response: 116116904
Conc: 51.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.209 min
Delta R.T.: -0.014 min
Response: 61286348
Conc: 54.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.681 min
Delta R.T.: -0.005 min
Response: 66978755
Conc: 42.67 ng/ml



#3 AR-1016-1


R.T.: 5.563 min
Delta R.T.: -0.007 min
Response: 807306
Conc: 12.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.718 min
Delta R.T.: -0.003 min
Response: 1031118
Conc: 13.84 ng/ml




#4 AR-1016-2

R.T.: 5.592 min
Delta R.T.: 0.000 min
Response: 1111488
Conc: 12.36 ng/ml



#4 AR-1016-2


R.T.: 4.737 min
Delta R.T.: -0.003 min
Response: 730068
Conc: 7.06 ng/ml



#5 AR-1016-3


R.T.: 5.649 min
Delta R.T.: -0.007 min
Response: 403537
Conc: 7.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.916 min
Delta R.T.: -0.002 min
Response: 577694
Conc: 9.80 ng/ml




#6 AR-1016-4

R.T.: 5.747 min
Delta R.T.: -0.008 min
Response: 442074
Conc: 9.57 ng/ml



#6 AR-1016-4


R.T.: 4.957 min
Delta R.T.: -0.003 min
Response: 22855680
Conc: 485.89 ng/ml



#7 AR-1016-5


R.T.: 6.045 min
Delta R.T.: -0.007 min
Response: 11115075
Conc: 238.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.173 min
Delta R.T.: -0.004 min
Response: 14914122
Conc: 251.18 ng/ml




#8 AR-1221-1

R.T.: 4.599 min
Delta R.T.: 0.000 min
Response: 34443
Conc: 1.19 ng/ml



#8 AR-1221-1


R.T.: 3.835 min
Delta R.T.: -0.003 min
Response: 75523
Conc: 2.56 ng/ml



#9 AR-1221-2


R.T.: 4.693 min
Delta R.T.: 0.007 min
Response: 1782104
Conc: 82.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.925 min
Delta R.T.: 0.000 min
Response: 1919583
Conc: 86.18 ng/ml




#10 AR-1221-3

R.T.: 4.756 min
Delta R.T.: -0.006 min
Response: 211712
Conc: 3.51 ng/ml



#10 AR-1221-3


R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 277891
Conc: 4.36 ng/ml



#11 AR-1232-1


R.T.: 4.756 min
Delta R.T.: -0.006 min
Response: 211712
Conc: 4.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.002 min
Delta R.T.: 0.000 min
Response: 277891
Conc: 6.02 ng/ml




#12 AR-1232-2

R.T.: 5.294 min
Delta R.T.: -0.004 min
Response: 393775
Conc: 15.10 ng/ml



#12 AR-1232-2


R.T.: 4.737 min
Delta R.T.: -0.002 min
Response: 730068
Conc: 15.43 ng/ml



#13 AR-1232-3


R.T.: 5.592 min
Delta R.T.: 0.001 min
Response: 1111488
Conc: 27.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.916 min
Delta R.T.: -0.001 min
Response: 577694
Conc: 21.22 ng/ml




#14 AR-1232-4

R.T.: 5.747 min
Delta R.T.: -0.006 min
Response: 442074
Conc: 20.76 ng/ml



#14 AR-1232-4


R.T.: 4.999 min
Delta R.T.: -0.003 min
Response: 7469009
Conc: 306.38 ng/ml



#15 AR-1232-5


R.T.: 5.840 min
Delta R.T.: -0.006 min
Response: 19456031
Conc: 1065.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.173 min
Delta R.T.: -0.003 min
Response: 14914122
Conc: 539.35 ng/ml




#16 AR-1242-1

R.T.: 5.563 min
Delta R.T.: -0.005 min
Response: 807306
Conc: 16.95 ng/ml



#16 AR-1242-1


R.T.: 4.718 min
Delta R.T.: -0.002 min
Response: 1031118
Conc: 17.71 ng/ml



#17 AR-1242-2


R.T.: 5.592 min
Delta R.T.: 0.002 min
Response: 1111488
Conc: 16.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.737 min
Delta R.T.: -0.002 min
Response: 730068
Conc: 9.15 ng/ml




#18 AR-1242-3

R.T.: 5.649 min
Delta R.T.: -0.005 min
Response: 403537
Conc: 9.27 ng/ml



#18 AR-1242-3


R.T.: 4.916 min
Delta R.T.: 0.000 min
Response: 577694
Conc: 12.32 ng/ml



#19 AR-1242-4


R.T.: 5.747 min
Delta R.T.: -0.006 min
Response: 442074
Conc: 12.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.999 min
Delta R.T.: -0.003 min
Response: 7469009
Conc: 160.97 ng/ml




#20 AR-1242-5

R.T.: 6.496 min
Delta R.T.: -0.002 min
Response: 11138545
Conc: 343.77 ng/ml



#20 AR-1242-5


R.T.: 5.528 min
Delta R.T.: -0.002 min
Response: 53274162
Conc: 952.86 ng/ml



#21 AR-1248-1


R.T.: 5.563 min
Delta R.T.: -0.006 min
Response: 807306
Conc: 19.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.718 min
Delta R.T.: -0.002 min
Response: 1031118
Conc: 20.04 ng/ml




#22 AR-1248-2

R.T.: 5.840 min
Delta R.T.: -0.006 min
Response: 19456031
Conc: 298.40 ng/ml



#22 AR-1248-2


R.T.: 4.957 min
Delta R.T.: -0.003 min
Response: 22855680
Conc: 318.31 ng/ml



#23 AR-1248-3


R.T.: 6.045 min
Delta R.T.: -0.006 min
Response: 11115075
Conc: 160.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.999 min
Delta R.T.: -0.003 min
Response: 7469009
Conc: 98.72 ng/ml




#24 AR-1248-4

R.T.: 6.452 min
Delta R.T.: -0.006 min
Response: 14246065
Conc: 222.91 ng/ml



#24 AR-1248-4


R.T.: 5.173 min
Delta R.T.: -0.003 min
Response: 14914122
Conc: 167.20 ng/ml



#25 AR-1248-5


R.T.: 6.496 min
Delta R.T.: -0.003 min
Response: 11138545
Conc: 173.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.570 min
Delta R.T.: -0.003 min
Response: 6729820
Conc: 80.94 ng/ml




#26 AR-1254-1

R.T.: 6.427 min
Delta R.T.: -0.006 min
Response: 33255551
Conc: 495.48 ng/ml



#26 AR-1254-1


R.T.: 5.528 min
Delta R.T.: -0.003 min
Response: 53274162
Conc: 477.11 ng/ml



#27 AR-1254-2


R.T.: 6.648 min
Delta R.T.: -0.006 min
Response: 51388225
Conc: 543.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.677 min
Delta R.T.: -0.003 min
Response: 47308944
Conc: 474.02 ng/ml




#28 AR-1254-3

R.T.: 7.017 min
Delta R.T.: -0.006 min
Response: 54805822
Conc: 601.68 ng/ml



#28 AR-1254-3


R.T.: 6.084 min
Delta R.T.: -0.004 min
Response: 70488198
Conc: 450.54 ng/ml



#29 AR-1254-4


R.T.: 7.305 min
Delta R.T.: -0.007 min
Response: 36709235
Conc: 616.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.315 min
Delta R.T.: -0.003 min
Response: 40841078
Conc: 442.71 ng/ml




#30 AR-1254-5

R.T.: 7.728 min
Delta R.T.: -0.007 min
Response: 40119775
Conc: 584.39 ng/ml



#30 AR-1254-5


R.T.: 6.736 min
Delta R.T.: -0.004 min
Response: 62068965
Conc: 428.77 ng/ml



#31 AR-1260-1


R.T.: 7.183 min
Delta R.T.: -0.008 min
Response: 27993561
Conc: 354.81 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.215 min
Delta R.T.: -0.004 min
Response: 36508476
Conc: 332.06 ng/ml




#32 AR-1260-2

R.T.: 7.443 min
Delta R.T.: -0.006 min
Response: 26789686
Conc: 334.38 ng/ml



#32 AR-1260-2


R.T.: 6.405 min
Delta R.T.: -0.004 min
Response: 29168123
Conc: 229.50 ng/ml



#33 AR-1260-3


R.T.: 7.789 min
Delta R.T.: -0.024 min
Response: 7770311
Conc: 136.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.559 min
Delta R.T.: -0.005 min
Response: 29546451
Conc: 240.96 ng/ml




#34 AR-1260-4

R.T.: 8.019 min
Delta R.T.: -0.021 min
Response: 15311481
Conc: 230.44 ng/ml



#34 AR-1260-4


R.T.: 7.035 min
Delta R.T.: -0.005 min
Response: 3337965
Conc: 38.20 ng/ml



#35 AR-1260-5


R.T.: 8.355 min
Delta R.T.: -0.009 min
Response: 5661117
Conc: 52.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.277 min
Delta R.T.: -0.005 min
Response: 6956368
Conc: 37.66 ng/ml




#36 AR-1262-1

R.T.: 7.789 min
Delta R.T.: -0.023 min
Response: 7770311
Conc: 90.23 ng/ml



#36 AR-1262-1


R.T.: 6.826 min
Delta R.T.: -0.005 min
Response: 2945555
Conc: 46.36 ng/ml



#37 AR-1262-2


R.T.: 8.355 min
Delta R.T.: -0.007 min
Response: 5661117
Conc: 46.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.035 min
Delta R.T.: -0.004 min
Response: 3337965
Conc: 26.93 ng/ml




#38 AR-1262-3

R.T.: 8.687 min
Delta R.T.: 0.006 min
Response: 4333074
Conc: 49.84 ng/ml



#38 AR-1262-3


R.T.: 7.561 min
Delta R.T.: -0.006 min
Response: 413486
Conc: 4.41 ng/ml



#39 AR-1262-4


R.T.: 8.798 min
Delta R.T.: 0.025 min
Response: 100880
Conc: 1.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.626 min
Delta R.T.: -0.006 min
Response: 7384532
Conc: 43.35 ng/ml




#40 AR-1262-5

R.T.: 9.429 min
Delta R.T.: -0.013 min
Response: 168758
Conc: 4.41 ng/ml



#40 AR-1262-5


R.T.: 8.126 min
Delta R.T.: -0.004 min
Response: 839012
Conc: 11.40 ng/ml



#41 AR-1268-1


R.T.: 8.687 min
Delta R.T.: 0.007 min
Response: 4333074
Conc: 27.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.561 min
Delta R.T.: -0.007 min
Response: 413486
Conc: 1.57 ng/ml




#42 AR-1268-2

R.T.: 8.798 min
Delta R.T.: 0.022 min
Response: 100880
Conc: 0.77 ng/ml



#42 AR-1268-2


R.T.: 7.626 min
Delta R.T.: -0.007 min
Response: 7384532
Conc: 30.16 ng/ml



#43 AR-1268-3


R.T.: 9.013 min
Delta R.T.: 0.003 min
Response: 61339
Conc: 0.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.837 min
Delta R.T.: -0.005 min
Response: 58358
Conc: 0.29 ng/ml




#44 AR-1268-4

R.T.: 9.455 min
Delta R.T.: 0.011 min
Response: 65045
Conc: 1.54 ng/ml



#44 AR-1268-4


R.T.: 8.126 min
Delta R.T.: -0.005 min
Response: 839012
Conc: 10.13 ng/ml



#45 AR-1268-5

R.T.: 9.868 min
Delta R.T.: -0.003 min
Response: 297506
Conc: 0.91 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.422 min
Delta R.T.: -0.006 min
Response: 362290
Conc: 0.70 ng/ml