

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052423\
 Data File : PP057866.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 May 2023 01:07
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 25 05:48:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.384	3.622	111.0E6	117.1E6	50.589	51.628
2) SA Decachlor...	10.210	8.681	60451419	64682795	53.344	41.206
Target Compounds						
26) L6 AR-1254-1	6.426	5.528	34158921	52310474	508.939	468.483
27) L6 AR-1254-2	6.647	5.677	50877203	46241691	537.824	463.327
28) L6 AR-1254-3	7.016	6.085	49068907	69904286	538.695	446.808
29) L6 AR-1254-4	7.305	6.314	31065726	39989922	521.569	433.486
30) L6 AR-1254-5	7.728	6.736	38058522	60890995	554.369	420.635



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052423\
Data File : PP057866.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 May 2023 01:07
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 25 05:48:30 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.384 min
Delta R.T.: -0.006 min
Response: 110985194
Conc: 50.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: -0.001 min
Response: 117110332
Conc: 51.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.210 min
Delta R.T.: -0.013 min
Response: 60451419
Conc: 53.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.681 min
Delta R.T.: -0.005 min
Response: 64682795
Conc: 41.21 ng/ml



#26 AR-1254-1


R.T.: 6.426 min
Delta R.T.: -0.007 min
Response: 34158921
Conc: 508.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.528 min
Delta R.T.: -0.003 min
Response: 52310474
Conc: 468.48 ng/ml




#27 AR-1254-2

R.T.: 6.647 min
Delta R.T.: -0.007 min
Response: 50877203
Conc: 537.82 ng/ml



#27 AR-1254-2


R.T.: 5.677 min
Delta R.T.: -0.003 min
Response: 46241691
Conc: 463.33 ng/ml



#28 AR-1254-3


R.T.: 7.016 min
Delta R.T.: -0.007 min
Response: 49068907
Conc: 538.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.085 min
Delta R.T.: -0.004 min
Response: 69904286
Conc: 446.81 ng/ml




#29 AR-1254-4

R.T.: 7.305 min
Delta R.T.: -0.007 min
Response: 31065726
Conc: 521.57 ng/ml



#29 AR-1254-4


R.T.: 6.314 min
Delta R.T.: -0.004 min
Response: 39989922
Conc: 433.49 ng/ml



#30 AR-1254-5

R.T.: 7.728 min
Delta R.T.: -0.007 min
Response: 38058522
Conc: 554.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.736 min
Delta R.T.: -0.004 min
Response: 60890995
Conc: 420.63 ng/ml