

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052423\
 Data File : PP057896.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 May 2023 10:05
 Operator : YP\AJ
 Sample : 02895-11
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 4B

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/25/2023
 Supervised By :Ankita Jodhani 05/25/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 25 11:00:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.382	3.621	38248700	41342318	17.435	18.226
2) SA Decachlor...	10.207	8.679	16979500	21973178	14.983	13.998

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052423\
Data File : PP057896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 May 2023 10:05
Operator : YP\AJ
Sample : 02895-11
Misc :
ALS Vial : 47 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
4B


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/25/2023
Supervised By :Ankita Jodhani 05/25/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 25 11:00:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.382 min
Delta R.T.: -0.008 min
Response: 38248700
Conc: 17.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
4B


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/25/2023
Supervised By :Ankita Jodhani 05/25/2023




#1 Tetrachloro-m-xylene

R.T.: 3.621 min
Delta R.T.: -0.003 min
Response: 41342318
Conc: 18.23 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.207 min
Delta R.T.: -0.017 min
Response: 16979500
Conc: 14.98 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.679 min
Delta R.T.: -0.007 min
Response: 21973178
Conc: 14.00 ng/ml