

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052521\
 Data File : PP036107.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 May 2021 20:46
 Operator : DD\AJ
 Sample : PP052521ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 00:50:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 26 00:48:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	44.509	44.334 E3	0.4	100	0.00
2 SA Decachlorobiphenyl	42.737	41.085 E3	3.9	98	0.00
3 L1 AR-1016-1	1.603	1.562 E3	2.6	99	0.00
4 L1 AR-1016-2	2.266	2.230 E3	1.6	100	0.00
5 L1 AR-1016-3	1.448	1.448 E3	0.0	102	0.00
6 L1 AR-1016-4	1.211	1.193 E3	1.5	100	0.00
7 L1 AR-1016-5	1.184	1.177 E3	0.6	99	0.00
31 L7 AR-1260-1	2.026	1.969 E3	2.8	99	0.00
32 L7 AR-1260-2	2.443	2.363 E3	3.3	99	0.00
33 L7 AR-1260-3	1.857	1.776 E3	4.4	98	0.00
34 L7 AR-1260-4	2.233	2.061 E3	7.7	95	0.00
35 L7 AR-1260-5	4.270	4.345 E3	-1.8	105	0.00

Signal #2

1 SA Tetrachloro-m-xylene	27.909	28.326 E3	-1.5	100	0.00
2 SA Decachlorobiphenyl	22.942	23.043 E3	-0.4	99	0.00
3 L1 AR-1016-1	931.461	949.611	-1.9	100	0.00
4 L1 AR-1016-2	1.306	1.310 E3	-0.3	99	0.00
5 L1 AR-1016-3	707.124	729.348	-3.1	101	0.00
6 L1 AR-1016-4	531.215	547.356	-3.0	101	0.00
7 L1 AR-1016-5	685.136	703.137	-2.6	99	0.00
31 L7 AR-1260-1	1.248	1.236 E3	1.0	100	0.00
32 L7 AR-1260-2	1.507	1.497 E3	0.7	100	0.00
33 L7 AR-1260-3	1.405	1.404 E3	0.1	100	0.00
34 L7 AR-1260-4	1.169	1.167 E3	0.2	101	0.00
35 L7 AR-1260-5	2.768	2.803 E3	-1.3	100	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	549.558	0.000	100.0#	0#	-5.21#
9 L2 AR-1221-2	416.829	0.000	100.0#	0#	-5.31#
10 L2 AR-1221-3	1.255	0.000 E3	100.0#	0#	-5.40#
11 L3 AR-1232-1	966.277	0.000	100.0#	0#	-5.40#
12 L3 AR-1232-2	502.507	0.000	100.0#	0#	-5.99#
13 L3 AR-1232-3	931.984	0.000	100.0#	0#	-6.31#
14 L3 AR-1232-4	483.490	0.000	100.0#	0#	-6.48#
15 L3 AR-1232-5	321.425	0.000	100.0#	0#	-6.58#
16 L4 AR-1242-1	1.204	0.000 E3	100.0#	0#	-6.28#
17 L4 AR-1242-2	1.684	0.000 E3	100.0#	0#	-6.31#
18 L4 AR-1242-3	1.098	0.000 E3	100.0#	0#	-6.37#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052521\
 Data File : PP036107.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 May 2021 20:46
 Operator : DD\AJ
 Sample : PP052521ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 00:50:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 26 00:48:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	913.208	0.000	100.0#	0#	-6.48#
20 L4 AR-1242-5	953.293	0.000	100.0#	0#	-7.26#
21 L5 AR-1248-1	930.982	0.000	100.0#	0#	-6.28#
22 L5 AR-1248-2	1.196	0.000 E3	100.0#	0#	-6.58#
23 L5 AR-1248-3	1.449	0.000 E3	100.0#	0#	-6.80#
24 L5 AR-1248-4	1.593	0.000 E3	100.0#	0#	-7.22#
25 L5 AR-1248-5	1.561	0.000 E3	100.0#	0#	-7.26#
26 L6 AR-1254-1	1.620	0.000 E3	100.0#	0#	-7.19#
27 L6 AR-1254-2	2.463	0.000 E3	100.0#	0#	-7.42#
28 L6 AR-1254-3	2.638	0.000 E3	100.0#	0#	-7.80#
29 L6 AR-1254-4	2.018	0.000 E3	100.0#	0#	-8.10#
30 L6 AR-1254-5	2.269	0.000 E3	100.0#	0#	-8.53#
36 L8 AR-1262-1	2.521	0.000 E3	100.0#	0#	-8.60#
37 L8 AR-1262-2	4.506	0.000 E3	100.0#	0#	-9.16#
38 L8 AR-1262-3	3.227	0.000 E3	100.0#	0#	-9.47#
39 L8 AR-1262-4	2.682	0.000 E3	100.0#	0#	-9.56#
40 L8 AR-1262-5	1.781	0.000 E3	100.0#	0#	-10.23#
41 L9 AR-1268-1	5.178	0.000 E3	100.0#	0#	-9.47#
42 L9 AR-1268-2	4.998	0.000 E3	100.0#	0#	-9.57#
43 L9 AR-1268-3	4.172	0.000 E3	100.0#	0#	-9.79#
44 L9 AR-1268-4	1.951	0.000 E3	100.0#	0#	-10.24#
45 L9 AR-1268-5	13.512	0.000 E3	100.0#	0#	-10.65#

Signal #2

8 L2 AR-1221-1	317.696	0.000	100.0#	0#	-4.22#
9 L2 AR-1221-2	240.757	0.000	100.0#	0#	-4.32#
10 L2 AR-1221-3	771.825	0.000	100.0#	0#	-4.41#
11 L3 AR-1232-1	584.950	0.000	100.0#	0#	-4.41#
12 L3 AR-1232-2	524.224	0.000	100.0#	0#	-5.24#
13 L3 AR-1232-3	277.027	0.000	100.0#	0#	-5.43#
14 L3 AR-1232-4	226.909	0.000	100.0#	0#	-5.53#
15 L3 AR-1232-5	249.220	0.000	100.0#	0#	-5.71#
16 L4 AR-1242-1	719.990	0.000	100.0#	0#	-5.22#
17 L4 AR-1242-2	978.283	0.000	100.0#	0#	-5.24#
18 L4 AR-1242-3	536.376	0.000	100.0#	0#	-5.43#
19 L4 AR-1242-4	493.497	0.000	100.0#	0#	-5.53#
20 L4 AR-1242-5	647.129	0.000	100.0#	0#	-6.09#
21 L5 AR-1248-1	536.456	0.000	100.0#	0#	-5.22#
22 L5 AR-1248-2	697.836	0.000	100.0#	0#	-5.48#
23 L5 AR-1248-3	719.701	0.000	100.0#	0#	-5.53#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052521\
 Data File : PP036107.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 May 2021 20:46
 Operator : DD\AJ
 Sample : PP052521ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 00:50:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 26 00:48:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	883.194	0.000	100.0#	0#	-5.71#
25 L5 AR-1248-5	879.383	0.000	100.0#	0#	-6.14#
26 L6 AR-1254-1	1.307	0.000 E3	100.0#	0#	-6.09#
27 L6 AR-1254-2	1.153	0.000 E3	100.0#	0#	-6.25#
28 L6 AR-1254-3	1.904	0.000 E3	100.0#	0#	-6.67#
29 L6 AR-1254-4	1.194	0.000 E3	100.0#	0#	-6.91#
30 L6 AR-1254-5	1.706	0.000 E3	100.0#	0#	-7.34#
36 L8 AR-1262-1	904.060	0.000	100.0#	0#	-7.34#
37 L8 AR-1262-2	2.977	0.000 E3	100.0#	0#	-7.89#
38 L8 AR-1262-3	1.188	0.000 E3	100.0#	0#	-8.18#
39 L8 AR-1262-4	2.205	0.000 E3	100.0#	0#	-8.25#
40 L8 AR-1262-5	1.017	0.000 E3	100.0#	0#	-8.74#
41 L9 AR-1268-1	3.396	0.000 E3	100.0#	0#	-8.18#
42 L9 AR-1268-2	3.088	0.000 E3	100.0#	0#	-8.25#
43 L9 AR-1268-3	2.573	0.000 E3	100.0#	0#	-8.46#
44 L9 AR-1268-4	1.102	0.000 E3	100.0#	0#	-8.74#
45 L9 AR-1268-5	8.539	0.000 E3	100.0#	0#	-9.03#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0