

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052522\
 Data File : PR054376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2022 06:12
 Operator : AJ\MA
 Sample : N2921-05
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 07:22:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 25 15:27:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.649	2.922	83996257	43319367	22.773	22.933
2)	SA Decachlor...	9.529	8.158	57128816	46219047	37.393	32.094
Target Compounds							
3)	L1 AR-1016-1	4.868	4.035	164.6E6	89728414	1312.620	1447.875
4)	L1 AR-1016-2	4.891	4.053	153.5E6	69539450	892.557	807.436
5)	L1 AR-1016-3	4.957	4.232	18451057	50041867	175.317	1109.947 #
6)	L1 AR-1016-4	5.056	4.282	88589049	699.8E6	1034.991	19154.909 #
7)	L1 AR-1016-5	5.371	4.498	602.9E6	371.0E6	7258.860	8012.003
8)	L2 AR-1221-1	3.857	3.140	5757467	1272407	137.773	61.475 #
9)	L2 AR-1221-2	3.959	3.223	1181873	804950	38.042	51.892 #
10)	L2 AR-1221-3	4.036	3.309	612354	333870	6.504	7.013
11)	L3 AR-1232-1	4.036	3.309	612354	333870	6.938	7.473
12)	L3 AR-1232-2	4.586	4.053	14142455	69539450	328.949	1748.493 #
13)	L3 AR-1232-3	4.891	4.232	153.5E6	50041867	1930.435	2405.645
14)	L3 AR-1232-4	5.056	4.323	88589049	267.6E6	2197.133	14631.498 #
15)	L3 AR-1232-5	5.160	4.498	1248.7E6	371.0E6	42296.935	17090.362 #
16)	L4 AR-1242-1	4.868	4.035	164.6E6	89728414	1763.440	1913.792
17)	L4 AR-1242-2	4.891	4.053	153.5E6	69539450	1179.061	1060.971
18)	L4 AR-1242-3	4.957	4.232	18451057	50041867	229.643	1447.268 #
19)	L4 AR-1242-4	5.056	4.323	88589049	267.6E6	1370.933	8092.279 #
20)	L4 AR-1242-5	5.844	4.867	1634.2E6	5784.7E6	25596.119	136638.277 #
21)	L5 AR-1248-1	4.868	4.035	164.6E6	89728414	2314.453	2570.213
22)	L5 AR-1248-2	5.160	4.282	1248.7E6	699.8E6	13549.641	14704.387
23)	L5 AR-1248-3	5.371	4.323	602.9E6	267.6E6	5718.262	5456.162
24)	L5 AR-1248-4	5.803	4.498	1642.5E6	371.0E6	14910.710	6327.245 #
25)	L5 AR-1248-5	5.844	4.908	1634.2E6	937.7E6	15066.178	15801.150
26)	L6 AR-1254-1	5.775	4.867	8154.8E6	5784.7E6	68921.806	62420.363
27)	L6 AR-1254-2	6.013	5.025	11507.1E6	7494.3E6	66050.564	92866.700 #
28)	L6 AR-1254-3	6.402	5.446	7623.8E6	5662.6E6	44929.285	43559.390
29)	L6 AR-1254-4	6.712	5.689	5072.0E6	3252.0E6	42850.409	39869.485
30)	L6 AR-1254-5	7.167	6.135	28554.8E6	29345.4E6	238153.819	257925.062
31)	L7 AR-1260-1	6.583	5.587	22192.8E6	19709.7E6	193139.373	240940.626
32)	L7 AR-1260-2	6.865	5.790	22883.6E6	21107.2E6	175948.354	213938.809
33)	L7 AR-1260-3	7.254	5.948	16595.5E6	22098.0E6	178002.689	242020.842 #
34)	L7 AR-1260-4	7.496	6.453	21719.5E6	19146.2E6	206403.496	249446.193
35)	L7 AR-1260-5	7.834	6.720	32925.9E6	38036.7E6	174954.569	213325.139
36)	L8 AR-1262-1	7.254	6.180	16595.5E6	18007.0E6	123173.253	157416.443 #
37)	L8 AR-1262-2	7.834	6.453	32925.9E6	19146.2E6	155737.232	186100.113
38)	L8 AR-1262-3	8.144	7.020	27213.1E6	10734.3E6	190401.536	134629.726 #
39)	L8 AR-1262-4	8.222	7.087	7259.2E6	32744.3E6	102466.242	219596.578 #
40)	L8 AR-1262-5	8.842	7.623	13238.2E6	13601.3E6	177953.959	198626.867
41)	L9 AR-1268-1	8.144	7.020	27213.1E6	10734.3E6	112492.032	45555.955 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052522\
 Data File : PR054376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2022 06:12
 Operator : AJ\MA
 Sample : N2921-05
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 07:22:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 25 15:27:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.222	7.087	7259.2E6	32744.3E6	33122.531	153203.042 #
43)	L9 AR-1268-3	8.441	7.309	414.2E6	461.3E6	2284.843	2560.956
44)	L9 AR-1268-4	8.842	7.623	13238.2E6	13601.3E6	163513.926	176899.745
45)	L9 AR-1268-5	9.223	7.916	2600.0E6	2803.6E6	4793.738	4948.212



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052522\
Data File : PR054376.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2022 06:12
Operator : AJ\MA
Sample : N2921-05
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 26 07:22:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 25 15:27:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.649 min
Delta R.T.: 0.000 min
Response: 83996257
Conc: 22.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.922 min
Delta R.T.: 0.000 min
Response: 43319367
Conc: 22.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.529 min
Delta R.T.: -0.011 min
Response: 57128816
Conc: 37.39 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.158 min
Delta R.T.: -0.004 min
Response: 46219047
Conc: 32.09 ng/ml



#3 AR-1016-1


R.T.: 4.868 min
Delta R.T.: -0.001 min
Response: 164591089
Conc: 1312.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.035 min
Delta R.T.: 0.000 min
Response: 89728414
Conc: 1447.87 ng/ml




#4 AR-1016-2

R.T.: 4.891 min
Delta R.T.: -0.001 min
Response: 153541424
Conc: 892.56 ng/ml



#4 AR-1016-2


R.T.: 4.053 min
Delta R.T.: -0.001 min
Response: 69539450
Conc: 807.44 ng/ml



#5 AR-1016-3


R.T.: 4.957 min
Delta R.T.: 0.001 min
Response: 18451057
Conc: 175.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.232 min
Delta R.T.: 0.000 min
Response: 50041867
Conc: 1109.95 ng/ml




#6 AR-1016-4

R.T.: 5.056 min
Delta R.T.: -0.001 min
Response: 88589049
Conc: 1034.99 ng/ml



#6 AR-1016-4


R.T.: 4.282 min
Delta R.T.: -0.001 min
Response: 699802951
Conc: 19154.91 ng/ml



#7 AR-1016-5


R.T.: 5.371 min
Delta R.T.: -0.002 min
Response: 602878469
Conc: 7258.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.498 min
Delta R.T.: -0.001 min
Response: 371041384
Conc: 8012.00 ng/ml




#8 AR-1221-1

R.T.: 3.857 min
Delta R.T.: -0.013 min
Response: 5757467
Conc: 137.77 ng/ml



#8 AR-1221-1


R.T.: 3.140 min
Delta R.T.: -0.007 min
Response: 1272407
Conc: 61.48 ng/ml



#9 AR-1221-2


R.T.: 3.959 min
Delta R.T.: 0.000 min
Response: 1181873
Conc: 38.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.223 min
Delta R.T.: -0.010 min
Response: 804950
Conc: 51.89 ng/ml




#10 AR-1221-3

R.T.: 4.036 min
Delta R.T.: -0.001 min
Response: 612354
Conc: 6.50 ng/ml



#10 AR-1221-3


R.T.: 3.309 min
Delta R.T.: 0.000 min
Response: 333870
Conc: 7.01 ng/ml



#11 AR-1232-1


R.T.: 4.036 min
Delta R.T.: 0.000 min
Response: 612354
Conc: 6.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.309 min
Delta R.T.: 0.000 min
Response: 333870
Conc: 7.47 ng/ml




#12 AR-1232-2

R.T.: 4.586 min
Delta R.T.: 0.000 min
Response: 14142455
Conc: 328.95 ng/ml



#12 AR-1232-2


R.T.: 4.053 min
Delta R.T.: 0.000 min
Response: 69539450
Conc: 1748.49 ng/ml



#13 AR-1232-3


R.T.: 4.891 min
Delta R.T.: -0.002 min
Response: 153541424
Conc: 1930.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.232 min
Delta R.T.: 0.000 min
Response: 50041867
Conc: 2405.64 ng/ml




#14 AR-1232-4

R.T.: 5.056 min
Delta R.T.: -0.002 min
Response: 88589049
Conc: 2197.13 ng/ml



#14 AR-1232-4


R.T.: 4.323 min
Delta R.T.: -0.001 min
Response: 267635291
Conc: 14631.50 ng/ml



#15 AR-1232-5


R.T.: 5.160 min
Delta R.T.: -0.002 min
Response: 1248665288
Conc: 42296.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.498 min
Delta R.T.: -0.001 min
Response: 371041384
Conc: 17090.36 ng/ml




#16 AR-1242-1

R.T.: 4.868 min
Delta R.T.: -0.002 min
Response: 164591089
Conc: 1763.44 ng/ml



#16 AR-1242-1


R.T.: 4.035 min
Delta R.T.: 0.000 min
Response: 89728414
Conc: 1913.79 ng/ml



#17 AR-1242-2


R.T.: 4.891 min
Delta R.T.: -0.001 min
Response: 153541424
Conc: 1179.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.053 min
Delta R.T.: 0.000 min
Response: 69539450
Conc: 1060.97 ng/ml




#18 AR-1242-3

R.T.: 4.957 min
Delta R.T.: 0.000 min
Response: 18451057
Conc: 229.64 ng/ml



#18 AR-1242-3


R.T.: 4.232 min
Delta R.T.: 0.000 min
Response: 50041867
Conc: 1447.27 ng/ml



#19 AR-1242-4


R.T.: 5.056 min
Delta R.T.: -0.002 min
Response: 88589049
Conc: 1370.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.323 min
Delta R.T.: -0.002 min
Response: 267635291
Conc: 8092.28 ng/ml




#20 AR-1242-5

R.T.: 5.844 min
Delta R.T.: -0.001 min
Response: 1634201623
Conc: 25596.12 ng/ml



#20 AR-1242-5


R.T.: 4.867 min
Delta R.T.: 0.000 min
Response: 5784727997
Conc: 136638.28 ng/ml



#21 AR-1248-1


R.T.: 4.868 min
Delta R.T.: -0.002 min
Response: 164591089
Conc: 2314.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.035 min
Delta R.T.: 0.000 min
Response: 89728414
Conc: 2570.21 ng/ml




#22 AR-1248-2

R.T.: 5.160 min
Delta R.T.: -0.001 min
Response: 1248665288
Conc: 13549.64 ng/ml



#22 AR-1248-2


R.T.: 4.282 min
Delta R.T.: 0.000 min
Response: 699802951
Conc: 14704.39 ng/ml



#23 AR-1248-3


R.T.: 5.371 min
Delta R.T.: -0.002 min
Response: 602878469
Conc: 5718.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.323 min
Delta R.T.: -0.001 min
Response: 267635291
Conc: 5456.16 ng/ml




#24 AR-1248-4

R.T.: 5.803 min
Delta R.T.: -0.002 min
Response: 1642538747
Conc: 14910.71 ng/ml



#24 AR-1248-4


R.T.: 4.498 min
Delta R.T.: 0.000 min
Response: 371041384
Conc: 6327.24 ng/ml



#25 AR-1248-5


R.T.: 5.844 min
Delta R.T.: -0.002 min
Response: 1634201623
Conc: 15066.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 4.908 min
Delta R.T.: 0.000 min
Response: 937714507
Conc: 15801.15 ng/ml




#26 AR-1254-1

R.T.: 5.775 min
Delta R.T.: -0.002 min
Response: 8154832244
Conc: 68921.81 ng/ml



#26 AR-1254-1


R.T.: 4.867 min
Delta R.T.: 0.000 min
Response: 5784727997
Conc: 62420.36 ng/ml



#27 AR-1254-2


R.T.: 6.013 min
Delta R.T.: -0.001 min
Response: 11507084945
Conc: 66050.56 ng/ml

Instrument : ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.025 min
Delta R.T.: 0.000 min
Response: 7494316514
Conc: 92866.70 ng/ml




#28 AR-1254-3

R.T.: 6.402 min
Delta R.T.: -0.002 min
Response: 7623765769
Conc: 44929.28 ng/ml



#28 AR-1254-3


R.T.: 5.446 min
Delta R.T.: 0.000 min
Response: 5662622784
Conc: 43559.39 ng/ml



#29 AR-1254-4


R.T.: 6.712 min
Delta R.T.: -0.004 min
Response: 5072039095
Conc: 42850.41 ng/ml

Instrument : ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.689 min
Delta R.T.: 0.000 min
Response: 3251950471
Conc: 39869.49 ng/ml




#30 AR-1254-5

R.T.: 7.167 min
Delta R.T.: -0.005 min
Response: 28554805428
Conc: 238153.82 ng/ml



#30 AR-1254-5


R.T.: 6.135 min
Delta R.T.: 0.000 min
Response: 29345435504
Conc: 257925.06 ng/ml



#31 AR-1260-1


R.T.: 6.583 min
Delta R.T.: -0.002 min
Response: 22192817662
Conc: 193139.37 ng

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.587 min
Delta R.T.: 0.001 min
Response: 19709689979
Conc: 240940.63 ng/ml




#32 AR-1260-2

R.T.: 6.865 min
Delta R.T.: -0.002 min
Response: 22883596992
Conc: 175948.35 ng/ml



#32 AR-1260-2


R.T.: 5.790 min
Delta R.T.: 0.001 min
Response: 21107159818
Conc: 213938.81 ng/ml



#33 AR-1260-3


R.T.: 7.254 min
Delta R.T.: -0.004 min
Response: 16595506117
Conc: 178002.69 ng

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 5.948 min
Delta R.T.: 0.000 min
Response: 22098019941
Conc: 242020.84 ng/ml




#34 AR-1260-4

R.T.: 7.496 min
Delta R.T.: -0.003 min
Response: 21719496008
Conc: 206403.50 ng/ml



#34 AR-1260-4


R.T.: 6.453 min
Delta R.T.: 0.000 min
Response: 19146200249
Conc: 249446.19 ng/ml



#35 AR-1260-5


R.T.: 7.834 min
Delta R.T.: -0.001 min
Response: 32925884523
Conc: 174954.57 ng

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 6.720 min
Delta R.T.: 0.003 min
Response: 38036678515
Conc: 213325.14 ng/ml




#36 AR-1262-1

R.T.: 7.254 min
Delta R.T.: -0.004 min
Response: 16595506117
Conc: 123173.25 ng/ml



#36 AR-1262-1


R.T.: 6.180 min
Delta R.T.: 0.000 min
Response: 18007007422
Conc: 157416.44 ng/ml



#37 AR-1262-2


R.T.: 7.834 min
Delta R.T.: -0.004 min
Response: 32925884523
Conc: 155737.23 ng

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.453 min
Delta R.T.: 0.000 min
Response: 19146200249
Conc: 186100.11 ng/ml




#38 AR-1262-3

R.T.: 8.144 min
Delta R.T.: 0.000 min
Response: 27213127266
Conc: 190401.54 ng/ml



#38 AR-1262-3


R.T.: 7.020 min
Delta R.T.: 0.000 min
Response: 10734321013
Conc: 134629.73 ng/ml



#39 AR-1262-4


R.T.: 8.222 min
Delta R.T.: -0.008 min
Response: 7259236823
Conc: 102466.24 ng

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.087 min
Delta R.T.: 0.000 min
Response: 32744348992
Conc: 219596.58 ng/ml




#40 AR-1262-5

R.T.: 8.842 min
Delta R.T.: -0.006 min
Response: 13238202440
Conc: 177953.96 ng/ml



#40 AR-1262-5


R.T.: 7.623 min
Delta R.T.: 0.000 min
Response: 13601254584
Conc: 198626.87 ng/ml



#41 AR-1268-1


R.T.: 8.144 min
Delta R.T.: 0.005 min
Response: 27213127266
Conc: 112492.03 ng

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.020 min
Delta R.T.: 0.000 min
Response: 10734321013
Conc: 45555.96 ng/ml




#42 AR-1268-2

R.T.: 8.222 min
Delta R.T.: -0.007 min
Response: 7259236823
Conc: 33122.53 ng/ml



#42 AR-1268-2


R.T.: 7.087 min
Delta R.T.: -0.004 min
Response: 32744348992
Conc: 153203.04 ng/ml



#43 AR-1268-3


R.T.: 8.441 min
Delta R.T.: -0.007 min
Response: 414157006
Conc: 2284.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.309 min
Delta R.T.: -0.003 min
Response: 461289055
Conc: 2560.96 ng/ml




#44 AR-1268-4

R.T.: 8.842 min
Delta R.T.: -0.007 min
Response: 13238202440
Conc: 163513.93 ng/ml



#44 AR-1268-4


R.T.: 7.623 min
Delta R.T.: -0.002 min
Response: 13601254584
Conc: 176899.75 ng/ml



#45 AR-1268-5

R.T.: 9.223 min
Delta R.T.: -0.008 min
Response: 2600028956
Conc: 4793.74 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 7.916 min
Delta R.T.: -0.003 min
Response: 2803629318
Conc: 4948.21 ng/ml