

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052523\
 Data File : PP057951.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2023 03:46
 Operator : YP\AJ
 Sample : 02934-04 10X
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 06:06:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.383	3.621	2478649	2840208	1.130	1.252
2)	SA Decachlor...	10.207	8.679	2088573	2187552	1.843	1.394
Target Compounds							
3)	L1 AR-1016-1	5.583	4.718	79074	97648	1.269	1.310
4)	L1 AR-1016-2	5.599	4.737	42915	41896	0.477	0.405
5)	L1 AR-1016-3	5.665	4.916	23866	90640	0.422	1.537 #
6)	L1 AR-1016-4	5.750	4.957	36111	276598	0.782	5.880 #
7)	L1 AR-1016-5	6.061	5.171	47315	155889	1.015	2.625 #
8)	L2 AR-1221-1	4.604	0.000	86302	0	2.982	N.D. #
9)	L2 AR-1221-2	4.698	3.924	164535	93039	7.592	4.177 #
10)	L2 AR-1221-3	4.771	3.976f	90524	27627	1.503	0.434 #
11)	L3 AR-1232-1	4.771	3.976f	90524	27627	2.067	0.599 #
12)	L3 AR-1232-2	5.292	4.737	103406	41896	3.965	0.885 #
13)	L3 AR-1232-3	5.599	4.916	42915	90640	1.051	3.330 #
14)	L3 AR-1232-4	5.750	5.001	36111	104222	1.696	4.275 #
15)	L3 AR-1232-5	5.839	5.171	215921	155889	11.823	5.637 #
16)	L4 AR-1242-1	5.557	4.718	182125	97648	3.823	1.677 #
17)	L4 AR-1242-2	5.599	4.737	42915	41896	0.628	0.525
18)	L4 AR-1242-3	5.644	4.916	66822	90640	1.535	1.933 #
19)	L4 AR-1242-4	5.750	5.001	36111	104222	1.025	2.246 #
20)	L4 AR-1242-5	6.488	5.527	211460	532985	6.526	9.533 #
21)	L5 AR-1248-1	5.557	4.718	182125	97648	4.328	1.898 #
22)	L5 AR-1248-2	5.839	4.957	215921	276598	3.312	3.852
23)	L5 AR-1248-3	6.061	5.001	47315	104222	0.681	1.378 #
24)	L5 AR-1248-4	6.454	5.171	320107	155889	5.009	1.748 #
25)	L5 AR-1248-5	6.488	5.565	211460	139063	3.296	1.673 #
26)	L6 AR-1254-1	6.427	5.527	363653	532985	5.418	4.773
27)	L6 AR-1254-2	6.647	5.676	714654	542360	7.555	5.434 #
28)	L6 AR-1254-3	7.016	6.081	1297682	666570	14.246	4.261 #
29)	L6 AR-1254-4	7.304	6.312	763460	445920	12.818	4.834 #
30)	L6 AR-1254-5	7.727	6.735	1096202	918648	15.968	6.346 #
31)	L7 AR-1260-1	7.180	6.213	651050	366330	8.252	3.332 #
32)	L7 AR-1260-2	7.443	6.400	613343	682269	7.655	5.368 #
33)	L7 AR-1260-3	7.827	6.558	196850	493542	3.455	4.025
34)	L7 AR-1260-4	8.047	7.042	173658	70267	2.614	0.804 #
35)	L7 AR-1260-5	8.358	7.274	252990	72681	2.356	0.394 #
36)	L8 AR-1262-1	7.827f	6.812f	196850	716354	2.286	11.275 #
37)	L8 AR-1262-2	8.358	7.042	252990	70267	2.078	0.567 #
38)	L8 AR-1262-3	8.696	7.556	118911	201608	1.368	2.149 #
39)	L8 AR-1262-4	8.767	7.623	41114	215275	0.613	1.264 #
40)	L8 AR-1262-5	0.000	8.132	0	379740	N.D.	5.160 #
41)	L9 AR-1268-1	8.696f	7.556	118911	201608	0.764	0.766

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052523\
Data File : PP057951.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2023 03:46
Operator : YP\AJ
Sample : 02934-04 10X
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 26 06:06:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.767	7.623	41114	215275	0.315	0.879 #
43)	L9 AR-1268-3	9.012	7.853	137207	52311	1.063	0.258 #
44)	L9 AR-1268-4	0.000	8.132	0	379740	N.D.	4.585 #
45)	L9 AR-1268-5	9.865	8.425	45947	128896	0.141	0.249 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052523\
Data File : PP057951.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2023 03:46
Operator : YP\AJ
Sample : 02934-04 10X
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 26 06:06:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.383 min
Delta R.T.: -0.007 min
Response: 2478649
Conc: 1.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.621 min
Delta R.T.: -0.002 min
Response: 2840208
Conc: 1.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.207 min
Delta R.T.: -0.017 min
Response: 2088573
Conc: 1.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.679 min
Delta R.T.: -0.007 min
Response: 2187552
Conc: 1.39 ng/ml



#3 AR-1016-1


R.T.: 5.583 min
Delta R.T.: 0.013 min
Response: 79074
Conc: 1.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.718 min
Delta R.T.: -0.003 min
Response: 97648
Conc: 1.31 ng/ml




#4 AR-1016-2

R.T.: 5.599 min
Delta R.T.: 0.006 min
Response: 42915
Conc: 0.48 ng/ml



#4 AR-1016-2


R.T.: 4.737 min
Delta R.T.: -0.004 min
Response: 41896
Conc: 0.41 ng/ml



#5 AR-1016-3


R.T.: 5.665 min
Delta R.T.: 0.010 min
Response: 23866
Conc: 0.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.916 min
Delta R.T.: -0.002 min
Response: 90640
Conc: 1.54 ng/ml




#6 AR-1016-4

R.T.: 5.750 min
Delta R.T.: -0.004 min
Response: 36111
Conc: 0.78 ng/ml



#6 AR-1016-4


R.T.: 4.957 min
Delta R.T.: -0.003 min
Response: 276598
Conc: 5.88 ng/ml



#7 AR-1016-5


R.T.: 6.061 min
Delta R.T.: 0.009 min
Response: 47315
Conc: 1.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.171 min
Delta R.T.: -0.006 min
Response: 155889
Conc: 2.63 ng/ml




#8 AR-1221-1

R.T.: 4.604 min
Delta R.T.: 0.005 min
Response: 86302
Conc: 2.98 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 3.839 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.698 min
Delta R.T.: 0.012 min
Response: 164535
Conc: 7.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.924 min
Delta R.T.: 0.000 min
Response: 93039
Conc: 4.18 ng/ml




#10 AR-1221-3

R.T.: 4.771 min
Delta R.T.: 0.009 min
Response: 90524
Conc: 1.50 ng/ml



#10 AR-1221-3


R.T.: 3.976 min
Delta R.T.: -0.026 min
Response: 27627
Conc: 0.43 ng/ml



#11 AR-1232-1


R.T.: 4.771 min
Delta R.T.: 0.008 min
Response: 90524
Conc: 2.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.976 min
Delta R.T.: -0.026 min
Response: 27627
Conc: 0.60 ng/ml




#12 AR-1232-2

R.T.: 5.292 min
Delta R.T.: -0.006 min
Response: 103406
Conc: 3.97 ng/ml



#12 AR-1232-2


R.T.: 4.737 min
Delta R.T.: -0.002 min
Response: 41896
Conc: 0.89 ng/ml



#13 AR-1232-3


R.T.: 5.599 min
Delta R.T.: 0.007 min
Response: 42915
Conc: 1.05 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.916 min
Delta R.T.: -0.001 min
Response: 90640
Conc: 3.33 ng/ml




#14 AR-1232-4

R.T.: 5.750 min
Delta R.T.: -0.002 min
Response: 36111
Conc: 1.70 ng/ml



#14 AR-1232-4


R.T.: 5.001 min
Delta R.T.: 0.000 min
Response: 104222
Conc: 4.28 ng/ml



#15 AR-1232-5


R.T.: 5.839 min
Delta R.T.: -0.007 min
Response: 215921
Conc: 11.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.171 min
Delta R.T.: -0.005 min
Response: 155889
Conc: 5.64 ng/ml




#16 AR-1242-1

R.T.: 5.557 min
Delta R.T.: -0.012 min
Response: 182125
Conc: 3.82 ng/ml



#16 AR-1242-1


R.T.: 4.718 min
Delta R.T.: -0.002 min
Response: 97648
Conc: 1.68 ng/ml



#17 AR-1242-2


R.T.: 5.599 min
Delta R.T.: 0.008 min
Response: 42915
Conc: 0.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.737 min
Delta R.T.: -0.003 min
Response: 41896
Conc: 0.52 ng/ml




#18 AR-1242-3

R.T.: 5.644 min
Delta R.T.: -0.010 min
Response: 66822
Conc: 1.53 ng/ml



#18 AR-1242-3


R.T.: 4.916 min
Delta R.T.: -0.001 min
Response: 90640
Conc: 1.93 ng/ml



#19 AR-1242-4


R.T.: 5.750 min
Delta R.T.: -0.003 min
Response: 36111
Conc: 1.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.001 min
Delta R.T.: -0.001 min
Response: 104222
Conc: 2.25 ng/ml




#20 AR-1242-5

R.T.: 6.488 min
Delta R.T.: -0.009 min
Response: 211460
Conc: 6.53 ng/ml



#20 AR-1242-5


R.T.: 5.527 min
Delta R.T.: -0.004 min
Response: 532985
Conc: 9.53 ng/ml



#21 AR-1248-1


R.T.: 5.557 min
Delta R.T.: -0.012 min
Response: 182125
Conc: 4.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.718 min
Delta R.T.: -0.002 min
Response: 97648
Conc: 1.90 ng/ml




#22 AR-1248-2

R.T.: 5.839 min
Delta R.T.: -0.007 min
Response: 215921
Conc: 3.31 ng/ml



#22 AR-1248-2


R.T.: 4.957 min
Delta R.T.: -0.003 min
Response: 276598
Conc: 3.85 ng/ml



#23 AR-1248-3


R.T.: 6.061 min
Delta R.T.: 0.009 min
Response: 47315
Conc: 0.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.001 min
Delta R.T.: -0.001 min
Response: 104222
Conc: 1.38 ng/ml




#24 AR-1248-4

R.T.: 6.454 min
Delta R.T.: -0.005 min
Response: 320107
Conc: 5.01 ng/ml



#24 AR-1248-4


R.T.: 5.171 min
Delta R.T.: -0.005 min
Response: 155889
Conc: 1.75 ng/ml



#25 AR-1248-5


R.T.: 6.488 min
Delta R.T.: -0.010 min
Response: 211460
Conc: 3.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.565 min
Delta R.T.: -0.007 min
Response: 139063
Conc: 1.67 ng/ml




#26 AR-1254-1

R.T.: 6.427 min
Delta R.T.: -0.007 min
Response: 363653
Conc: 5.42 ng/ml



#26 AR-1254-1


R.T.: 5.527 min
Delta R.T.: -0.005 min
Response: 532985
Conc: 4.77 ng/ml



#27 AR-1254-2


R.T.: 6.647 min
Delta R.T.: -0.007 min
Response: 714654
Conc: 7.55 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.676 min
Delta R.T.: -0.005 min
Response: 542360
Conc: 5.43 ng/ml




#28 AR-1254-3

R.T.: 7.016 min
Delta R.T.: -0.008 min
Response: 1297682
Conc: 14.25 ng/ml



#28 AR-1254-3


R.T.: 6.081 min
Delta R.T.: -0.007 min
Response: 666570
Conc: 4.26 ng/ml



#29 AR-1254-4


R.T.: 7.304 min
Delta R.T.: -0.008 min
Response: 763460
Conc: 12.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.312 min
Delta R.T.: -0.006 min
Response: 445920
Conc: 4.83 ng/ml




#30 AR-1254-5

R.T.: 7.727 min
Delta R.T.: -0.009 min
Response: 1096202
Conc: 15.97 ng/ml



#30 AR-1254-5


R.T.: 6.735 min
Delta R.T.: -0.005 min
Response: 918648
Conc: 6.35 ng/ml



#31 AR-1260-1


R.T.: 7.180 min
Delta R.T.: -0.011 min
Response: 651050
Conc: 8.25 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.213 min
Delta R.T.: -0.007 min
Response: 366330
Conc: 3.33 ng/ml




#32 AR-1260-2

R.T.: 7.443 min
Delta R.T.: -0.006 min
Response: 613343
Conc: 7.66 ng/ml



#32 AR-1260-2


R.T.: 6.400 min
Delta R.T.: -0.009 min
Response: 682269
Conc: 5.37 ng/ml



#33 AR-1260-3


R.T.: 7.827 min
Delta R.T.: 0.015 min
Response: 196850
Conc: 3.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.558 min
Delta R.T.: -0.005 min
Response: 493542
Conc: 4.02 ng/ml




#34 AR-1260-4

R.T.: 8.047 min
Delta R.T.: 0.007 min
Response: 173658
Conc: 2.61 ng/ml



#34 AR-1260-4


R.T.: 7.042 min
Delta R.T.: 0.003 min
Response: 70267
Conc: 0.80 ng/ml



#35 AR-1260-5


R.T.: 8.358 min
Delta R.T.: -0.006 min
Response: 252990
Conc: 2.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.274 min
Delta R.T.: -0.009 min
Response: 72681
Conc: 0.39 ng/ml




#36 AR-1262-1

R.T.: 7.827 min
Delta R.T.: 0.016 min
Response: 196850
Conc: 2.29 ng/ml



#36 AR-1262-1


R.T.: 6.812 min
Delta R.T.: -0.019 min
Response: 716354
Conc: 11.27 ng/ml



#37 AR-1262-2


R.T.: 8.358 min
Delta R.T.: -0.004 min
Response: 252990
Conc: 2.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.042 min
Delta R.T.: 0.003 min
Response: 70267
Conc: 0.57 ng/ml




#38 AR-1262-3

R.T.: 8.696 min
Delta R.T.: 0.015 min
Response: 118911
Conc: 1.37 ng/ml



#38 AR-1262-3


R.T.: 7.556 min
Delta R.T.: -0.011 min
Response: 201608
Conc: 2.15 ng/ml



#39 AR-1262-4


R.T.: 8.767 min
Delta R.T.: -0.005 min
Response: 41114
Conc: 0.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.623 min
Delta R.T.: -0.008 min
Response: 215275
Conc: 1.26 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 9.442 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 8.132 min
Delta R.T.: 0.002 min
Response: 379740
Conc: 5.16 ng/ml



#41 AR-1268-1


R.T.: 8.696 min
Delta R.T.: 0.016 min
Response: 118911
Conc: 0.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.556 min
Delta R.T.: -0.013 min
Response: 201608
Conc: 0.77 ng/ml




#42 AR-1268-2

R.T.: 8.767 min
Delta R.T.: -0.009 min
Response: 41114
Conc: 0.32 ng/ml



#42 AR-1268-2


R.T.: 7.623 min
Delta R.T.: -0.010 min
Response: 215275
Conc: 0.88 ng/ml



#43 AR-1268-3


R.T.: 9.012 min
Delta R.T.: 0.003 min
Response: 137207
Conc: 1.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.853 min
Delta R.T.: 0.011 min
Response: 52311
Conc: 0.26 ng/ml




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 9.445 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 8.132 min
Delta R.T.: 0.001 min
Response: 379740
Conc: 4.59 ng/ml



#45 AR-1268-5

R.T.: 9.865 min
Delta R.T.: -0.006 min
Response: 45947
Conc: 0.14 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.425 min
Delta R.T.: -0.002 min
Response: 128896
Conc: 0.25 ng/ml