

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052523\
 Data File : PP057900.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 May 2023 11:11
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 25 12:41:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.381	3.620	109.3E6	115.9E6	49.825	51.109
2) SA Decachlor...	10.205	8.677	65862949	72261750	58.119	46.034
Target Compounds						
16) L4 AR-1242-1	5.560	4.715	24749339	29022559	519.530	498.575
17) L4 AR-1242-2	5.582	4.734	35707465	40264089	522.139	504.447
18) L4 AR-1242-3	5.646	4.912	23518098	24323352	540.144	518.642
19) L4 AR-1242-4	5.744	4.997	18397264	23667595	522.101	510.073
20) L4 AR-1242-5	6.488	5.525	17887141	27077165	552.050	484.302



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052523\
Data File : PP057900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 May 2023 11:11
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 25 12:41:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.381 min
Delta R.T.: -0.009 min
Response: 109308496
Conc: 49.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.620 min
Delta R.T.: -0.004 min
Response: 115932239
Conc: 51.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.205 min
Delta R.T.: -0.018 min
Response: 65862949
Conc: 58.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.677 min
Delta R.T.: -0.008 min
Response: 72261750
Conc: 46.03 ng/ml



#16 AR-1242-1


R.T.: 5.560 min
Delta R.T.: -0.008 min
Response: 24749339
Conc: 519.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.715 min
Delta R.T.: -0.005 min
Response: 29022559
Conc: 498.57 ng/ml




#17 AR-1242-2

R.T.: 5.582 min
Delta R.T.: -0.009 min
Response: 35707465
Conc: 522.14 ng/ml



#17 AR-1242-2


R.T.: 4.734 min
Delta R.T.: -0.005 min
Response: 40264089
Conc: 504.45 ng/ml



#18 AR-1242-3


R.T.: 5.646 min
Delta R.T.: -0.008 min
Response: 23518098
Conc: 540.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.912 min
Delta R.T.: -0.005 min
Response: 24323352
Conc: 518.64 ng/ml




#19 AR-1242-4

R.T.: 5.744 min
Delta R.T.: -0.009 min
Response: 18397264
Conc: 522.10 ng/ml



#19 AR-1242-4


R.T.: 4.997 min
Delta R.T.: -0.005 min
Response: 23667595
Conc: 510.07 ng/ml



#20 AR-1242-5

R.T.: 6.488 min
Delta R.T.: -0.009 min
Response: 17887141
Conc: 552.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.525 min
Delta R.T.: -0.006 min
Response: 27077165
Conc: 484.30 ng/ml