

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052523\
 Data File : PP057939.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 May 2023 23:30
 Operator : YP\AJ
 Sample : 02933-01 10X
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 17A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 05:59:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.383	3.621	4145652	4536933	1.890	2.000
2) SA Decachlor...	10.206	8.680	3061959	3041184	2.702	1.937 #
Target Compounds						
26) L6 AR-1254-1	6.426	5.527	7995727	11325268	119.130	101.427
27) L6 AR-1254-2	6.646	5.676	11569475	10499991	122.301	105.207
28) L6 AR-1254-3	7.015	6.084	10875594	14138301	119.396	90.368
29) L6 AR-1254-4	7.303	6.313	7024384	7904522	117.934	85.684 #
30) L6 AR-1254-5	7.727	6.735	8403736	12203441	122.411	84.301 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052523\
Data File : PP057939.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 May 2023 23:30
Operator : YP\AJ
Sample : 02933-01 10X
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
17A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 26 05:59:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.383 min
Delta R.T.: -0.007 min
Response: 4145652
Conc: 1.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
17A




#1 Tetrachloro-m-xylene

R.T.: 3.621 min
Delta R.T.: -0.002 min
Response: 4536933
Conc: 2.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.206 min
Delta R.T.: -0.017 min
Response: 3061959
Conc: 2.70 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.680 min
Delta R.T.: -0.006 min
Response: 3041184
Conc: 1.94 ng/ml



#26 AR-1254-1


R.T.: 6.426 min
Delta R.T.: -0.008 min
Response: 7995727
Conc: 119.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
17A




#26 AR-1254-1

R.T.: 5.527 min
Delta R.T.: -0.004 min
Response: 11325268
Conc: 101.43 ng/ml




#27 AR-1254-2

R.T.: 6.646 min
Delta R.T.: -0.009 min
Response: 11569475
Conc: 122.30 ng/ml



#27 AR-1254-2


R.T.: 5.676 min
Delta R.T.: -0.004 min
Response: 10499991
Conc: 105.21 ng/ml



#28 AR-1254-3


R.T.: 7.015 min
Delta R.T.: -0.008 min
Response: 10875594
Conc: 119.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
17A




#28 AR-1254-3

R.T.: 6.084 min
Delta R.T.: -0.005 min
Response: 14138301
Conc: 90.37 ng/ml




#29 AR-1254-4

R.T.: 7.303 min
Delta R.T.: -0.008 min
Response: 7024384
Conc: 117.93 ng/ml



#29 AR-1254-4


R.T.: 6.313 min
Delta R.T.: -0.005 min
Response: 7904522
Conc: 85.68 ng/ml



#30 AR-1254-5

R.T.: 7.727 min
Delta R.T.: -0.009 min
Response: 8403736
Conc: 122.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
17A



#30 AR-1254-5

R.T.: 6.735 min
Delta R.T.: -0.005 min
Response: 12203441
Conc: 84.30 ng/ml