

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052523\
 Data File : PP057946.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2023 02:24
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 06:03:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.383	3.622	95901652	104.5E6	43.714	46.080
2) SA Decachlor...	10.206	8.679	57508737	65308399	50.747	41.604
Target Compounds						
26) L6 AR-1254-1	6.425	5.527	30545366	48865164	455.100	437.627
27) L6 AR-1254-2	6.646	5.676	46196108	43221606	488.340	433.067
28) L6 AR-1254-3	7.016	6.083	47431991	67223569	520.724	429.674
29) L6 AR-1254-4	7.303	6.313	30910023	38390587	518.954	416.149
30) L6 AR-1254-5	7.726	6.734	36462110	59949024	531.115	414.127



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052523\
Data File : PP057946.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2023 02:24
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 26 06:03:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.383 min
Delta R.T.: -0.007 min
Response: 95901652
Conc: 43.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: -0.002 min
Response: 104525204
Conc: 46.08 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.206 min
Delta R.T.: -0.017 min
Response: 57508737
Conc: 50.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.679 min
Delta R.T.: -0.007 min
Response: 65308399
Conc: 41.60 ng/ml



#26 AR-1254-1


R.T.: 6.425 min
Delta R.T.: -0.008 min
Response: 30545366
Conc: 455.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.527 min
Delta R.T.: -0.005 min
Response: 48865164
Conc: 437.63 ng/ml




#27 AR-1254-2

R.T.: 6.646 min
Delta R.T.: -0.008 min
Response: 46196108
Conc: 488.34 ng/ml



#27 AR-1254-2


R.T.: 5.676 min
Delta R.T.: -0.005 min
Response: 43221606
Conc: 433.07 ng/ml



#28 AR-1254-3


R.T.: 7.016 min
Delta R.T.: -0.008 min
Response: 47431991
Conc: 520.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.083 min
Delta R.T.: -0.005 min
Response: 67223569
Conc: 429.67 ng/ml




#29 AR-1254-4

R.T.: 7.303 min
Delta R.T.: -0.008 min
Response: 30910023
Conc: 518.95 ng/ml



#29 AR-1254-4


R.T.: 6.313 min
Delta R.T.: -0.005 min
Response: 38390587
Conc: 416.15 ng/ml



#30 AR-1254-5

R.T.: 7.726 min
Delta R.T.: -0.009 min
Response: 36462110
Conc: 531.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.734 min
Delta R.T.: -0.006 min
Response: 59949024
Conc: 414.13 ng/ml