

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052523\
 Data File : PP057966.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2023 08:53
 Operator : YP\AJ
 Sample : 02939-04 10X
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HEE595N-2-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 13:16:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 08 19:18:05 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.382	3.643	3576237	3596182	1.630	1.585
2) SA Decachlor...	10.206	8.702	2724027	2536173	2.404	1.616 #
Target Compounds						
26) L6 AR-1254-1	6.424	5.548f	9277202	12261559	138.223	109.812
27) L6 AR-1254-2	6.645	5.696f	13961674	11978876	147.589	120.025
28) L6 AR-1254-3	7.014	6.104f	13944845	17413682	153.091	111.303 #
29) L6 AR-1254-4	7.302	6.334f	9518386	9875533	159.806	107.050 #
30) L6 AR-1254-5	7.725	6.756f	12793186	17212593	186.348	118.904 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052523\
Data File : PP057966.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2023 08:53
Operator : YP\AJ
Sample : 02939-04 10X
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HEE595N-2-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 26 13:16:30 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP050823.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 08 19:18:05 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.382 min
Delta R.T.: -0.008 min
Response: 3576237
Conc: 1.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEE595N-2-2




#1 Tetrachloro-m-xylene

R.T.: 3.643 min
Delta R.T.: 0.020 min
Response: 3596182
Conc: 1.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.206 min
Delta R.T.: -0.017 min
Response: 2724027
Conc: 2.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.702 min
Delta R.T.: 0.016 min
Response: 2536173
Conc: 1.62 ng/ml



#26 AR-1254-1


R.T.: 6.424 min
Delta R.T.: -0.009 min
Response: 9277202
Conc: 138.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEE595N-2-2




#26 AR-1254-1

R.T.: 5.548 min
Delta R.T.: 0.016 min
Response: 12261559
Conc: 109.81 ng/ml




#27 AR-1254-2

R.T.: 6.645 min
Delta R.T.: -0.009 min
Response: 13961674
Conc: 147.59 ng/ml



#27 AR-1254-2


R.T.: 5.696 min
Delta R.T.: 0.016 min
Response: 11978876
Conc: 120.02 ng/ml



#28 AR-1254-3


R.T.: 7.014 min
Delta R.T.: -0.009 min
Response: 13944845
Conc: 153.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEE595N-2-2




#28 AR-1254-3

R.T.: 6.104 min
Delta R.T.: 0.016 min
Response: 17413682
Conc: 111.30 ng/ml




#29 AR-1254-4

R.T.: 7.302 min
Delta R.T.: -0.009 min
Response: 9518386
Conc: 159.81 ng/ml



#29 AR-1254-4


R.T.: 6.334 min
Delta R.T.: 0.015 min
Response: 9875533
Conc: 107.05 ng/ml



#30 AR-1254-5

R.T.: 7.725 min
Delta R.T.: -0.011 min
Response: 12793186
Conc: 186.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
HEE595N-2-2



#30 AR-1254-5

R.T.: 6.756 min
Delta R.T.: 0.016 min
Response: 17212593
Conc: 118.90 ng/ml