

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052722\
 Data File : PP048105.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2022 18:05
 Operator : YP\AJ
 Sample : PB144739BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB144739BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 00:04:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:53:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.038	3.222	41547255	36102819	17.742	18.319
2)	SA Decachlor...	10.391	8.668	35153687	30632218	18.087	18.950
Target Compounds							
3)	L1 AR-1016-1	5.349f	4.389	129177	59150	1.532	0.879 #
4)	L1 AR-1016-2	5.349	4.413	129177	79921	1.092	0.858
5)	L1 AR-1016-3	5.407	4.612f	329070	71205	4.870	1.390 #
6)	L1 AR-1016-4	5.530	4.640	261417	20823	4.241	0.547 #
7)	L1 AR-1016-5	5.837	4.888f	230755	43708	4.097	0.915 #
8)	L2 AR-1221-1	4.281	3.438f	95915	39216	3.628	1.892 #
9)	L2 AR-1221-2	4.372	3.540	779515	481429	38.060	31.344
10)	L2 AR-1221-3	4.476f	0.000	160131	0	2.558	N.D. #
11)	L3 AR-1232-1	4.476f	0.000	160131	0	3.083	N.D. #
12)	L3 AR-1232-2	5.011f	4.399	631568	32698	26.307	0.812 #
13)	L3 AR-1232-3	5.349	4.612f	129177	71205	2.603	3.451 #
14)	L3 AR-1232-4	5.530	4.685	261417	92471	10.936	5.328 #
15)	L3 AR-1232-5	5.623	4.888f	108745	43708	6.656	2.231 #
16)	L4 AR-1242-1	5.349f	4.389	129177	59150	2.310	1.240 #
17)	L4 AR-1242-2	5.349	4.413	129177	79921	1.571	1.203
18)	L4 AR-1242-3	5.427	4.612f	505717	71205	10.302	2.053 #
19)	L4 AR-1242-4	5.530	4.685	261417	92471	6.606	2.879 #
20)	L4 AR-1242-5	6.323f	5.253	1169145	79650	27.452	1.907 #
21)	L5 AR-1248-1	5.349f	4.389	129177	59150	2.995	1.616 #
22)	L5 AR-1248-2	5.623	4.640	108745	20823	1.902	0.437 #
23)	L5 AR-1248-3	5.837	4.685	230755	92471	3.409	1.867 #
24)	L5 AR-1248-4	6.323f	4.888f	1169145	43708	15.034	0.750 #
25)	L5 AR-1248-5	6.323f	5.295	1169145	26207	15.370	0.434 #
26)	L6 AR-1254-1	6.238f	5.247	69047	38522	0.951	0.439 #
27)	L6 AR-1254-2	6.513	5.409	337328	41046	2.835	0.546 #
28)	L6 AR-1254-3	0.000	5.835	0	325342	N.D.	2.652 #
29)	L6 AR-1254-4	0.000	6.104	0	66449	N.D.	0.885 #
30)	L6 AR-1254-5	7.712	6.558	620620	74263	6.697	0.734 #
31)	L7 AR-1260-1	7.102	5.986	462731	50349	4.958	0.588 #
32)	L7 AR-1260-2	7.366f	6.184f	165675	18038	1.482	0.175 #
33)	L7 AR-1260-3	7.802f	6.370	1504873	151471	17.569	1.530 #
34)	L7 AR-1260-4	8.030	6.883	4353785	57529	41.961	0.691 #
35)	L7 AR-1260-5	8.395	7.149	129777	42980	0.629	0.217 #
36)	L8 AR-1262-1	7.802f	6.600	1504873	93810	13.195	0.868 #
37)	L8 AR-1262-2	8.395	6.883	129777	57529	0.609	0.581
38)	L8 AR-1262-3	0.000	7.465	0	248514	N.D.	3.302 #
39)	L8 AR-1262-4	8.837	7.534	1071030	95498	18.013	0.656 #
40)	L8 AR-1262-5	9.562	8.068	245013	183108	3.096	2.713
41)	L9 AR-1268-1	0.000	7.465	0	248514	N.D.	1.064 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052722\
 Data File : PP048105.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2022 18:05
 Operator : YP\AJ
 Sample : PB144739BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB144739BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 00:04:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:53:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.837	7.534	1071030	95498	4.435	0.441 #
43) L9	AR-1268-3	9.118f	7.754	1553182	128427	7.555	0.691 #
44) L9	AR-1268-4	9.562	8.068	245013	183108	2.728	2.229
45) L9	AR-1268-5	10.023	8.385	588109	131534	0.833	0.217 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052722\
Data File : PP048105.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2022 18:05
Operator : YP\AJ
Sample : PB144739BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB144739BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 00:04:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 27 23:53:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.038 min
Delta R.T.: 0.000 min
Response: 41547255
Conc: 17.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#1 Tetrachloro-m-xylene

R.T.: 3.222 min
Delta R.T.: -0.001 min
Response: 36102819
Conc: 18.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.391 min
Delta R.T.: -0.001 min
Response: 35153687
Conc: 18.09 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.668 min
Delta R.T.: -0.001 min
Response: 30632218
Conc: 18.95 ng/ml



#3 AR-1016-1


R.T.: 5.349 min
Delta R.T.: 0.024 min
Response: 129177
Conc: 1.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#3 AR-1016-1

R.T.: 4.389 min
Delta R.T.: 0.002 min
Response: 59150
Conc: 0.88 ng/ml




#4 AR-1016-2

R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 129177
Conc: 1.09 ng/ml



#4 AR-1016-2


R.T.: 4.413 min
Delta R.T.: 0.006 min
Response: 79921
Conc: 0.86 ng/ml



#5 AR-1016-3


R.T.: 5.407 min
Delta R.T.: -0.009 min
Response: 329070
Conc: 4.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#5 AR-1016-3

R.T.: 4.612 min
Delta R.T.: 0.018 min
Response: 71205
Conc: 1.39 ng/ml




#6 AR-1016-4

R.T.: 5.530 min
Delta R.T.: 0.004 min
Response: 261417
Conc: 4.24 ng/ml



#6 AR-1016-4


R.T.: 4.640 min
Delta R.T.: -0.002 min
Response: 20823
Conc: 0.55 ng/ml



#7 AR-1016-5


R.T.: 5.837 min
Delta R.T.: -0.014 min
Response: 230755
Conc: 4.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#7 AR-1016-5

R.T.: 4.888 min
Delta R.T.: 0.018 min
Response: 43708
Conc: 0.92 ng/ml




#8 AR-1221-1

R.T.: 4.281 min
Delta R.T.: 0.009 min
Response: 95915
Conc: 3.63 ng/ml



#8 AR-1221-1


R.T.: 3.438 min
Delta R.T.: -0.019 min
Response: 39216
Conc: 1.89 ng/ml



#9 AR-1221-2


R.T.: 4.372 min
Delta R.T.: 0.003 min
Response: 779515
Conc: 38.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#9 AR-1221-2

R.T.: 3.540 min
Delta R.T.: -0.007 min
Response: 481429
Conc: 31.34 ng/ml




#10 AR-1221-3

R.T.: 4.476 min
Delta R.T.: 0.024 min
Response: 160131
Conc: 2.56 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 3.628 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.476 min
Delta R.T.: 0.024 min
Response: 160131
Conc: 3.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.627 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.011 min
Delta R.T.: -0.018 min
Response: 631568
Conc: 26.31 ng/ml



#12 AR-1232-2


R.T.: 4.399 min
Delta R.T.: -0.007 min
Response: 32698
Conc: 0.81 ng/ml



#13 AR-1232-3


R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 129177
Conc: 2.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#13 AR-1232-3

R.T.: 4.612 min
Delta R.T.: 0.020 min
Response: 71205
Conc: 3.45 ng/ml




#14 AR-1232-4

R.T.: 5.530 min
Delta R.T.: 0.004 min
Response: 261417
Conc: 10.94 ng/ml



#14 AR-1232-4


R.T.: 4.685 min
Delta R.T.: -0.001 min
Response: 92471
Conc: 5.33 ng/ml



#15 AR-1232-5


R.T.: 5.623 min
Delta R.T.: -0.002 min
Response: 108745
Conc: 6.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#15 AR-1232-5

R.T.: 4.888 min
Delta R.T.: 0.019 min
Response: 43708
Conc: 2.23 ng/ml




#16 AR-1242-1

R.T.: 5.349 min
Delta R.T.: 0.023 min
Response: 129177
Conc: 2.31 ng/ml



#16 AR-1242-1


R.T.: 4.389 min
Delta R.T.: 0.002 min
Response: 59150
Conc: 1.24 ng/ml



#17 AR-1242-2


R.T.: 5.349 min
Delta R.T.: -0.001 min
Response: 129177
Conc: 1.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#17 AR-1242-2

R.T.: 4.413 min
Delta R.T.: 0.007 min
Response: 79921
Conc: 1.20 ng/ml




#18 AR-1242-3

R.T.: 5.427 min
Delta R.T.: 0.009 min
Response: 505717
Conc: 10.30 ng/ml



#18 AR-1242-3


R.T.: 4.612 min
Delta R.T.: 0.019 min
Response: 71205
Conc: 2.05 ng/ml



#19 AR-1242-4


R.T.: 5.530 min
Delta R.T.: 0.004 min
Response: 261417
Conc: 6.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#19 AR-1242-4

R.T.: 4.685 min
Delta R.T.: -0.001 min
Response: 92471
Conc: 2.88 ng/ml




#20 AR-1242-5

R.T.: 6.323 min
Delta R.T.: -0.017 min
Response: 1169145
Conc: 27.45 ng/ml



#20 AR-1242-5


R.T.: 5.253 min
Delta R.T.: 0.003 min
Response: 79650
Conc: 1.91 ng/ml



#21 AR-1248-1


R.T.: 5.349 min
Delta R.T.: 0.025 min
Response: 129177
Conc: 2.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#21 AR-1248-1

R.T.: 4.389 min
Delta R.T.: 0.003 min
Response: 59150
Conc: 1.62 ng/ml




#22 AR-1248-2

R.T.: 5.623 min
Delta R.T.: -0.002 min
Response: 108745
Conc: 1.90 ng/ml



#22 AR-1248-2


R.T.: 4.640 min
Delta R.T.: -0.001 min
Response: 20823
Conc: 0.44 ng/ml



#23 AR-1248-3


R.T.: 5.837 min
Delta R.T.: -0.013 min
Response: 230755
Conc: 3.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#23 AR-1248-3

R.T.: 4.685 min
Delta R.T.: -0.001 min
Response: 92471
Conc: 1.87 ng/ml




#24 AR-1248-4

R.T.: 6.323 min
Delta R.T.: 0.026 min
Response: 1169145
Conc: 15.03 ng/ml



#24 AR-1248-4


R.T.: 4.888 min
Delta R.T.: 0.019 min
Response: 43708
Conc: 0.75 ng/ml



#25 AR-1248-5


R.T.: 6.323 min
Delta R.T.: -0.017 min
Response: 1169145
Conc: 15.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#25 AR-1248-5

R.T.: 5.295 min
Delta R.T.: 0.002 min
Response: 26207
Conc: 0.43 ng/ml




#26 AR-1254-1

R.T.: 6.238 min
Delta R.T.: -0.030 min
Response: 69047
Conc: 0.95 ng/ml



#26 AR-1254-1


R.T.: 5.247 min
Delta R.T.: -0.003 min
Response: 38522
Conc: 0.44 ng/ml



#27 AR-1254-2


R.T.: 6.513 min
Delta R.T.: 0.003 min
Response: 337328
Conc: 2.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#27 AR-1254-2

R.T.: 5.409 min
Delta R.T.: -0.003 min
Response: 41046
Conc: 0.55 ng/ml




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.917 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 5.835 min
Delta R.T.: -0.013 min
Response: 325342
Conc: 2.65 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 7.237 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#29 AR-1254-4

R.T.: 6.104 min
Delta R.T.: 0.007 min
Response: 66449
Conc: 0.88 ng/ml




#30 AR-1254-5

R.T.: 7.712 min
Delta R.T.: 0.010 min
Response: 620620
Conc: 6.70 ng/ml



#30 AR-1254-5


R.T.: 6.558 min
Delta R.T.: 0.004 min
Response: 74263
Conc: 0.73 ng/ml



#31 AR-1260-1


R.T.: 7.102 min
Delta R.T.: 0.006 min
Response: 462731
Conc: 4.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#31 AR-1260-1

R.T.: 5.986 min
Delta R.T.: -0.007 min
Response: 50349
Conc: 0.59 ng/ml




#32 AR-1260-2

R.T.: 7.366 min
Delta R.T.: -0.018 min
Response: 165675
Conc: 1.48 ng/ml



#32 AR-1260-2


R.T.: 6.184 min
Delta R.T.: -0.015 min
Response: 18038
Conc: 0.17 ng/ml



#33 AR-1260-3


R.T.: 7.802 min
Delta R.T.: 0.021 min
Response: 1504873
Conc: 17.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#33 AR-1260-3

R.T.: 6.370 min
Delta R.T.: 0.005 min
Response: 151471
Conc: 1.53 ng/ml




#34 AR-1260-4

R.T.: 8.030 min
Delta R.T.: -0.003 min
Response: 4353785
Conc: 41.96 ng/ml



#34 AR-1260-4


R.T.: 6.883 min
Delta R.T.: 0.000 min
Response: 57529
Conc: 0.69 ng/ml



#35 AR-1260-5


R.T.: 8.395 min
Delta R.T.: 0.000 min
Response: 129777
Conc: 0.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#35 AR-1260-5

R.T.: 7.149 min
Delta R.T.: 0.000 min
Response: 42980
Conc: 0.22 ng/ml




#36 AR-1262-1

R.T.: 7.802 min
Delta R.T.: 0.022 min
Response: 1504873
Conc: 13.19 ng/ml



#36 AR-1262-1


R.T.: 6.600 min
Delta R.T.: 0.002 min
Response: 93810
Conc: 0.87 ng/ml



#37 AR-1262-2


R.T.: 8.395 min
Delta R.T.: -0.001 min
Response: 129777
Conc: 0.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#37 AR-1262-2

R.T.: 6.883 min
Delta R.T.: 0.002 min
Response: 57529
Conc: 0.58 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.739 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.465 min
Delta R.T.: 0.006 min
Response: 248514
Conc: 3.30 ng/ml



#39 AR-1262-4


R.T.: 8.837 min
Delta R.T.: -0.004 min
Response: 1071030
Conc: 18.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#39 AR-1262-4

R.T.: 7.534 min
Delta R.T.: 0.006 min
Response: 95498
Conc: 0.66 ng/ml




#40 AR-1262-5

R.T.: 9.562 min
Delta R.T.: -0.012 min
Response: 245013
Conc: 3.10 ng/ml



#40 AR-1262-5


R.T.: 8.068 min
Delta R.T.: -0.003 min
Response: 183108
Conc: 2.71 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 8.738 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#41 AR-1268-1

R.T.: 7.465 min
Delta R.T.: 0.007 min
Response: 248514
Conc: 1.06 ng/ml




#42 AR-1268-2

R.T.: 8.837 min
Delta R.T.: -0.007 min
Response: 1071030
Conc: 4.43 ng/ml



#42 AR-1268-2


R.T.: 7.534 min
Delta R.T.: 0.006 min
Response: 95498
Conc: 0.44 ng/ml



#43 AR-1268-3


R.T.: 9.118 min
Delta R.T.: 0.026 min
Response: 1553182
Conc: 7.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL




#43 AR-1268-3

R.T.: 7.754 min
Delta R.T.: -0.002 min
Response: 128427
Conc: 0.69 ng/ml




#44 AR-1268-4

R.T.: 9.562 min
Delta R.T.: -0.011 min
Response: 245013
Conc: 2.73 ng/ml



#44 AR-1268-4


R.T.: 8.068 min
Delta R.T.: -0.002 min
Response: 183108
Conc: 2.23 ng/ml



#45 AR-1268-5

R.T.: 10.023 min
Delta R.T.: -0.002 min
Response: 588109
Conc: 0.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB144739BL



#45 AR-1268-5

R.T.: 8.385 min
Delta R.T.: -0.004 min
Response: 131534
Conc: 0.22 ng/ml