

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052722\
 Data File : PP048091.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2022 12:37
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 23:31:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 27 23:28:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.038	3.222	220.4E6	193.1E6	98.673	98.765
2) SA Decachlor...	10.392	8.667	190.0E6	155.1E6	98.395	100.516
Target Compounds						
26) L6 AR-1254-1	6.266	5.250	74164736	88850957	986.159	982.016
27) L6 AR-1254-2	6.509	5.411	116.2E6	75269087	977.879	977.982
28) L6 AR-1254-3	6.917	5.848	123.1E6	125.2E6	987.427	985.480
29) L6 AR-1254-4	7.235	6.098	90286778	77505157	985.596	993.284
30) L6 AR-1254-5	7.701	6.555	97816418	106.8E6	989.423	985.138

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052722\
Data File : PP048091.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2022 12:37
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 23:31:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 27 23:28:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

