

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052725\
 Data File : PP072365.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2025 10:11
 Operator : YP\AJ
 Sample : Q2110-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 60458

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/28/2025
 Supervised By :mohammad ahmed 05/30/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 11:14:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.795	16237753	13376797	8.075m	8.369m
2) SA Decachlor...	10.199	8.813	32377993	23249766	19.874m	21.963

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052725\
Data File : PP072365.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2025 10:11
Operator : YP\AJ
Sample : Q2110-02
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
60458


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/28/2025
Supervised By :mohammad ahmed 05/30/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 11:14:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.500 min
Delta R.T.: 0.002 min
Response: 16237753
Conc: 8.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
60458


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/28/2025
Supervised By :mohammad ahmed 05/30/2025




#1 Tetrachloro-m-xylene

R.T.: 3.795 min
Delta R.T.: 0.002 min
Response: 13376797
Conc: 8.37 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.199 min
Delta R.T.: -0.013 min
Response: 32377993
Conc: 19.87 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.813 min
Delta R.T.: -0.009 min
Response: 23249766
Conc: 21.96 ng/ml