

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052824\
 Data File : PP065282.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 May 2024 21:47
 Operator : YP\AJ
 Sample : P2622-19
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 S02-04A

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/29/2024
 Supervised By :Ankita Jodhani 05/29/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 29 04:34:42 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 21 18:19:41 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.468	3.750	32291041	30806552	20.459m	20.293
2) SA Decachlor...	10.207	8.819	51272083	41712300	39.431	31.646
Target Compounds						
26) L6 AR-1254-1	6.488	5.661	233.8E6	247.9E6	3786.129	3107.321
27) L6 AR-1254-2	6.704	5.807	373.8E6	235.4E6	4074.825	3338.383
28) L6 AR-1254-3	7.074	6.215	1007.9E6	918.1E6	10834.551	8330.655
29) L6 AR-1254-4	7.355	6.442	509.5E6	401.2E6	8404.594	7107.236
30) L6 AR-1254-5	7.792f	6.862	3035.6E6	2808.1E6	43212.045	30136.137 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP052824\
Data File : PP065282.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 May 2024 21:47
Operator : YP\AJ
Sample : P2622-19
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
S02-04A

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/29/2024
Supervised By :Ankita Jodhani 05/29/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 29 04:34:42 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP052124.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 21 18:19:41 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

